



Certificate of Analysis

Product name

KIT COBAS 6800/8800 MPX 96T CE-IVD

Material No.

06997708190

Lot No.

G32991

Date of Manufacturing

2021-04-02

Expiry Date

2022-11

Parameter	Result	Unit
HIVM-2.5X-Sens/Repeat Reps-R	100	%
HBV-2.5X-Sens/Repeat Reps-R	100	%
HCV-2.5X-Sens/Repeat Reps-R	100	%
HIV1O-2.5X-Sens/Repeat Reps-R	100	%
HIV2-2.5X-Sens/Repeat Reps-R	100	%
HIVM-0.5X-SensReps-R	100	%
HBV-0.5X-SensReps-R	95	%
HCV-0.5X-SensReps-R	86	%
HIV1O-0.5X-SensReps-R	90	%
HIV2-0.5X-SensReps-R	86	%
NHP-Specificity-NR	100	%
Components	1	

Issue date 2021-05-06

Page 1 of 2

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771



Certificate of Analysis

Product name

KIT COBAS 6800/8800 MPX 96T CE-IVD

Material No.

06997708190

Lot No.

G32991

Date of Manufacturing

2021-04-02

Expiry Date

2022-11

Parameter**Result****Unit**

This lot meets our specification

Disposition by

Vanessa Nieto

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2021-05-06

Page 2 of 2

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771