

## DECLARATION OF CONFORMITY

No 271/2021

**Manufacturer:**

**FORMED PRO Sp. z o.o.**  
ul. Ceramiczna 2; 34-325 Łodygowice  
tel./fax 0048 33 863 18 00  
email: formed-pro@formed-pro.pl

www.formed-pro.pl

**We declare that:**

**Medical device:**                   **Bedsides cabinet**

**Type:**                               **EMU, EMU PLUS** - Variants are specified in the technical documentation of the product

**Class:**                               **Class I** by rule 1 according to annex IX Medical Directive 93/42/EEC

**This product carries the CE Mark, it is medical device according to Council Directive 93/42/EEC concerning medical devices and fulfils the essential requirements set out in Annex I of this Directive.**

**Procedure of assessing conformity:**     Annex VII of Medical Devices Directive 93/42/EEC

Product meets the requirements of harmonized standards with Medical Devices Directive 93/42/EEC:

**PN-EN 60601-1:2011, PN-EN ISO 14971:2012, PN-EN 1041:2010, PN-EN ISO 15223-1:2017,**

**PN-EN 62366:2008.**

Łodygowice, 08.03.2021 r.



**FORMED PRO Sp. z o.o.**  
34-325 Łodygowice, ul. Ceramiczna 2  
**Regon 241634241**  
**NIP 553-247-83-19**  
**tel. +48 33 863 18 00**  
**( 1 )**

Stamp

*Prezes Zarządu*

*mgr inż. Tomasz Kania*

.....  
Signature

President of the  
**FORMED PRO EXECUTIVE BOARD**