



贝恩医疗设备（广州）有限公司

地址：中国广州市经济技术开发区东区骏成路10号 邮编：510760

Bain Medical Equipment(Guangzhou) Co., Ltd.

Add:No.10, Juncheng Road, Eastern Section, Guangzhou Economic & Technological Development District, Guangzhou, China 510760

Tel: 86-20-82265249 Fax: 86-20-32067500

Website: www.bainmedical.com

Date: Dec 17, 2015

Declaration

To whom it may concern:

We, Bain Medical Equipment (Guangzhou) Co., Ltd., located in No.10, Juncheng Road, Eastern Section, Guangzhou Economic & Technological Development District, Guangzhou, China 510760, hereby declare that the DORA Hemodialyzers manufactured by us meet the following clearance data:

Product name: Hemodialyzer

Brand: DORA

Manufacturer: Bain Medical Equipment (Guangzhou) Co., Ltd.

Low flux

Measured with QB=300, QD=500, QF/TMP=0 (ml/min)

| Surface m2 | 1.4 | 1.6 | 1.8 | 2.0 | 2.2 |
|------------|-----|-----|-----|-----|-----|
| Urea | 232 | 239 | 246 | 251 | 261 |
| Creatinine | 207 | 212 | 225 | 230 | 240 |
| Phosphate | 168 | 180 | 190 | 198 | 206 |
| Vit.B12 | 103 | 106 | 118 | 128 | 138 |

High flux

Measured with QB=300, QD=500, QF/TMP=0 (ml/min)

| Surface m2 | 1.4 | 1.6 | 1.8 | 2.0 | 2.2 |
|------------|-----|-----|-----|-----|-----|
| Urea | 255 | 265 | 270 | 275 | 280 |
| Creatinine | 234 | 247 | 255 | 265 | 272 |
| Phosphate | 218 | 230 | 248 | 260 | 268 |
| Vit.B12 | 168 | 172 | 192 | 200 | 211 |
| Inulin | 118 | 117 | 137 | 152 | 171 |

Yours sincerely,

For and on behalf of
BAIN MEDICAL EQUIPMENT (GUANGZHOU) CO., LTD.
贝恩医疗设备（广州）有限公司

Authorized Signature(s)

Li Zhongming

Quality Manager

BAIN MEDICAL EQUIPMENT (GUANGZHOU) CO., LTD.