

Anti-HBc IgM

IgM antibodies to hepatitis B core antigen (anti-HBc IgM)

cobas[®]

REF		SYSTEM
11820567 122	100	Elecsys 2010 MODULAR ANALYTICS E170 cobas e 411 cobas e 601 cobas e 602

English

Intended use

Immunoassay for the in vitro qualitative determination of IgM antibodies to the hepatitis B core antigen in human serum and plasma.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on Elecsys and **cobas e** immunoassay analyzers.

Summary

References^{1,2,3,4,5,6,7,8,9,10}

Hepatitis B core antigen (HBcAg) is a non-glycosylated protein (p22) which forms the nucleocapsid (virus core) of the hepatitis B virus. The virus core encloses the HBV-DNA (virus genome) and the DNA-polymerase. In the cytosol of virus-producing hepatocytes the nucleocapsid is enveloped by the hepatitis B surface antigen (HBsAg) to form virions. Free HBcAg or non-enveloped virus cores are not detectable in serum.

IgM antibodies to HBcAg occur in serum during proliferation of active hepatitis B virus and can still be detected weeks to months after viral proliferation has ceased. High anti-HBc IgM concentrations can be found in acute hepatitis B and in attacks during chronic hepatitis B.

Tests for detecting anti-HBc IgM antibodies are used, in conjunction with HBsAg determinations, to identify acute hepatitis B viral infections. An acute attack of hitherto non-diagnosed chronic hepatitis B clinically resembles an acute hepatitis B infection and cannot be distinguished from this with certainty by determining the anti-HBc IgM. Follow-up studies, imaging procedures and liver biopsies are useful in differentiating between these two clinical pictures.

Test principle

μ-Capture test principle. Total duration of assay: 18 minutes.

- 1st incubation: Pretreatment of 10 μL of sample (automatically prediluted 1:400 with Diluent Universal) with anti-Fdy reagent to block specific IgG.
- 2nd incubation: Biotinylated monoclonal h-IgM-specific antibodies, HBcAg labeled with a ruthenium complex^{a)} and streptavidin-coated microparticles are added to the pretreated sample. Anti-HBc IgM antibodies present in the sample react with the ruthenium-labeled HBc antigen and the biotinylated anti-h-IgM to form a sandwich complex which becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined automatically by the software by comparing the electrochemiluminescence signal obtained from the reaction product of the sample with the signal of the cutoff value previously obtained by calibration.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)₃)²⁺

Reagents - working solutions

The reagent rackpack (M, R1, R2) is labeled as A-HBCIGM.

- M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL:
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 Pretreatment anti-HBc IgM (gray cap), 1 bottle, 10 mL:
Sample pretreatment reagent: Anti-human-Fdy-antibody (sheep)
> 0.05 mg/mL; phosphate buffer 100 mmol/L, pH 7.4; preservative.

R2 Anti-h-IgM-Ab-biotin; HBcAg-Ru(bpy)₃²⁺ (black cap), 1 bottle, 10 mL:
Biotinylated monoclonal anti-h-IgM antibody (mouse) > 600 ng/mL;
HBcAg (E. coli, rDNA), labeled with ruthenium complex > 200 ng/mL;
phosphate buffer 100 mmol/L, pH 7.4; preservative.

A-HBCIGM Cal1 Negative calibrator 1 (white cap), 2 bottles of 1.0 mL each:
Human serum, preservative.

A-HBCIGM Cal2 Positive calibrator 2 (black cap), 2 bottles of 1.0 mL each:
Anti-HBc IgM (human) > 100 PEI-U/mL^{b)} in human serum; preservative.

b) Paul-Ehrlich-Institute units

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:

n-Octyl-N,N-dimethyl-3-ammonio-1-propanesulfonate

EUH 208 May produce an allergic reaction.

Product safety labeling primarily follows EU GHS guidance.

All human material should be considered potentially infectious.

The negative calibrator (A-HBCIGM Cal1) has been prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV. The testing methods applied were FDA-approved or cleared in compliance with the European Directive 98/79/EC, Annex II, List A.

Positive calibrator (A-HBCIGM Cal2): Materials of human origin were tested for HIV and hepatitis C. The findings were negative. The serum containing anti-HBc IgM was inactivated using β-propiolactone and UV-radiation.

However, as no inactivation or testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.^{11,12}

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Reagent handling

The reagents in the kit are ready for use and are supplied in bottles compatible with the system.

Elecsys 2010 and **cobas e 411** analyzers: The calibrators should only be left on the analyzers during calibration at 20-25 °C. After use, close the bottles as soon as possible and store upright at 2-8 °C.

Due to possible evaporation effects, not more than 5 calibration procedures per bottle set should be performed.

MODULAR ANALYTICS E170, **cobas e 601** and **cobas e 602** analyzers: Unless the entire volume is necessary for calibration on the analyzers, transfer aliquots of the ready-for-use calibrators into empty snap-cap bottles (CalSet Vials). Attach the supplied labels to these additional bottles. Store the aliquots at 2-8 °C for later use.

Perform **only one** calibration procedure per aliquot.

All information required for correct operation is read in from the respective reagent barcodes.

Anti-HBc IgM

IgM antibodies to hepatitis B core antigen (anti-HBc IgM)

cobas[®]

Please note: Both the vial labels, and the additional labels (if available) contain 2 different barcodes. The barcode between the yellow markers is for **cobas** 8000 systems only. If using a **cobas** 8000 system, please turn the vial cap 180° into the correct position so the barcode can be read by the system. Place the vial on the instrument as usual.

Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability of the reagent rackpack	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	8 weeks
on Elecsys 2010 and cobas e 411	4 weeks
on MODULAR ANALYTICS E170, cobas e 601 and cobas e 602	8 weeks

Stability of the calibrators	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	8 weeks
on Elecsys 2010 and cobas e 411 at 20-25 °C	up to 5 hours
on MODULAR ANALYTICS E170, cobas e 601 and cobas e 602 at 20-25 °C	use only once

Store calibrators **upright** in order to prevent the calibrator solution from adhering to the snap-cap.

Specimen collection and preparation

Only the specimens listed below were tested in a sufficient number and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel.

Sodium heparin, Li-heparin, K₃-EDTA and sodium citrate plasma. Do not use plasma treated with sodium fluoride and potassium oxalate.

Criterion: Correct assignment of negative and positive samples.

Stable for 6 days at 2-8 °C, 3 months at -20 °C. The samples may be frozen 5 times.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates and frozen samples before performing the assay. Heat-inactivated samples may be used.

Do not use samples and controls stabilized with azide.

Ensure the samples, calibrators and controls are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples and calibrators on the analyzers should be analyzed/measured within 2 hours.

Materials provided

See "Reagents – working solutions" section for reagents.

- 2 x 6 bottle labels

Materials required (but not provided)

- [REF] 11876333122, PreciControl Anti-HBc IgM, 8 x 1.0 mL each of PreciControl Anti-HBc IgM 1 and 2
- [REF] 11732277122, Diluent Universal, 2 x 16 mL sample diluent or [REF] 03183971122, Diluent Universal, 2 x 36 mL sample diluent
- [REF] 11776576322, CalSet Vials, 2 x 56 empty snap-cap bottles

- General laboratory equipment
 - Elecsys 2010, MODULAR ANALYTICS E170 or **cobas e** analyzer
- Accessories for Elecsys 2010 and **cobas e** 411 analyzers:
- [REF] 11662988122, ProCell, 6 x 380 mL system buffer
 - [REF] 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
 - [REF] 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
 - [REF] 11933159001, Adapter for SysClean
 - [REF] 11706802001, Elecsys 2010 AssayCup, 60 x 60 reaction vessels
 - [REF] 11706799001, Elecsys 2010 AssayTip, 30 x 120 pipette tips

Accessories for MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers:

- [REF] 04880340190, ProCell M, 2 x 2 L system buffer
- [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- [REF] 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use
- [REF] 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change
- [REF] 03004899190, PreClean M, 5 x 600 mL detection cleaning solution
- [REF] 12102137001, AssayTip/AssayCup Combimagazine M, 48 magazines x 84 reaction vessels or pipette tips, waste bags
- [REF] 03023150001, WasteLiner, waste bags
- [REF] 03027651001, SysClean Adapter M

Accessories for all analyzers:

- [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers.

MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers: PreClean M solution is necessary.

Bring the cooled reagents to approximately 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

Place the calibrators in the sample zone.

All the information necessary for calibrating the assay is automatically read into the analyzer.

After calibration has been performed, store the calibrators at 2-8 °C or discard (MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers).

Calibration

Traceability: This method has been standardized against the "HBc Reference Serum 84 (anti-HBc IgM)" of the Paul-Ehrlich-Institute, Langen (Germany). For the Elecsys Anti-HBc IgM assay, the cutoff (cutoff index 1.0) was set to approx. 100 PEI-U/mL.¹³

Calibration frequency: Calibration must be performed once per reagent lot using A-HBCIGM Cal1, A-HBCIGM Cal2 and fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer). Renewed calibration is recommended as follows:

- after 1 month (28 days) when using the same reagent lot
- after 7 days (when using the same reagent kit on the analyzer)
- as required: e.g.: quality control findings with PreciControl Anti-HBc IgM outside the defined limits
- more frequently when this is required by pertinent regulations.

Anti-HBc IgM

IgM antibodies to hepatitis B core antigen (anti-HBc IgM)



Range for the electrochemiluminescence signals (counts) for the calibrators:

Negative calibrator (A-HBCIGM Cal1): 600-3500 (Elecsys 2010 and **cobas e 411** analyzers), 400-3500 (MODULAR ANALYTICS E170, **cobas e 601** and **cobas e 602** analyzers).

Positive calibrator (A-HBCIGM Cal2): 24000-150000 (Elecsys 2010 and **cobas e 411** analyzers), 18000-130000 (MODULAR ANALYTICS E170, **cobas e 601** and **cobas e 602** analyzers).

Quality control

For quality control, use PreciControl Anti-HBc IgM.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

If necessary, repeat the measurement of the samples concerned.

Follow the applicable government regulations and local guidelines for quality control.

Note:

For technical reasons re-assigned target values valid only for a specific reagent and control lot combination, must be entered manually on all analyzers (except for the **cobas e 602** analyzer). Therefore always refer to the value sheet included in the reagent kit or PreciControl kit to make sure that the correct target values are used.

When a new reagent or control lot is used, the analyzer will use the original values encoded in the control barcodes.

Calculation

The analyzer automatically calculates the cutoff based on the measurement of A-HBCIGM Cal1 and A-HBCIGM Cal2.

The result of a sample is given either as reactive or non-reactive as well as in the form of a cutoff index (signal sample/cutoff).

Interpretation of the results

Samples with a cutoff index ≥ 1.0 are reactive in the Elecsys Anti-HBc IgM assay. These samples are considered positive for anti-HBc IgM.

Samples with a cutoff index < 1.0 are non-reactive in the Elecsys Anti-HBc IgM assay. These samples are considered negative.

Note: According to the recommendations of the Paul-Ehrlich-Institute, Langen (Germany), an equivocal range should be allowed for the assessment of results from anti-HBc IgM tests.

For the Elecsys Anti-HBc IgM assay the equivocal cutoff index range is 0.9-1.1.

Limitations - interference

The assay is unaffected by icterus (bilirubin $< 428 \mu\text{mol/L}$ or $< 25 \text{ mg/dL}$), hemolysis (Hb $< 1.2 \text{ mmol/L}$ or $< 2.0 \text{ g/dL}$), lipemia (Intralipid $< 1500 \text{ mg/dL}$) and biotin ($< 409 \text{ nmol/L}$ or $< 100 \text{ ng/mL}$).

Criterion: Correct assignment of negative and positive samples.

Samples should not be taken from patients receiving therapy with high biotin doses (i.e. $> 5 \text{ mg/day}$) until at least 8 hours following the last biotin administration.

No interference was observed from rheumatoid factors up to a concentration of 4200 IU/mL.

In vitro tests were performed on 19 commonly used pharmaceuticals. No interference with the assay was found.

In rare cases, interference due to extremely high titers of antibodies to immunological components, streptavidin and ruthenium can occur.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

Limits and ranges

Detection limit: $\leq 3.0 \text{ PEI-U/mL}$

The detection limit represents the lowest measurable analyte level that can be distinguished from zero. It is calculated as the value lying two standard deviations above that of the negative calibrator (negative calibrator + 2 SD, repeatability study, $n = 21$).

Dilution

Use Diluent Universal for automatic sample predilution.

Expected values

For the Elecsys Anti-HBc IgM assay, the cutoff (cutoff index 1.0) was set to approx. 100 PEI-U/mL. In acute HBV infections the anti-HBc IgM level is generally far above this limit. After recovery from hepatitis B disease the anti-HBc IgM levels are below this. Chronic hepatitis can produce values in the vicinity of the cutoff.

Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using Elecsys reagents, human sera and controls.

Elecsys 2010 and cobas e 411 analyzers						
	Repeatability ^{c)}			Intermediate precision ^{d)}		
Sample	Mean COI ^{e)}	SD COI	CV %	Mean COI	SD COI	CV %
HS ^{f)} , negative	0.123	0.003	2.2	0.069	0.002	2.8
HS, weakly positive	1.14	0.040	3.5	1.15	0.021	1.8
HS, positive	3.58	0.131	3.7	3.98	0.135	3.4
PC ^{g)} A-HBCIGM1	0.053	0.001	1.7	0.063	0.002	3.3
PC A-HBCIGM2	1.39	0.063	4.5	1.55	0.056	3.6

c) Repeatability = within-run precision ($n = 20/21$)

d) Intermediate precision = between-run ($n = 10$)

e) COI = cutoff index

f) HS = human serum

g) PC = PreciControl

MODULAR ANALYTICS E170, cobas e 601 and cobas e 602 analyzers						
	Repeatability ^{h)}			Intermediate precision ⁱ⁾		
Sample	Mean COI	SD COI	CV %	Mean COI	SD COI	CV %
HS, negative	0.037	0.001	3.4	0.037	0.002	3.9
HS, weakly positive	1.32	0.032	2.4	1.33	0.054	4.1
HS, positive	4.91	0.080	1.6	5.11	0.171	3.3
PC A-HBCIGM1	0.032	0.001	1.5	0.042	0.001	1.9
PC A-HBCIGM2	1.67	0.043	2.6	1.80	0.069	3.8

h) Repeatability = within-run precision ($n = 20/21$)

i) Intermediate precision = within-laboratory (modified protocol (EP5-A) of the CLSI (Clinical and Laboratory Standards Institute): 6 times daily for 10 days ($n = 60$))

Analytical specificity

No cross-reactions with HAV, HCV, HIV 1+2, CMV, EBV, HSV, E. coli, Toxoplasma gondii, Rubella, and Treponema pallidum were observed.

Measurements were performed on each of the pathogens listed above using ≥ 8 serum or plasma samples which were positive for antibodies to the above-mentioned pathogens or contained autoantibodies (SLE, ANA).

Cutoff sensitivity

Approx. 100 PEI-U/mL for the Elecsys Anti-HBc IgM assay. Assays of other manufacturers may be set differently.

Anti-HBc IgM

IgM antibodies to hepatitis B core antigen (anti-HBc IgM)

cobas[®]

Clinical sensitivity

	Elecys Anti-HBc IgM assay			Anti-HBc IgM comparison test		
	pos ^{j)}	neg ^{k)}	discrepant	pos	neg	discrepant
Acute HBV infection clinically and serologically manifested	48	-	-	48	-	-
Acute HBV infection clinically manifested	31	4	3 (neg)	31	4	3 (pos) ^{j)}
Acute HBV infection serologically manifested	57	6	16 (neg) 1 (pos)	57	6	16 (pos) 1 (neg)
Serologically manifested no clinical information	145	292	44 (neg)	145	292	44 (pos)

j) positive

k) negative

l) In the comparison test, the discrepant samples were weakly positive.

Clinical specificity

To investigate the specificity, samples from randomly selected blood donors were tested with the Elecys Anti-HBc IgM assay in comparison to licensed enzyme immunoassays.

1003/1003 samples from blood donors were negative with the Elecys Anti-HBc IgM assay (100 % specificity for this cohort).

990/1003 were negative with a comparison test (98.7 % specificity).

242/242 samples from hospitalized patients, pregnant women, and dialysis patients with no indication of a HBV infection were negative with both the Elecys Anti-HBc IgM assay and the comparison test (100 % specificity for this cohort).

References

- Gudat F, Bianchi L, Sonnabend W, et al. Pattern of core and surface expression in liver tissue reflects state of specific immune response in hepatitis B. *Laboratory investigations* 1975;32(1):1-9.
- Gerlich WH, Robinson WS. Hepatitis B virus contains protein attached to the 5' terminus of its complete DNA strand. *Cell* 1980;21:801-809.
- Gerlich WH, Goldmann U, Müller R, et al. Specificity and localization of the hepatitis B virus-associated protein kinase. *Journal of Virology* 1982;42(3):761-766.
- Roossinck MJ, Siddiqui A. In vivo phosphorylation and protein analysis of hepatitis B virus core antigen. *J Virol* 1987;61:955-961.
- Gallina A, Bonelli F, Zentilin L, et al. A recombinant hepatitis B core antigen polypeptide with the protamine-like domain deleted self-assembles into capsid particles but fails to bind nucleic acids. *J Virol* 1989;63(11):4645-4652.
- Zhou S, Stranding DN. Hepatitis B virus capsid particles are assembled from core-protein dimer precursors. *Proc Natl Acad Sci USA* 1992;89:10046-10050.
- Bimbaum BS, Nassal M. Hepatitis B virus nucleocapsid assembly: primary structure requirements in the core protein. *J Virol* 1990;64(7):3025-3330.
- Hollinger FB. Hepatitis B virus. In Fields BN, Knipe DM (eds) *Virology* 2nd ed., New York, Raven Press 1990:2171-2236.
- Frösner G. *Moderne Hepatitisdiagnostik*. Kilian Verlag, Marburg 1996.
- Hoofnagle Biscieglie MD. Serologic diagnosis of acute and chronic viral hepatitis. *Seminars in Liver Disease* 1991;11(2):73-83.
- Occupational Safety and Health Standards: bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.

- Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.
- Hadziyannis JS, Hadziyannis AS, Dourakis S, et al. Clinical Significance of Quantitative Anti-HBc IgM assay in Acute and Chronic HBV Infection. *Hepato Gastroenterol* 1993;40:588-592.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

	Contents of kit
	Analyzers/Instruments on which reagents can be used
	Reagent
	Calibrator
	Volume after reconstitution or mixing
	Global Trade Item Number

COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2015, Roche Diagnostics

0123



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

