



04840763001V7.0

TSH CalSet

cobas[®]

REF 04738551 190

4 x 1.3 mL

English

Intended use

TSH CalSet is used for calibrating the quantitative Elecsys TSH assay on the Elecsys and **cobas e** immunoassay analyzers.

Summary

TSH CalSet is an equine serum matrix (TSH Cal1) and a human serum matrix with human TSH (TSH Cal2) in two concentration ranges.

The CalSet can be used with all reagent lots.

Reagents - working solutions

- TSH Cal1: 2 bottles, each containing 1.3 mL of calibrator 1
- TSH Cal2: 2 bottles, each containing 1.3 mL of calibrator 2

The concentration of TSH Cal1 in the equine serum matrix is approx. 0 µIU/mL; TSH Cal2 contains approx. 1.5 µIU/mL TSH (human) in the human serum matrix.

The exact lot-specific calibrator values are encoded in the barcode as well as printed on the enclosed (or electronically available) calibrator barcode sheet.

Calibrator values

Traceability: The Elecsys TSH assay has been standardized against the 2nd IRP WHO Reference Standard 80/558.

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV. The testing methods applied were FDA-approved or cleared in compliance with the European Directive 98/79/EC, Annex II, List A.

However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.^{1,2}

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Handling

The calibrators are supplied ready-for-use in bottles compatible with the system.

Elecsys 2010 and **cobas e** 411 analyzers: The calibrators should only be left on the analyzers during calibration at 20-25 °C. After use, close the bottles as soon as possible and store at 2-8 °C.

Due to possible evaporation effects, not more than 5 calibration procedures per bottle set should be performed.

MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers: Unless the entire volume is necessary for calibration on the analyzers, transfer aliquots of the ready-for-use calibrators into empty snap-cap bottles (CalSet Vials). Attach the supplied labels to these additional bottles. Store the aliquots at 2-8 °C for later use.

Perform **only one** calibration procedure per aliquot.

Storage and stability

Store at 2-8 °C.

Stability:	
unopened at 2-8 °C	up to the stated expiration date
after opening/in aliquots at 2-8 °C	12 weeks
on Elecsys 2010 and cobas e 411 analyzers at 20-25 °C	up to 5 hours

Stability:	
on MODULAR ANALYTICS E170, cobas e 601 and cobas e 602 analyzers	use only once

Store calibrators **upright** in order to prevent the calibrator solution from adhering to the snap-cap.

Materials provided

- TSH CalSet, barcode card, calibrator barcode sheet, 2 x 4 bottle labels

Materials required (but not provided)

- REF 11776576322, CalSet Vials, 2 x 56 empty snap-cap bottles
- Elecsys 2010, MODULAR ANALYTICS E170 or **cobas e** immunoassay analyzers and Elecsys TSH assay reagents.

See the assay Method Sheet and the operator's manual for additionally required material.

Assay

Place the bottles in the sample zone.

Read in all the information necessary for calibrating the assay.

Ensure the calibrators are at 20-25 °C prior to measurement.

References

- 1 Occupational Safety and Health Standards: bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- 2 Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

CONTENT	Contents of kit
SYSTEM	Analyzers/Instruments on which reagents can be used
REAGENT	Reagent
CALIBRATOR	Calibrator
→	Volume after reconstitution or mixing

FOR US CUSTOMERS ONLY: LIMITED WARRANTY

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS E, ELECSYS and MODULAR are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Significant additions or changes are indicated by a change bar in the margin.

© 2013, Roche Diagnostics



04840763001V7.0

TSH CalSet

cobas®



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com



Distribution in USA by:

Roche Diagnostics, Indianapolis, IN

US Customer Technical Support 1-800-428-2336

