

**EG-Konformitätserklärung/EC Declaration of Conformity**

gemäß Anhang III der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998  
*as per Annex III of Directive 98/79/EC of the European Parliament and Council of 27 October 1998*

**Hersteller/Manufacturer:** Roche Diagnostics GmbH  
**Adresse/Address:** Roche Centralized Diagnostics  
Sandhofer Straße 116  
D-68305 Mannheim

Die Roche Diagnostics GmbH erklärt, dass das Produkt/die Produktfamilie (bei rezepturgleichen Produkten)  
*Roche Diagnostics GmbH declares that the product/the product line (in case of products manufactured by identical recipes)*

**Produktname/Product name:** ISE Diluent Gen.2

**Art.-Nr./Id. No.:** 04522630

**Beschreibung/Description:** Das ISE-Modul der Roche/Hitachi **cobas c** Systeme dient zur quantitativen Bestimmung von Natrium, Kalium und Chlorid in Serum, Plasma oder Urin mittels ionenselektiver Elektroden.  
*The ISE module of the Roche/Hitachi cobas c systems is intended for the quantitative determination of sodium, potassium, and chloride in serum, plasma or urine using ion-selective electrodes.*

auf das/die sich diese Erklärung bezieht, den Forderungen der EG-Richtlinie 98/79/EG des Rates vom 27. Oktober 1998 (bzw. seine Umsetzung in nationales Recht der Mitgliedsstaaten in welchen das Produkt vermarktet werden soll) über In-vitro-Diagnostica entspricht.  
*to which this declaration relates fulfils the requirements of EC Directive 98/79/EC of the Council of 27 October 1998 (and its relevant transposition into the national laws of the Member States in which the device is intended to be placed on the market) concerning in-vitro diagnostic devices.*

Mannheim, 12.06.2006

Roche Diagnostics GmbH

ppa./on behalf of the company i. V./on behalf of the company

Dr. M. Thein  
Head of Quality Management &  
Regulatory Affairs  
Centralized Diagnostics

A. Schenkel  
Head of Quality Assurance  
Centralized Diagnostics

**Kontaktadresse/Contact address:** Roche Centralized Diagnostics  
Abt./Dept. Regulatory Affairs  
Sandhofer Straße 116  
D-68305 Mannheim  
Fax: +49 621/759 1448

isediluentgen2\_1.doc-AJ