



07052472001V1.0

PreciControl Syphilis

cobas[®]

REF 06923364 190

→ 4 x 2.0 mL

English

Intended use

PreciControl Syphilis is used for quality control of the Elecsys Syphilis immunoassay on the Elecsys and **cobas e** immunoassay analyzers.

Summary

PreciControl Syphilis is a lyophilized control serum based on human serum. The controls are used for monitoring the accuracy of the Elecsys Syphilis immunoassay.

Reagents - working solutions

- PC Syphilis1: 2 bottles, each for 2.0 mL of control serum
Human serum, negative for syphilis antibodies; preservative.
Target value for the cutoff index: approximately 0.15
- PC Syphilis2: 2 bottles, each for 2.0 mL of control serum
Human serum, positive for syphilis IgG; preservative.
Target value for the cutoff index: approximately 5.00

Note: The controls are not barcode-labeled and therefore have to be run like external controls. All values and ranges have to be entered manually. Please refer to the section "QC" in the operator's manual or to the online help of the instrument software.

Non-barcode labeled controls: Only one target value and range for each control level can be entered in the analyzer. The reagent lot-specific target values have to be re-entered each time when a specific reagent lot with different control target values and ranges is used. Two reagent lots with different control target values and ranges cannot be used in parallel in the same run.

The exact lot-specific target values and ranges are printed on the enclosed (or electronically available) value sheet included in the reagent kit or PreciControl kit.

Please make sure that the correct values are used.

Target values and ranges

The target values and ranges were determined and evaluated by Roche. They were obtained using the Elecsys Syphilis assay reagents and analyzers available at the time of testing.

Results must be within the specified ranges. In the event that increasing or decreasing trends, or any other suddenly occurring deviations beyond the range limits are observed, all test steps must be checked.

When necessary, measurement of the patient sample tested should be repeated.

Traceability information is given in the Method Sheet of the relevant Elecsys assay.

Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV.

The testing methods applied were FDA-approved or cleared in compliance with the European Directive 98/79/EC, Annex II, List A.

However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.^{1,2}

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Handling

Carefully dissolve the contents of one bottle by adding exactly 2.0 mL of distilled or deionized water and allow to stand closed for 15 minutes to reconstitute. Mix carefully, avoiding foam formation.

When measuring a non-barcoded control, use only recommended sample tubes, cup on tube or cup on rack.

Transfer an aliquot of the reconstituted control into an appropriate tube and use the aliquot immediately.

Perform **only one** control procedure per aliquot.

Close the glass vial carefully and store the remainder of the reconstituted control at 2-8 °C or freeze immediately at -20 °C for later use.

Due to possible evaporation effects, not more than 5 quality control procedures per glass vial should be performed.

Storage and stability

Store at 2-8 °C.

The lyophilized control serum is stable up to the stated expiration date.

Stability of the reconstituted control serum:	
either at -20 °C	6 months (3 freeze/thaw cycles possible)
or at 2-8 °C	28 days
on the analyzers at 20-25 °C	up to 5 hours

Store controls **upright** in order to prevent the control solution from adhering to the lid of the glass vial.

Materials provided

- PreciControl Syphilis

Materials required (but not provided)

- Elecsys 2010, MODULAR ANALYTICS E170 or **cobas e** immunoassay analyzers and assay reagents
- Distilled or deionized water

See the assay Method Sheet and the operator's manual for additionally required material.

Assay

Treat the reconstituted control serum in the system-compatible labeled bottles for analysis in the same way as patient samples.

The control values and ranges have to be entered manually. Please refer to the corresponding section in the operator's manual.

Ensure the controls are at 20-25 °C prior to measurement.

Run controls daily in parallel with patient samples, once per reagent kit, and whenever a calibration is performed. The control intervals and limits should be adapted to each laboratory's individual requirements.

Follow the applicable government regulations and local guidelines for quality control.

References

- Occupational Safety and Health Standards: bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

CONTENT	Contents of kit
SYSTEM	Analyzers/Instruments on which reagents can be used
REAGENT	Reagent
CALIBRATOR	Calibrator



PreciControl Syphilis

cobas[®]

Volume after reconstitution or mixing

COBAS, COBAS E, ELECSYS, MODULAR and PRECICONTROL are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Significant additions or changes are indicated by a change bar in the margin.

© 2013, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

