

BLAYCO®

Adhesive electrode cleaning device

References

TCG01: Electrode cleaning device

**Date**

January 2012

Specification N.

ETP/024/6

TECHNICAL SPECIFICATIONS**Materials**

- Foam: Polyurethane.
- Scratch surface: Silicon dioxide.
- Adhesive: Latex free.
- Release liner: Siliconized paper.
- Radioopaque check system: Barium thread.
- Pouch:
 - Cellulose, medical grade
 - Film PET/PE

**Physical and dimensional properties**

- Surface (mm) : 50 x 50
- Thickness (mm): $6,0 \pm 0,5$
- Total weight: 3g

Sterilization and expiry date

- Sterile by ethylene oxide
- Expiry date: 3 years

Biocompatibility

- No cytotoxic (ISO 10993-5)
- Negative intracutaneous irritation test (ISO 10993-10)
- No sensitizing (ISO 10993-10)

Packaging

- 1 unit per pouch
- 250 pouches per service box

