
From: [REDACTED]
Date: 2023-06-12 17:29
To: [REDACTED]
CC: [REDACTED]
Subject: Re: Re: new tender

Dear [REDACTED],

Glad to hear from you. Ambulanc (Shenzhen) Tech. Co., Ltd as the manufacturer declares that E6 CPR machine is compliant with below requirements:

- 1. The principle of operation is chest compression (presses) with active decompression (lifts). Decompression is performed by suctioning the membrane (vacuum) after each compression.*
- 2. Wireless connection for data (CPR process report) transfer*

Best regards,

Best Regards,



<http://en.amoulmed.com/>

Ambulanc (Shenzhen) Tech. Co., Ltd

Address: 8F, Building C, Skyworth Innovation Ind. Park, BaoAn District, ShenZhen 518108, P.R.C

Whatsapp: [REDACTED]

Email: [REDACTED]