

Order information

REF	CONTENT	Analyzer(s) on which cobas c pack(s) can be used
08056960190	Bilirubin Total Gen.3 (1050 tests)	System-ID 2031 001 cobas c 303, cobas c 503
10759350190	Calibrator f.a.s. (12 x 3 mL)	Code 20401
10158046122	Precibil (4 x 2 mL)	Code 20306
05117003190	PreciControl ClinChem Multi 1 (20 x 5 mL)	Code 20391
05947626190	PreciControl ClinChem Multi 1 (4 x 5 mL)	Code 20391
05117216190	PreciControl ClinChem Multi 2 (20 x 5 mL)	Code 20392
05947774190	PreciControl ClinChem Multi 2 (4 x 5 mL)	Code 20392
08063494190	Diluent NaCl 9 % (123 mL)	System-ID 2906 001

English**System information****BILT3: ACN 20310****Intended use**

In vitro test for the quantitative determination of total bilirubin in serum and plasma of adults and neonates on Roche/Hitachi **cobas c** systems.

Summary¹

Bilirubin is formed in the reticuloendothelial system during the degradation of aged erythrocytes. The heme portion from hemoglobin and from other heme-containing proteins is removed, metabolized to bilirubin, and transported as a complex with serum albumin to the liver. In the liver, bilirubin is conjugated with glucuronic acid for solubilization and subsequent transport through the bile duct and elimination via the digestive tract.

Diseases or conditions which, through hemolytic processes, produce bilirubin faster than the liver can metabolize it, cause the levels of unconjugated (indirect) bilirubin to increase in the circulation. Liver immaturity and several other diseases in which the bilirubin conjugation mechanism is impaired cause similar elevations of circulating unconjugated bilirubin. Bile duct obstruction or damage to hepatocellular structure causes increases in the levels of both conjugated (direct) and unconjugated (indirect) bilirubin in the circulation.

Test principle²

Colorimetric diazo method

Total bilirubin, in the presence of a suitable solubilizing agent, is coupled with 3,5-dichlorophenyl diazonium in a strongly acidic medium.



The color intensity of the red azo dye formed is directly proportional to the total bilirubin and can be determined photometrically.

Reagents - working solutions**R1** Phosphate: 50 mmol/L; detergents; stabilizers; pH 1.0**R3** 3,5-dichlorophenyl diazonium salt: ≥ 1.35 mmol/L

R1 is in position B and R3 is in position C.

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:



Danger

H290 May be corrosive to metals.**H314** Causes severe skin burns and eye damage.**H360FD** May damage fertility. May damage the unborn child.**Prevention:****P201** Obtain special instructions before use.**P280** Wear protective gloves/ protective clothing/ eye protection/ face protection.**Response:****P303 + P361 + P353** IF ON SKIN (or hair): Take off immediately all contaminated clothing. Rinse skin with water.**P304 + P340 + P310** IF INHALED: Remove person to fresh air and keep comfortable for breathing. Immediately call a POISON CENTER/doctor.**P305 + P351 + P338 + P310** IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a POISON CENTER/doctor.**P308 + P313** IF exposed or concerned: Get medical advice/attention.

Product safety labeling follows EU GHS guidance.

Contact phone: all countries: +49-621-7590

Reagent handling

Ready for use

Storage and stabilityShelf life at 2-8 °C: See expiration date on **cobas c** pack label.

On-board in use and refrigerated on the analyzer: 6 weeks

Specimen collection and preparation

For specimen collection and preparation only use suitable tubes or collection containers.

Only the specimens listed below were tested and found acceptable.

Serum

Plasma: Li-heparin and K₂-, K₃-EDTA plasma

(The use of EDTA-plasma with elevated hematocrit may lead to slightly lower values.)

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

See the limitations and interferences section for details about possible sample interferences.

BILT3

Bilirubin Total Gen.3



Stability: ^{a),3}	1 day at 15-25 °C
	7 days at 2-8 °C
	6 months at (-15)-(-25) °C

a) If care is taken to prevent exposure to light

Materials provided

See "Reagents – working solutions" section for reagents.

Materials required (but not provided)

See "Order information" section

General laboratory equipment

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

The performance of applications not validated by Roche is not warranted and must be defined by the user.

Application for serum and plasma

Test definition

Reporting time	10 min		
Wavelength (sub/main)	600/546 nm		
Reagent pipetting		Diluent (H ₂ O)	
R1	78 µL	–	
R3	16 µL	–	
Sample volumes	Sample	Sample dilution	
		Sample	Diluent (NaCl)
Normal	1.3 µL	–	–
Decreased	2.6 µL	20 µL	60 µL
Increased	1.3 µL	–	–

For further information about the assay test definitions refer to the application parameters setting screen of the corresponding analyzer and assay.

Calibration

Calibrators	S1: H ₂ O S2: C.f.a.s.
Calibration mode	Linear
Calibration frequency	Automatic full calibration - after reagent lot change Full calibration - as required following quality control procedures

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Traceability: The method was standardized against the Doumas method.⁴

Quality control

For quality control, use control materials as listed in the "Order information" section. In addition, other suitable control material can be used.

The control intervals and limits should be adapted to each laboratory's individual requirements. It is recommended to perform quality control always after lot calibration and subsequently at least every 6 weeks.

Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

Calculation

cobas c systems automatically calculate the analyte concentration of each sample in the unit µmol/L (mg/dL, mg/L).

Conversion factors:	µmol/L x 0.0585 = mg/dL
	µmol/L x 0.585 = mg/L

Limitations - interference

Criterion: Recovery within ± 3.4 µmol/L (0.199 mg/dL) of initial values of samples ≤ 34 µmol/L (1.99 mg/dL) and within ± 10 % for samples > 34 µmol/L.

Hemolysis:⁵ No significant interference up to an H index of 800 (approximate hemoglobin concentration: 497 µmol/L or 800 mg/dL).

Immunoglobulins: No significant interference from immunoglobulins up to a concentration of 28 g/L (187 µmol/L) (simulated by human immunoglobulin G).

Criterion: Recovery within ± 1.7 µmol/L (0.099 mg/dL) of initial values of samples ≤ 17 µmol/L (0.995 mg/dL) and within ± 10 % for samples > 17 µmol/L.

Hemolysis in neonates:⁵ No significant interference up to an H index of 1000 (approximate hemoglobin concentration: 621 µmol/L or 1000 mg/dL).

Lipemia (Intralipid):⁵ No significant interference up to an L index of 1000. There is poor correlation between the L index (corresponds to turbidity) and triglycerides concentration.

Drugs: No interference was found at therapeutic concentrations using common drug panels.^{6,7}

Indican: No significant interference from indican up to a concentration of 0.12 mmol/L (3 mg/dL).

Cyanokit (Hydroxocobalamin) may cause falsely low results.

Samples containing indocyanine green must not be measured.

Results from certain multiple myeloma patients may show a positive bias in recovery. Not all multiple myeloma patients show the bias and the severity of the bias may vary between patients.

In very rare cases, gammopathy, in particular type IgM (Waldenström's macroglobulinemia), may cause unreliable results.⁸

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

In certain cases specimens may give a direct bilirubin result slightly greater than the total bilirubin result. This is observed in patient samples when nearly all the reacting bilirubin is in the direct form. In such cases the result for the total bilirubin should be reported for both D-bilirubin and total bilirubin values.

ACTION REQUIRED

Special Wash Programming: The use of special wash steps is mandatory when certain test combinations are run together on **cobas c** systems. All special wash programming necessary for avoiding carry-over is available via the **cobas** link. The latest version of the carry-over evasion list can be found with the NaOHD/SMS/SCCS Method Sheet for information. For further instructions refer to the operator's manual.

Limits and ranges

Measuring range

2.5-650 µmol/L (0.146-38.0 mg/dL)

Determine samples having higher concentrations via the rerun function. Dilution of samples via the rerun function is a 1:2 dilution. Results from samples diluted using the rerun function are automatically multiplied by a factor of 2.

Lower limits of measurement

Limit of Blank, Limit of Detection and Limit of Quantitation

Limit of Blank	= 1.7 µmol/L (0.099 mg/dL)
Limit of Detection	= 2.5 µmol/L (0.146 mg/dL)
Limit of Quantitation	= 2.5 µmol/L (0.146 mg/dL)

The Limit of Blank, Limit of Detection and Limit of Quantitation were determined in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP17-A2 requirements.

The Limit of Blank is the 95th percentile value from $n \geq 60$ measurements of analyte-free samples over several independent series. The Limit of Blank corresponds to the concentration below which analyte-free samples are found with a probability of 95 %.

The Limit of Detection is determined based on the Limit of Blank and the standard deviation of low concentration samples.

The Limit of Detection corresponds to the lowest analyte concentration which can be detected (value above the Limit of Blank with a probability of 95 %).

The Limit of Quantitation is the lowest analyte concentration that can be reproducibly measured with a total error of 30 %. It has been determined using low concentration bilirubin samples.

Expected values**μmol/L**

Adults ⁹	up to 21 μmol/L
Children with age ≥ 1 month ⁹	up to 17 μmol/L
Reference range study with 500 well-characterized human serum samples: ¹⁰	
Males	up to 24 μmol/L
Females	up to 15 μmol/L
High risk for developing clinically significant hyperbilirubinemia:	
Newborns: Term and near-term ¹¹	

Age of newborn:

24 hours	≥ 137 μmol/L ^{b)}
48 hours	≥ 222 μmol/L ^{b)}
84 hours	≥ 290 μmol/L ^{b)}

b) 95th percentile

Levels > 95th percentile: Such levels of hyperbilirubinemia have been deemed significant and are generally considered to require close supervision, possible further evaluation, and sometimes intervention.

mg/dL


Adults ⁹	up to 1.2 mg/dL
Children with age ≥ 1 month ⁹	up to 1.0 mg/dL
Reference range study with 500 well-characterized human serum samples: ¹⁰	
Males	up to 1.4 mg/dL
Females	up to 0.9 mg/dL
High risk for developing clinically significant hyperbilirubinemia:	
Newborns: Term and near-term ¹¹	

Age of newborn:

24 hours	≥ 8.0 mg/dL ^{b)}
48 hours	≥ 13.0 mg/dL ^{b)}
84 hours	≥ 17.0 mg/dL ^{b)}

b) 95th percentile

Levels > 95th percentile: Such levels of hyperbilirubinemia have been deemed significant and are generally considered to require close supervision, possible further evaluation, and sometimes intervention.

Nomogram for designation of risk in 2840 well newborns¹¹**Serum bilirubin*** 95th percentile**A** High risk zone **C** Low intermediate risk zone**B** High intermediate risk zone **D** Low risk zone

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

Specific performance data

Representative performance data on the analyzers are given below. These data represent the performance of the analytical procedure itself.

Results obtained in individual laboratories may differ due to heterogenous sample materials, aging of analyzer components and mixture of reagents running on the analyzer.

Precision

Precision was determined using human samples and controls in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP05-A3 requirements with repeatability (n = 84) and intermediate precision (2 aliquots per run, 2 runs per day, 21 days). Results for repeatability and intermediate precision were obtained on the **cobas c 503** analyzer.

Repeatability	Mean	SD	CV
	μmol/L	μmol/L	%
PCCC1 ^{c)}	16.2	0.256	1.6
PCCC2 ^{d)}	61.4	0.315	0.5
Human serum 1	5.43	0.211	3.9
Human serum 2	21.5	0.228	1.1
Human serum 3	91.6	0.507	0.6
Human serum 4	295	1.24	0.4
Human serum 5	519	1.97	0.4
Intermediate precision	Mean	SD	CV
	μmol/L	μmol/L	%
PCCC1 ^{c)}	16.2	0.372	2.3
PCCC2 ^{d)}	60.9	0.630	1.0
Human serum 1	5.43	0.222	4.1
Human serum 2	21.4	0.269	1.3
Human serum 3	91.6	0.706	0.8
Human serum 4	295	1.57	0.5
Human serum 5	516	3.26	0.6

c) PreciControl ClinChem Multi 1

BILT3

Bilirubin Total Gen.3

d) PreciControl ClinChem Multi 2

Method comparison

Total bilirubin values for human serum and plasma samples obtained with the Roche Bilirubin Total Gen.3 reagent on a **cobas c** 503 analyzer (y) were compared with those determined using the corresponding reagent on a **cobas c** 501 analyzer (x).

Sample size (n) = 649

Passing/Bablok¹² $y = 1.000x - 0.0394 \mu\text{mol/L}$ $r = 0.979$

Linear regression

 $y = 1.002x - 0.339 \mu\text{mol/L}$ $r = 1.000$ The sample concentrations were between 2.51 and 622 $\mu\text{mol/L}$.

Total bilirubin values for human serum and plasma samples obtained with the Roche Bilirubin Total Gen.3 reagent on a **cobas c** 303 analyzer (y) were compared with those determined using the corresponding reagent on a **cobas c** 501 analyzer (x).

Sample size (n) = 67

Passing/Bablok¹² $y = 1.010x - 0.247 \mu\text{mol/L}$ $r = 0.966$

Linear regression

 $y = 1.008x - 0.264 \mu\text{mol/L}$ $r = 1.000$ The sample concentrations were between 2.90 and 615 $\mu\text{mol/L}$.

References

- 1 Balistreri WF, Shaw LM. Liver function. In: Tietz NW, ed. Fundamentals of Clinical Chemistry. 3rd ed. Philadelphia: WB Saunders 1987;729-761.
- 2 Wahlefeld AW, Herz G, Bernt E. Modification of the Malloy-Evelyn method for a simple, reliable determination of total bilirubin in serum. Scand J Clin Lab Invest 1972;29 Supplement 126:Abstract 11.12.
- 3 Quality of Diagnostic Samples, Recommendations of the Working Group on Preanalytical Quality of the German Society for Clinical Chemistry and Laboratory Medicine, 3rd completely revised ed. 2010.
- 4 Dumas BT, Kwok-Cheung PP, Perry BW, et al. Candidate Reference Method for Determination of Total Bilirubin in Serum: Development and Validation. Clin Chem 1985;31:1779-1789.
- 5 Glick MR, Ryder KW, Jackson SA. Graphical Comparisons of Interferences in Clinical Chemistry Instrumentation. Clin Chem 1986;32:470-475.
- 6 Breuer J. Report on the Symposium "Drug effects in Clinical Chemistry Methods". Eur J Clin Chem Clin Biochem 1996;34:385-386.
- 7 Sonntag O, Scholer A. Drug interference in clinical chemistry: recommendation of drugs and their concentrations to be used in drug interference studies. Ann Clin Biochem 2001;38:376-385.
- 8 Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. Clin Chem Lab Med 2007;45(9):1240-1243.
- 9 Thomas L, ed. Labor und Diagnose. Indikation und Bewertung von Laborbefunden für die Medizinische Diagnostik, 7th ed.: TH-Books Verlagsgesellschaft 2007:259-273.
- 10 Löhr B, El-Samalouti V, Junge W, et al. Reference Range Study for Various Parameters on Roche Clinical Chemistry Analyzers. Clin Lab 2009;55:465-471.
- 11 Subcommittee on Hyperbilirubinemia. Management of Hyperbilirubinemia in the Newborn Infant 35 or More Weeks of Gestation. Pediatrics 2004;114:297-316.
- 12 Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. J Clin Chem Clin Biochem 1988 Nov;26(11):783-790.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

cobas®

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog. Roche.com for definition of symbols used):

CONTENT

Contents of kit



Volume after reconstitution or mixing

GTIN

Global Trade Item Number

COBAS, COBAS C and PRECICONTROL are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2020, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.Roche.com

