



Product Specification
**Syringe filter, Filtropur S, PES, pore size: 0.2 µm,
for sterile filtration**

Page 1



Product description

Order number	83.1826.001
Product description	Syringe filter, Filtropur S, membrane: PES, filtration surface: 6.2 cm², pore size: 0.2 µm, for sterile filtration, connection: Luer lock, sterile, non-pyrogenic/endotoxin-free, non-cytotoxic, 1 piece(s)/blister

Product characteristics

Example of use	for sterile filtration
Filter	yes
Surface area of filter	6.2 cm²
Pore size	0,2
Connection	Luer lock

Size

Diameter	33 mm
----------	-------

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.





Product Specification
**Syringe filter, Filtropur S, PES, pore size: 0.2 µm,
for sterile filtration**

Page 2

Material & colours

Product material	Methacrylate-butadiene-styrene (MBS)
Colour of product	blue
Membrane material	Polyethersulfone (PES)

Purity & certification

Product category	no medical device IVD
Purity standard	sterile, non-pyrogenic/endotoxin-free, non-cytotoxic
Sterilisation	Gamma irradiation
Pyrogen-free	yes
Non-cytotoxic	yes
Batched	yes

Packaging

Minimum order qty.	50
Type of smallest subpackaging	blister
Piece(s) / inner box	50
Piece (s) / outer case	500
Piece(s) / pallet	12000
Depth of case	375 mm
Width of case	285 mm
Height of case	530 mm
Case volume	0.0566 cbm
Weight of product	0.0042 kg
Weight of case	3.1 kg
EAN of inner box	4038917133913
EAN case	4038917025256

This is the current specification for this product. Sarstedt reserves the right to make changes, in full or in part, at any time without prior notification.

This specification is confidential and the property of Sarstedt. It is neither to be duplicated nor made available to third parties without our prior written consent.

Date of issue:
2023-09-18

This document was prepared by EDP support and is valid without signature.



SARSTEDT