

## Manufacturer Declaration Letter

Date: 15th September, 2023

To whom it may concern,

We SonoScape Medical Corp., located at: 2F,12th Building, Shenzhen Software Park Phase II, Keji Middle 2nd Road, Nanshan District, Shenzhen, 518057, Guangdong, China, who are established and reputable manufacturer of Ultrasound systems model P60 do hereby confirm the following specifications of ultrasound system model P60:

- T10 1. Automated breast measurements that can locate a tumor in a defined area and automatically provide a standardized report on the qualification of a benign or malignant mass.
- T7 2. Automatic measurements of the thyroid gland, which can determine the boundaries of the lesion in the defined area and automatically generate a report on the features.
- 8 3. Automatic measurements of fetal biometrics
- T8 4. Automatic pelvic floor measurement without fixed point, including 2D and 3D
- T9 5. Automatic obstetric measurements with 12 measurement points.
- 3.1 6. LCD monitor is 21.5inch
- T1 7. Maximum frame rate in B mode: 2400 f/s
- 1.1 8. The system is mobile, with castors, and all castors have individually locking brakes
- 10.1 ir 10.2 9. Color doppler mode have power doppler mode and directional power doppler mode
- 6 10. Ability to save and transfer (export) research data (static images and dynamic images (movies) and results of measurements/calculations) in DICOM format
- 1.2 11. There is 5 active probe ports  
Yra 5 aktyvios davikliu jungtys

For any additional details, please don't hesitate to contact us

at: Fax:

E-mail: s