

Téléphone +33 (0) 3 88 63 36 00

Télécopie +33 (0) 3 88 94 12 82

Internet : <http://www.schiller-medical.com>

Email : [info@schiller.fr](mailto:info@schiller.fr)

## **Declaration of conformity with directives 93/42/CEE and 2014/53/UE**

**File N°: CE ARE 0127**

### **Product**

Name: **FRED PA-1**

Function: **Semi-automated or automated external defibrillator for Public access (PAD)**

Classification: **IIb** in accordance with rule 9 below of classification of medical device of Directive 93/42/CEE

"All active therapeutics devices intended to administer or exchange energy are in Class IIa unless their characteristics are such that they may administer or exchange energy to or from the human body in a potentially hazardous way, taking account of the nature, density and site of application of the energy, in which case they are in Class IIb.

All active devices intended to control or monitor the performance of active therapeutic device in Class IIb, or intended directly to influence the performance of such devices are in Class IIb."

Number: **Composition of number 12799sxxxxxx**

**127 : project number of the device**

**12799: FRED PA-1**

**s : device serial number**

**xxxxxx : Unit number**

### **Manufacturer:**

Manufacturer's address: SCHILLER MEDICAL  
4, rue Louis Pasteur  
67160 Wissembourg – France

### **Standards applied**

IEC 60601-1

IEC 60601-1-6

IEC 60601-2-4

IEC 60601-1-2

IEC 60601-1-11

### **Notified Body**

Number: 0459

Name: GMED

Address: 1, rue Gaston Boissier 75724 PARIS CEDEX 15-France



**THE ART OF DIAGNOSTICS**

Téléphone +33 (0) 3 88 63 36 00

Télécopie +33 (0) 3 88 94 12 82

Internet : <http://www.schiller-medical.com>  
Email : [info@schiller.fr](mailto:info@schiller.fr)

## PROOF OF CONFORMITY WITH MAIN REQUIREMENTS WITH DIRECTIVES USED

### Directive 93/42/CEE

Annex II: GMED certificate CE n° 23246 rev 9 issued on April 21<sup>h</sup>, 2021

### Directive 2014/53/UE

Standards applied

- Safety and health (ART 3.1.a): serial IEC 60601
- CEM (ART 3.1b) : IEC 60601-1-2
- Spectrum (ART 3.2) : ETSI EN 302 291-1 V1.1.1  
ETSI EN 302 291-2 V1.1.1  
ETSI EN 301 489 -1 v1.9.2  
ETSI EN 301 489 -3 v1.6.1  
ETSI EN 301 908-1 V11.1.1  
ETSI EN 300 330 v2.1.1  
ETSI EN 301 489-1v2.1.1  
ETSI EN 301 489-3 v2.1.0  
ETSI EN 301 489-24 v1.5.1

## ENGAGEMENT

As responsible for Regulatory Affairs at Schiller MEDICAL, I hereby certify that :

- The product above fulfils the main requirements set out in Directive 93/42/CEE appendix I, chapter 1 to 13
- The product above fulfils the main requirements set out in Directive 2014/53/UE
- CE labelling will be fixed in accordance with directives used.

Wissembourg, May 06, 2021

**SCHILLER MEDICAL**  
  
4 rue Louis Pasteur  
67160 WISSEMBOURG | FRANCE  
Tél: +33 (0) 3 88 63 36 00

Valérie ENGEL  
Regulatory Affairs manager



**THE ART OF DIAGNOSTICS**