



Certificate of Analysis

Product name

KIT COBAS 58/68/8800 MPX 480T IVD

Material No.

09040862190

Lot No.

K20606

Date of Manufacturing

2023-11-22

Expiry Date

2025-09

Parameter	Result	Unit
HIVM-2.5X-Sens/Repeat Reps-R	100	%
HBV-2.5X-Sens/Repeat Reps-R	100	%
HCV-2.5X-Sens/Repeat Reps-R	100	%
HIV1O-2.5X-Sens/Repeat Reps-R	100	%
HIV2-2.5X-Sens/Repeat Reps-R	100	%
HIVM-0.5X-SensReps-R	81	%
HBV-0.5X-SensReps-R	86	%
HCV-0.5X-SensReps-R	76	%
HIV2-0.5X-SensReps-R	71	%
NHP-Specificity-NR	100	%
HIV1O-0.5X-SensReps-R	71	%
Components	1	

Issue date 2024-01-09

Page 1 of 2

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771



Certificate of Analysis

Product name

KIT COBAS 58/68/8800 MPX 480T IVD

Material No.

09040862190

Lot No.

K20606

Date of Manufacturing

2023-11-22

Expiry Date

2025-09

Parameter**Result****Unit**

This lot meets our specification

Disposition by

Craig Cawley

This certificate is a computer printout and has therefore not been signed by hand.

Issue date 2024-01-09

Page 2 of 2

Roche Molecular Systems, Inc.
Quality Management
1080 U.S. Highway 202 South
Branchburg, NJ 08876-3771