

## Disposable Biopsy Forceps Comparison Chart

Measurements for biopsy forceps in mm											
	A	B	C	D	E	F	G	H	I	J	Volume in mm <sup>3</sup> *
<b>Olympus EndoJaw</b>											
EndoJaw 2.0 mm Scope Channel FB-211/221/231/241	1.8	1.5	2.8	3.7	9.0	1.3	0.8	0.20	1.2	5.0	4.0
EndoJaw 2.8 mm Scope Channel FB-210/220/230/240	2.2	1.8	3.0	4.4	10.6	1.2	0.9	0.20	1.0	6.5	6.1
EndoJaw Large 2.8 mm Scope Channel FB-215/225/235/245	2.4	2.0	3.0	5.1	11.3	1.7	1.0	0.2	1.0	8.3	9.5
EndoJaw 3.2 mm Scope Channel FB-214/224/234/244	2.8	2.4	3.4	4.9	11.1	1.0	1.2	0.20	1.0	7.8	11.8
EndoJaw 3.7 mm Scope Channel FB-212/222/232/242	3.3	2.9	4.0	5.5	11.7	1.1	1.5	0.20	1.0	8.7	20.0

  

\*The resected tissue size can deviate depending both on patient conditions and physicians handling.

For more information, please visit:

<http://et-catalogue.olympus.eu/>

As medical knowledge is constantly growing, technical modifications or changes of the product design, product specifications and accessories may be required.