


To whom it may concern,

We, STEPHANIX, who are official manufacturer of medical X-ray equipment, having factory at 10, rue Jean Moulin – ZI du Bayon – 42150 La Ricamarie hereby, confirm that:

1. As we constantly research most optimal equipment solutions, recently the software for the use camera integrated in the vertical column in our XTREME PREMIUM unit has been added to PRIMO S software and the technical documentation regarding this change has not been publicly announced yet. We are currently working on adjustments for the manual of the equipment where we will combine all the small changes and changes regarding the integrated camera use. As we declared before, our equipment has video camera integrated into vertical column which could be controlled from operator workstation and STEPHANIX will thoroughly instruct our partners on the operation of said integrated camera.
2. Quality control tests are performed by the manufacturer, (i.e. STEPHANIX) during the production of the XTREME PREMIUM unit with needed equipment and phantom (PMMA layers (25 cm x 25 cm x 20 cm)). We confirm that the client or final user will not have to perform these tests during installation or use of the equipment. There are no other quality control or calibration tests that must be performed by our partners during installation or the end user during the use of equipment.
3. As it was indicated in XTREME PREMIUM unit manual, the equipment recognizes the collimated area and displays only the region that was indicated before the procedure. This process may as well be called virtual collimation (different manufacturers call it differently). This virtual collimation may be performed automatically via afore mentioned algorithm which virtually indicates collimated area only or, if preferred, manually.

Valerie MARCADE  
Export Manager



La Ricamarie, July 17th, 2024

10, rue Jean Moulin - Z.I du Bayon - 42 150 La Ricamarie - FRANCE  
Tél. + 33 (0)4 77 47 81 60 - Fax + 33 (0)4 77 37 55 19 - e-mail : [contact@stephanix.com](mailto:contact@stephanix.com)

[www.stephanix.com](http://www.stephanix.com)

SAS au capital de 2 800 000 € - RCS Saint-Etienne - SIREN 332 390 566 - Code NAF 4646 Z - N°identification : FR 70 332 390 566

**STEPHANIX**  
ZI du Bayon - 10, rue J. Moulin  
42150 LA RICAMARIE  
Tél. +33(0)477 47 81 60 - Fax +33(0)477 37 55 19  
[contact@stephanix.com](mailto:contact@stephanix.com)  
SA au capital de 2 800 000 e  
SIRET 332 390 566 00054 - NAF 2660Z - FR 70 332 390 566