




PRODUCT TECHNICAL DATA SHEET

PRODUCT NAME T'LIFT®	DESIGNATION Single use retraction system
CODE(S): AW16280	
INDICATIONS The PETERS SURGICAL T'LIFT® is an instrument exclusively designed for laparoscopic surgery. It should only be used by qualified practitioners having gained due experience thereof. T'Lift® is designed to enable the suspension and presentation of anatomical sections during laparoscopic surgery.	
COMPOSITION (No presence of human or biological products – Latex & Phthalate free) T'LIFT®: PolyPropylene Applicator (needle & body): Stainless steel / PolyCarbonate / Silicone Locking clamp: PolyPropylene	
DESCRIPTION & USE T'Lift® is delivered in its packaging with: 2 T'Lift®, 2 locking clamps, 1 applicator (body and needle). T'Lift® loading: T'Lift® introduction from its distal thickest part into the applicator body (steps ① & ②) and push it till its 1 st blue mark. T'Lift® setting: perforation of the abdominal wall, then completely through the anatomic part to be suspended, checking the trajectory and the exit of the needle. Hold the applicator's body and push the T'Lift® rod till its 2 nd blue mark. T'Lift® locking: Finally, extract the applicator and lock the T'Lift® with the clamp onto the abdominal wall. Do not tow the anatomic piece with the T'Lift® only, but support it with an atraumatic grasping forceps.	
<div></div> <div></div>	
PACKAGING / STORAGE CONDITION Packaged in box of 25 units (box size: 32*23*8cm - Weight: 0,6Kg) - Box Code GTIN: 03700393301353 Keep away from humidity and direct sunlight. Avoid any prolonged exposure to low and/or high temperatures.	
STERILIZATION / EXPIRY DATE Sterilized by ethylene oxide. Shelf life: 5 years.	
CONTRAINDICATIONS Use of the PETERS SURGICAL T'LIFT® is strictly forbidden whenever contraindication exists for laparoscopic surgery. On no account may PETERS SURGICAL be held liable for the assessment of any such contraindication for laparoscopic surgery. Not intended for use on new-born and infant.	
CLASSIFICATIONS Medical Device Class: IIa (according to the European Council Directive 93/42/EEC). GMDN code: 61079 CND code: K01020106 Legal Manufacturer: PETERS Surgical – Immeuble Aurélium – 1 Cours de l'Ile Seguin - 92100 Boulogne-Billancourt - France Peters Surgical is certified ISO 13485.	

