

To whom it may concern

Contact: Gudrun Henke
Fon: +49 5661 71-2712
Fax: -
Email: gudrun.henke@bbraun.com
Internet: http://www.bbraun.com
Date: March 08, 2024

CONFIRMATION

We,

B. Braun Melsungen AG
Carl-Braun-Str. 1
34212 Melsungen
Germany

herewith confirm that The surface area for the hydrophobic filter paper of the following medical devices

4269071S	Vasofix Safety® PUR
4269098S	Vasofix Safety® PUR
4269217S	Vasofix Safety® PUR
4269110S	Vasofix Safety® PUR
4269330S	Vasofix Safety® PUR
4269136S	Vasofix Safety® PUR
4269152S	Vasofix Safety® PUR
4269179S	Vasofix Safety® PUR
4269225S	Vasofix Safety® PUR

is 43.98 mm².

For and on behalf of

B. Braun Melsungen AG

i. A.



Marie Fischer

Junior Manager Regulatory Affairs International
International CoE Infusion & Pain Therapy

i. A.



Charlotte Wisker

Administrator Regulatory Affairs
CoE Infusion & Pain Therapy