

Certificat d'analyse

Certificate of analysis

Méthanol SIEMENS

Product code : 00286794

Batch n° : N61215

Manufacturing date : 2024-02-20

Expiry date : 2027-02-20

Charateristics	Specifications	Results
Appearance	Clear, bright, free	Conforms
Platinum-cobalt color	≤5	2
Specific gravity at 20/20 °C	0.791 - 0.793	0.791
Water Karl Fisher	≤1000 mg/kg	150 mg/kg
Non volatile matter	≤10 mg/kg	2.0 mg/kg
Sulfide	< 0.5 mg/kg	0.0 mg/kg
Acidity as acetic acid	≤30 mg/kg	15 mg/kg
Chloride content	< 0.5 mg/kg	0.1 mg/kg
Permanganate times	≥60 min	> 60 min
Purity Methyl alcohol by GC	≥ 99.85 Mass Pct	99.94

L'analyste : Florence BEURIER

The analyst : Florence BEURIER

Date d'édition : 18/04/2024

Edition date : 2024-04-18

Visa :

CELLAVISION

Visa :

RAL Diagnostics



RAL Diagnostics | Site Montesquieu | 33650 Martillac
ral-diagnostics.fr | cellavision.com

T +33 (0) 557 96 04 04 | F +33 (0) 557 96 04 05

Siret 449 261 403 00041 | Code APE 2059 Z

TVA Intracommunautaire FR67449261403