

ABL90 FLEX analyzer

Specifications

Measured parameters

| Type | Parameter | Units | Range of indication |
|---------------|------------------------|-------------------|---------------------|
| pH | pH | pH scale | 6.3–8.0 |
| Blood gas | $p\text{CO}_2$ | mmHg; Torr | 5–250 |
| | | kPa | 0.67–33.3 |
| | $p\text{O}_2$ | mmHg; Torr | 0–800 |
| | | kPa | 0–107 |
| Electrolyte | cK^+ | mmol/L | 0.5–25 |
| | | meq/L | 0.5–25 |
| | cNa^+ | mmol/L | 7–350 |
| | | meq/L | 7–350 |
| | cCa^{2+} | mmol/L | 0.2–9.99 |
| | | meq/L | 0.4–19.98 |
| | | mg/dL | 0.8–40.04 |
| | cCl^- | mmol/L | 7–350 |
| | | meq/L | 7–350 |
| | Metabolite | cGlu | mmol/L |
| mg/dL | | | 0–1081 |
| cLac | | mmol/L | -0.1–31 |
| | | meq/L | -0.1–31 |
| | | mg/dL | -1–279 |
| | | | |
| Oximetry | sO_2 | % | -2–102 |
| | | fraction | -0.02–1.02 |
| | ctHb | g/dL | -0.48–27.7 |
| | | g/L | -4.8–277 |
| | | mmol/L | -0.30–17.2 |
| | FO_2Hb | % | -2–103 |
| | | fraction | -0.02–1.03 |
| | FCOHb | % | -2–103 |
| | | fraction | -0.02–1.03 |
| | fMetHb | % | -2–103 |
| | | fraction | -0.02–1.03 |
| | fHHb | % | -2–102 |
| | | fraction | -0.02–1.02 |
| | fHbF | % | -25–121 |
| | | fraction | -0.25–1.21 |
| | ctBil | $\mu\text{mol/L}$ | -20–1000 |
| | | mg/dL | -1.2–58.5 |
| | | mg/L | -12–585 |

The *Range of indication* for a parameter is the range within which the analyzer is physically capable of measuring. As defined in the 'International vocabulary of basic and general terms in the metrology' (VIM).

Derived parameters

$\text{pH}(T)$
 $\text{pCO}_2(T)$
 $\text{cHCO}_3(\text{P})$
 $\text{cBase}(\text{B})$
 $\text{cBase}(\text{B,ox})$
 $\text{cBase}(\text{Ecf})$
 $\text{cBase}(\text{Ecf,ox})$
 $\text{cHCO}_3(\text{P,st})$
 cH^+
 $\text{cH}^+(T)$
 $\text{ctCO}_2(\text{P})$
 $\text{ctCO}_2(\text{B})$
 $\text{pH}(\text{st})$
 $\text{pO}_2(T)$
 $\text{pO}_2(\text{A})$
 $\text{pO}_2(\text{A},T)$
 p50
 $\text{p50}(T)$
 $\text{p50}(\text{st})$
 $\text{pO}_2(\text{A-a})$
 $\text{pO}_2(\text{A-a},T)$
 $\text{pO}_2(\text{a/A})$
 $\text{pO}_2(\text{a/A},T)$
 $\text{pO}_2(\text{a})/\text{FO}_2(\text{I})$
 $\text{pO}_2(\text{a},T)/\text{FO}_2(\text{I})$
 $\text{cCa}^{2+}(\text{pH}=7.40)$
Anion Gap(K^+)
Anion Gap
 DO_2
Hct
 $\text{pO}_2(\text{x})$
 $\text{pO}_2(\text{x},T)$
 $\text{ctO}_2(\text{B})$
 $\text{ctO}_2(\text{a-v})$
 BO_2
 $\text{ctO}_2(\text{x})$
 fShunt
 $\text{fShunt}(T)$
RI
 $\text{RI}(T)$
 VO_2
 mOsm
 Qx
 Q_t
 $\text{V}(\text{B})$
 sO_2
 FO_2Hb

Measuring system

| | |
|---------------------------------|-----------------------|
| Sample volume (all parameters) | ~ 65 μL |
| Measuring time (all parameters) | 35 sec |
| Cycle time | 60 sec |
| Throughput | 44 samples/hour |
| Average uptime | more than 22hours/day |
| Startup time | down to 1 hour |

Security and QA features

Advanced planning of replacement and QC schedules
Optional automatic QC at startup and after replacements
Customizable QC and calibration schedule.
Continuous sensor status monitoring with corrective actions to get precise results.

Sensor cassette

| | |
|---------------------|-----------------------------------------------------------|
| In-use lifetime | 30 days |
| Shelf life | 4 months |
| Storage temperature | 2 – 10 °C |
| Automatic QC | Yes |
| Thermosat control | Sensor cassette: 37±0.15 °C Oximetry: 37±0.30 °C |

BG / OXI with QC,
BG / LYT / OXI with QC,
BG / LYT / MET / OXI with QC: 50/100/300/600/900 tests

Hardware

Computer specifications

Processor Intel Celeron® M 600 MHz with 512K L2 Cache
1 GB RAM
2 GB SolidState storage
8.4" color TFT-LCD, resolution 800 × 600 SVGA Touch screen
4" thermal-sensitive printer

Software

Software platform

Windows® XP Embedded
Sybase® SQL Anywhere

Data capacity

Patient log: 2000
Activity log: 5000
Calibration adjustment log: 1000
Data secured by password protection
8 different user profiles
Unlimited ID access verification

Printer display options

Auto print (on/off)
Select derived parameters
Select input variables
Reference ranges with results

Sample handling

Inlet

Left/right hand operation
Position for syringe as well as capillary and test tube
Aspiration from capillary without adapter
Aspiration time 5 seconds

Additional information

Dimensions

| | |
|--------|-------|
| Width | 25 cm |
| Height | 45 cm |
| Depth | 29 cm |
| Weight | 11 kg |

Data subject to change without notice.

Radiometer, the Radiometer logo, ABL, AQT, TCM, RADIANCE, AQUIRE, PICO, CLINITUBES and QUALICHECK are trademarks of Radiometer Medical ApS.

© Radiometer Medical ApS, 2700 Brønshøj, Denmark, 2012. All Rights Reserved. 939-053. 201202E.

Solution pack



| | |
|---------------------|-----------|
| In-use lifetime | 30 days |
| Shelf life | 4 months |
| Storage temperature | 2 – 25 °C |

Estimated lifetime of solution packs

| | | | | | | |
|----------------------------------|----|----|----|----|----|----|
| No of tests per day | 5 | 10 | 15 | 20 | 30 | 50 |
| Estimated in-use lifetime (days) | 30 | 30 | 24 | 20 | 15 | 10 |

Interface



Built-in barcode reader for operator & sampler ID
Accepted codes: UPC/EAN, Code 128, Code 39, Code 93, I
2 of 5, Discrete 2 of 5, Codabar and more
Serial interface RS232 with power for external barcode reader
3 USB connections
Optional external keyboard
Optional external mouse
Optional external barcode reader

Communication

HIS/LIS communication

High-level protocols:
ASTM
HL7
POCT1-A
Low-level serial protocols:
ASTM 1381-91, E1394-91
Serial RAW
Low-level network protocols:
TCP/IP

Radiometer IT solution

Interface via Ethernet adapter

Sample mixer

Mixing time 7 seconds
For safePICO samplers

Other

| | |
|-----------------------|---------------------------------|
| Operating environment | 15 – 32 °C |
| Altitude correction | 3000 m above sea level |
| Power | 100 – 240 VAC, 50/60 Hz, 130 VA |