

CERTIFICATE OF ANALYSIS**Product:** SINGLE SET**LOT**Lot number: **XXXXXX**Expiry date: **XXXX-XX****REF****RS450Ca**

	Tests	Requirements according to manufacturer's Specifications*	Test result
1.	Sterility	no living microbes present	Conforms with Specification
2.	Content of bacterial endotoxines	less than 5.56 EU/mL	Conforms with Specification
3.	Anticoagulant CPDA-1 solution		
3.1.	Colour	colourless	Conforms with Specification
3.2.	Solution volume fluctuation	63 mL \pm 10%	Conforms with Specification
3.3.	Reaction	pH = 5.0 \div 6.0	Conforms with Specification
3.4.	Chlorides	not more than 0.0035 %	Conforms with Specification
3.5.	Total citrate expressed as citric acid, anhydrous (C ₆ H ₈ O ₇)	1.916 \div 2.118 g/100 mL	Conforms with Specification
3.6.	Monobasic sodium phosphate (monohydrate; NaH ₂ PO ₄ ·H ₂ O) content	0.211 \div 0.233 g/100 mL	Conforms with Specification
3.7.	Sodium (Na) content	0.621 \div 0.686 g/100 mL	Conforms with Specification
3.8.	Dextrose (anhydrous; C ₆ H ₁₂ O ₆) content	2.755 \div 3.045 g/100 mL	Conforms with Specification
3.9.	Adenine content	0.0247 \div 0.0303 g/100 mL	Conforms with Specification

***SB/01, SB/02, SB/13**

SUMMARY:

Blood collection bags with anticoagulant CPDA-1 conform with specifications

.....
Quality Control Manager

8 F 5 : H

WARRANTY CONDITIONS:

The warranty covers all flaws of the product caused by faulty parts and/or production defects

Warranty covers only exchange of the product for the one free from flaws in the period of validity specified by "Expiry date". Warranty does not cover:

- damages made during transportation and reloading
- damages made by storage not according to the Instruction of Use
- malfunction or damages caused by inappropriate use or use not according to the Instruction of Use
- malfunction or damages caused by act of God