

DR. RER. NAT. HOLGER BRILL

C/O DR. BRILL + PARTNER GMBH
LABOR FÜR HYGIENE UND MIKROBIOLOGIE
PAPENREVE 61, 22453 HAMBURG
TELEFON 0049-40/557631-0
TELEFAX 0049-40/557631-11
EMAIL: INFO@BRILLHYGIENE.COM
INTERNET WWW.BRILLHYGIENE.COM

DR. H. BRILL - C/O DR. BRILL + PARTNER GMBH - PAPENREVE 61 - 22453 HAMBURG

Chemi-Pharm AS
Pollu 132
EST - 10917 Tallinn

Hamburg, 12 April 2006

Expert's report

Efficacy testing of the surgical skin disinfectant *CHEMISEPT G* according to EN 14348: 2005

The surgical skin disinfectant *CHEMISEPT G* was tested and evaluated in accordance with EN 14348: 2005. According to test report no. L 06/001 of Dr. Brill + Partner GmbH the test preparation proved to be tuberculocidal under low organic load.

A sufficient effect within the quantitative suspension test according to EN 14348: 2005 against *Mycobacterium terrae* was reached at the following concentration-time relationship:

100 % in 30 seconds.



Dr. Holger Brill