

# Installation I

1. For MCS+ plasma procedures which use a blow molded bowl (BMB), a centrifuge chuck adapter is required to secure the bowl.

**Note:** The adapter is not a disposable element and should not be discarded after use.



66.4 pd

2. The dual pump manifold adapter and the single pump manifold adapter should be snapped in place.

**Note:** The adapter is not a disposable element and should not be discarded after use.

