
Confirmation

SMP GmbH

Prüfen Validieren Forschen

Hechinger Straße 262
72072 Tübingen

has tested a cleaning process for the master products of the following instrument families

IS4000 (Xi) 8mm Basic Instruments
IS4000 (Xi) Single Site Instruments
IS4000 (Xi) Stapler 45
IS4000 (Xi) 2.0 8mm Endoscope
IS3000 (Si) 5mm Basic Instruments
IS3000 (Si) Single Site Instruments

manufactured by
Intuitive Surgical, Inc.
1266 Kifer Road, Sunnyvale, CA 94086

using the washer-disinfector
Washer-disinfector Getinge Aquadis 56 – REF 56
S/N: WAA097283
with a dedicated Load Carrier for DaVinci Instruments

Da Vinci XI wash cart (Getinge Disinfection AB), REF 6028142401
Da Vinci SI wash cart (Getinge Disinfection AB), REF 6028184101
4-level wash cart (Getinge Disinfection AB), REF 6027591901+ Base Plate C1377

and a dedicated Cleaning Program

and the detergents
Getinge Clean Enzymatic Plus (REF 6002473001)

manufactured by
Getinge

Based on the test results achieved in the cleaning examinations with the project number 16421 SMP GmbH declares that the acceptance criteria for successful cleaning as defined by the sponsor were fulfilled for the instrument families listed above. The investigations were performed according to the protocols and instructions provided by Intuitive Surgical and named in the reports.

Tübingen

Dr. Ludger Schnieder
President/CEO

