



EU Declaration of Conformity

In accordance with EU Regulation 2017/746 of the European Parliament and of the Council of 5 April 2017 on *in vitro* diagnostic medical devices.

Manufacturer: **Ventana Medical Systems Inc.**
1910 E Innovation Park Drive
Tucson, AZ 85755, USA

Single Registration Number (SRN) **US-MF-000016993**
 Manufacturer:

Authorized Representative: **Roche Diagnostics GmbH**
Sandhofer Strasse 116
68305 Mannheim
Germany

Single Registration Number (SRN) **DE-AR-000006262**
 Authorized Representative:

This declaration is issued under the sole responsibility of Ventana Medical Systems Inc.

Product Information

Part Number:	Product Name:	Basic UDI-DI:
05279771001 (alternative P/N: 950-102)	EZ Prep Concentrate (10X)	761333601901AT

Intended Purpose: Ventana Medical Systems' (Ventana) EZ Prep Concentrate (10X) solution (EZ Prep) is used for paraffin removal from tissue samples during immunohistochemistry and in situ hybridization reactions, and to dilute 2X SSC during stringency washes during in situ hybridization reactions carried out on Ventana automated slide stainers.

This product is designed for use on BenchMark Series automated slide stainers.

This product is intended for in vitro diagnostic (IVD) use.

Risk Class: Class A

Common Specifications: Not applicable as no Common Specifications exist for the concerned device.

Conformity of the product with EU Regulation 2017/746 and other applicable EU legislation has been established.

On behalf of Ventana Medical Systems Inc.

Place: Tucson, AZ 85755, USA

Place: Tucson, AZ 85755, USA

Date: 29-Mar-2022

Date: 28-Mar-2022

Jeff Boone

Ben Curson

Jeff Boone

Benjamin Curson

Site Head of Quality Function

Site Head of Regulatory Affairs Function