

**Ecolab GmbH & Co. OHG**  
To Mr. Rudolf Glasmacher  
Department L34  
Reisholzer Wertfstraße 38-42  
  
D-40589 Düsseldorf

Your ref.  
R&D/Health care

Your letter of  
06. 05.2008

Our ref.  
GiS/WHC

Date  
21.05.2008

## Material compatibility Sekusept® Aktiv

Dear Mr. Glasmacher,

we write to confirm, that the product Sekusept® Aktiv, with a new composition, can be used (in regard to concentration, duration and water quality), by reprocessing of certain medical products of the manufacturer KARL STORZ GmbH & CO.KG

The active cleaner with a highly efficient disinfection did not cause any damages to the instruments, made of high-grade steel and videoscopes.

We approved that the manual cleaning-and disinfection solution Sekusept® aktiv, for use with KARL STORZ certain instruments, regard the material compatibility.

Yours sincerely  
Karl Storz GmbH & Co. KG

by order  
Gisela Schnellen  
Reprocessing Specialist  
- Test Laboratory QT -