

Issued by:
RA Peter C Carlsson
Date:
2012-10-29

Reg. No.
GD2310
Revision A

EC Declaration of Conformity

Getinge Disinfection AB hereby declares that medical devices described below comply with the requirements of the European Medical Devices Directive 93/42/EEC (LVFS 2003:11), Machine Directive 2006/42/EG and Low Voltage Directive (LVD) 2006/95/EC.

LVFS 2003:11 is the Swedish implementation of MDD 93/42/EEC.

Compliance with the MDD 93/42/EEC annex II has been audited and validated by the Notified Body Intertek SEMKO AB, Torshamnsgatan 43, Box 1103, SE-164 22 Kista, Sweden.
Code 0413. Certificate No: 41314824.

Washer Disinfectors. These medical devices are part of class IIb.

<u>Name</u>	<u>Model</u>	<u>GMDN code</u>
Getinge WD 14 Tablo	GE14	17671
Getinge WD 15 Claro	GE15	17671
Getinge 46-Series	46-4, 46-5, 46-4T, 46-5T	17671
Getinge 86-Series	S-8666, S-8668	17671
Getinge 88-Series	88-5	17671
Getinge 9100-Series	9120, 9122, 9125, 9128	17671
Getinge CM320-Series	CM320	17671

Washer Disinfectors. This medical device is part of class IIa.

<u>Name</u>	<u>Model</u>	<u>GMDN code</u>
Getinge 9000-Series	9027	17671

Washer Disinfectors for human waste containers. These medical devices are part of class IIa.

<u>Name</u>	<u>Model</u>	<u>GMDN code</u>
Getinge 600-Series	S-607, S-608, Amigo	35318
Getinge 1600-Series	FD 1600, FD 1605, FD 1610, FD 1615, Ninjo	35318
Getinge 1800-Series	FD 1800, FD 1810, Tornado	35318
Getinge 6000	SP-6000, Typhoon	35318

Växjö 2012-10-29
Getinge Disinfection AB



Philip Nokrach
Quality Assurance & Regulatory Affairs Manager