

Certificate No: 41310376  
Date: July 7, 2014  
Handled by: Maria Eklycke  
E-mail: medtechsweden@intertek.com

**Getinge Sterilization AB**

Att: Romana Stefek-Svensson  
Ekebergsvägen 26  
SE-305 75 Getinge  
Sweden

**Purpose** Assessment of the application dated June 27, 2014 to add new products to be included in your certified quality system according to LVFS 2003:11, Annex II (Swedish implementation of MDD 93/42/EEC).

**Products concerned**

Product category	Type/Model designation	Class	Sterile	GMDN code (not mandatory)
Sterilizer	GSS67F	IIB	No	40583
Sterilizer	GSS67H	IIB	No	38671

**Conclusions/Decisions** The products are similar to previously accepted products. The application has been accepted and the products can be added. Application of the CE-mark is permitted when the company's own procedures for CE-marking are fulfilled.

**Follow-up assessments** At the next audit your auditor may follow-up on the implementation of the new products in the Quality system.

**Appeals** Any appeal against this decision will be processed by an appeals panel at Intertek. The appeal shall be submitted to Intertek Semko AB, PO-Box 1103, SE-164 22 Kista, Sweden.

**Intertek Semko AB**  
Notified Body MDD

Mats Premfors  
Certification Authority MDD