

## Certificate of Analysis

**Brand:** medaSEPT latex PF  
**Product:** Powder Free Latex Examination Gloves, Textured  
**AQL:** ≤ 1,5  
**Shape:** Ambidextrous  
**Color:** natural latex  
**Packaging:** 10 x 100 gloves  
**Sizes Assortment:** XS,S, M, L, XL

### 1. WATERTIGHT TEST- EN455:00 Part 1

**Inspection level: G1**

**AQL:1.5**

Size	Sample size (pcs)	Acceptance (pcs)	Defects (pcs)	Result
XS	200	7	2	Pass
S	315	10	3	Pass
M	315	10	5	Pass
L	315	10	4	Pass
XL	200	7	3	Pass

### 2. Dimension EN 455:15 Part 2

Size	XS		S		M		L		XL	
Dimension	Length	Width								
<b>Median</b>	240	76	240	84	240	94	240	104	240	112

Size	XS			S			M			L			XL		
Thickness	Cuff	Palm	Finger												
<b>Median</b>	0.07	0.09	0.12	0.07	0.10	0.12	0.07	0.10	0.12	0.07	0.10	0.12	0.07	0.09	0.11

### 3. Force at break & Elongation at break before accelerated ageing and after accelerated ageing

Size	Median Value					
	Before Accelerated Aging (N)			After Aging (N)		
	Force at break (Min 6 N)			Force at break (Min 6 N)		
XS	7.3			6.7		
S	7.0			6.5		
M	7.2			6.6		
L	8.1			6.9		
XL	8.0			6.8		

### 4. Protein content test- EN 455:15 PART 3

**Requirements: Max 50 µg/g**

Rozmiar	XS	S	M	L	XL
<b>protein content µg/g</b>	49.67	49.23	49.80	49.21	49.75
<b>Result</b>	Pass				

**5. Powder residue**

**Requirements: Max 2mg/glove**

Size	Powder content mg/glove	Result
XS	1.0	Pass
S	1.1	
M	1.3	
L	1.2	
XL	1.4	

6. According to EN 455 part 4 above mentioned parameters (after ageing) are guaranteed during whole period of glove shelf life

*Dyrektor ds. Zakupów / Produktów*  
*Jowita Boryta-Szymańska*  
Purchasing & Product Director

**MEDASEPT**  
Spółka z ograniczoną odpowiedzialnością  
61-362 Poznań, Forteczna 19  
NIP 7822557229 REGON 302347543