



04955170001V4.0

RPR Calibrator Set

SEKISUI

REF 04955170 190

5 x 1 mL Calibrator

English

System information

For use on Roche/Hitachi MODULAR P analyzers the calibrator codes are 931-935.

Please note: For technical reasons it is necessary to specify dummy units (mg/dL) on Roche/Hitachi analyzers. The values can be converted to the correct units via the host computer.

Intended use

RPR Calibrator Set is a set of 5 calibrators for use in the calibration of the Sekisui "Mediace RPR" immunoturbidimetric assay (Cat. No. 04955153).

Summary

The RPR Calibrator Set contains 5 liquid, ready-for-use calibrators based on human serum and bovine serum albumin (BSA).

The concentrations of the calibrator components have been adjusted to ensure optimal calibration of the Sekisui "Mediace RPR" assay on clinical chemistry analyzers.

Reagents – working solutions

Reactive components:

Human serum and bovine serum albumin (BSA) with chemical additives and material of biological origin as specified.

The origin of the biological additive is as follows:

Analyte	Origin
Syphilitic anti-lipid antibodies	human

Non-reactive components:

BSA, stabilizers and preservatives.

The antibody levels are lot-specific. The exact calibrator values are given on the label of each vial and in the enclosed value sheets

The values are also encoded in the enclosed calibrator barcode sheets for Roche/Hitachi MODULAR P analyzers.

Calibrator values¹

The calibrator values were determined using the Sekisui Mediacce RPR immunoturbidimetric method.

Traceability information is given in the relevant Method Sheets for the system reagents.

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV.

The testing methods applied were FDA-approved or cleared in compliance with the European Directive 98/79/EC, Annex II, List A.

However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.^{2,3}

Handling

The product is ready-for-use. Mix carefully before use. Avoid the formation of foam.

The enclosed barcoded labels are intended exclusively for Roche/Hitachi MODULAR automated analyzers to identify the calibrator. Attach the barcoded labels to the tubes carrying the sample cups containing the calibrator material.

Storage and stability¹

Store at 2-8 °C.

Criterion for the stability data stated:

Recovery within ± 15 % of initial value.

Stability:

Unopened: up to the stated expiration date at 2-8 °C

After opening: 1 day at 15-25 °C
4 weeks at 2-8 °C, provided that dispensing of the calibrator occurs without contamination

Do not freeze. Store calibrator tightly capped when not in use.

Materials provided

- See "Reagents – working solutions" section
- Barcoded labels

Materials required (but not provided)

- System reagents and clinical chemistry analyzers
- General laboratory equipment

Assay

Use RPR Calibrator Set as specified in the relevant Method Sheet for the system reagents.

References

- Data on file at Sekisui.
- Occupational Safety and Health Standards: bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

	Contents of kit
	Calibrator
	Volume after reconstitution or mixing

MODULAR is a trademark of Roche. MEDIACE is a trademark of Sekisui.

All other product names and trademarks are the property of their respective owners.

Significant additions or changes are indicated by a change bar in the margin.

© 2014, Roche Diagnostics



Sekisui Medical Co. Ltd.
13-5, Nihombashi 3-chome, Chuo-ku
Tokyo, 103-0027, Japan



Medical Device Safety Service GmbH (MDSS)
Schiffgraben 41
D-30175 Hannover, Germany

European Distributor:
Roche Diagnostics GmbH
Sandhofer Strasse 116
D-68305 Mannheim, Germany

