

TDM Control Set

REF 04521536 190

2 x 5 mL Control Level I

2 x 5 mL Control Level II

2 x 5 mL Control Level III

English

System information

For use on Roche/Hitachi MODULAR and **cobas c** analyzers the control codes are:

Level I	TDMC1	Code 310
Level II	TDMC2	Code 311
Level III	TDMC3	Code 312

For use on COBAS INTEGRA analyzers the system IDs are:

Level I	TDMC1	system ID 07 6900 2
Level II	TDMC2	system ID 07 6901 0
Level III	TDMC3	system ID 07 6902 9

Intended use

The TDM Control Set is for use in quality control by monitoring accuracy and precision for the quantitative method as specified in the value sheets.

Summary

The TDM Control Set consists of 3 liquid ready-for-use controls based on human serum.

The adjusted concentrations of the control components are usually in the normal range or at the normal/pathological threshold.

Some methods specified in the relevant value sheet may not be available in all countries.

Reagents – working solutions

Reactive components:

Human serum with added therapeutic drugs

Non-reactive components:

Preservative and stabilizer.

The concentrations of the control components are lot-specific. The exact target values are given in the electronically available or enclosed value sheets.

The values are also encoded in the enclosed control barcode sheets for Roche/Hitachi MODULAR and COBAS INTEGRA analyzers.

For the **cobas c** analyzers (except for the **cobas c 111** analyzer) the values are encoded in electronic files sent via the **cobas** link to the analyzers.

Target values and ranges

The target values were determined using the method stated in the electronically available or enclosed value sheets. Determinations for Roche methods were performed under strictly standardized conditions on Roche analyzers using Roche system reagents. The target value specified is the mean of all values obtained. The corresponding control range is calculated as the target value ± 3 standard deviations (the standard deviation being the value obtained from several target value determinations). Results should be within the defined ranges. Each laboratory should establish corrective measures to be taken if values fall outside the range.

A clinically insignificant difference may be seen between the value(s) listed on the value sheet and the value(s) obtained from the instrument readable data. This is caused by:

- the rounding of value(s) during conversion from the unit in the instrument readable data to the unit that is being used.
- the calculation of the ranges by the analyzer using the percentage values for the ranges encoded in the barcodes.

The traceability of the target value is given in the respective Method Sheets for the system reagents to be used in combination with the recommended calibrator.

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV.

The testing methods applied were FDA-approved or cleared in compliance with the European Directive 98/79/EC, Annex II, List A.

However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.^{1,2}

For US users only:

Warning: This reagent contains carbamazepine, phenobarbital, phenytoin, primidone and valproic acid, substances known to the State of California to cause cancer or reproductive harm.

Handling

The product is ready-for-use. Prior to use swirl bottle carefully to obtain a homogenous solution. Record the date that the bottle was opened on the relevant bottle label.

The enclosed barcoded labels are intended exclusively for the Roche/Hitachi MODULAR analyzers and **cobas c** systems to identify the control. Attach the barcoded labels to the tubes carrying the sample cups containing the control material.

Storage and stability

Store at 2-8 °C.

Stability:

Unopened: up to the stated expiration date at 2-8 °C

After opening: 12 months, or until the printed expiration date, whichever comes first, at 2-8 °C

Store controls tightly capped when not in use.

Materials provided

- See "Reagents – working solutions" section
- Barcoded labels

Materials required (but not provided)

- Roche system reagents and clinical chemistry analyzers
- General laboratory equipment

Assay

Dispense the required volume into a sample cup and analyze in the same way as patient samples.

The controls should be run daily in parallel with patient samples and after every calibration. Control intervals must be adapted to individual laboratory's requirements.

Follow the applicable government regulations and local guidelines for quality control.

References

- Occupational Safety and Health Standards: bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

CONTENT

Contents of kit

TDM Control Set

cobas[®]

Volume after reconstitution or mixing

FOR US CUSTOMERS ONLY: LIMITED WARRANTY

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS C, COBAS INTEGRA and MODULAR are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Significant additions or changes are indicated by a change bar in the margin.

© 2013, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com



Distribution in USA by:

Roche Diagnostics, Indianapolis, IN

US Customer Technical Support 1-800-428-2336