

C E R T I F I C A T E
of Conformity



Registration No.: AK 50440940 0001

Report No.: 50268676 001

Holder: **Azure Biosystems, Inc.**
6747 Sierra Court, Suite A-B,
Dublin, California 94568
USA

Product: Measuring Instrument
(Bioanalytical Imaging System)

Identification: Type Designation: Azure 150 Azure 200 Azure 300Q
Azure 280 Azure 300 Azure 400
Azure 500 Azure 600 Azure 500Q
Serial No.: Engineering sample
Remark: Refer to test report 50268676 001 for details.

Tested acc. to: EN 301489-1 V 2.2.0:2017
EN 301489-17 V 3.2.0:2017
EN 61326-1:2013
EN 300328 V 2.1.1:2016
EN 62311:2008
EN 61010-1:2010

The certificate of conformity refers to the above mentioned product. This is to certify that the specimen is in conformity with the assessment requirement mentioned above. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Date 08.07.2019



Certification Body

A handwritten signature in blue ink, appearing to read "Mai Miao".
Mai Miao

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg