

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>CO(A-L A2C) Alias: CO A-L A2C</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>CI(A-L A2C) Alias: CI A-L A2C</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>EF(MOD A2C) Alias: LVEF MOD A2C</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>SV(MOD A2C) Alias: SV MOD A2C</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT,</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		<p>“Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>SI(MOD A2C) Alias: SI MOD A2C</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>CO(MOD A2C) Alias: CO MOD A2C</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>CI(MOD A2C) Alias: CI MOD A2C</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVAd(LAX) Alias: LVAd LAX</p>	<p>(G-0375, SRT, “Left Ventricular Diastolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”)</p>
<p>LVLd(LAX) Alias: LVLd LAX</p>	<p>(18077-8, LN, “Left Ventricle diastolic major axis”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32011, SRT, “End Diastole”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”)</p>
<p>LVLs(LAX) Alias: LVLd LAX</p>	<p>(18076-0, LN, “Left Ventricle systolic major axis”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (109070, DCM, “End Systole”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”)</p>
<p>LVEDV(A-L LAX) Alias: LVEDV A-L LAX</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>LVEDV(MOD LAX) Alias: LVEDV MOD LAX</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>LVAAs(LAX) Alias: LVAAs LAX</p>	<p>(G-0374, SRT, “Left Ventricular Systolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVESV(A-L LAX) Alias: LVESV A-L LAX</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>LVESV(MOD LAX) Alias: LVESV MOD LAX</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>EF(A-L LAX) Alias: EF A-L LAX</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>SV(A-L LAX) Alias: SV A-L LAX</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)</p>
<p>SI(A-L LAX) Alias: SI A-L LAX</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		<p>Ventricle") (111031, DCM, "Image View") = (G-0395, SRT, "Apical long axis") (G-C036, SRT, "Measurement Method") = (125205, DCM, "Area-Length Single Plane")</p>
<p>CO(A-L LAX) Alias: CO A-L LAX</p>	<p>(F-32100, SRT, "Cardiac Output")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T-32600, SRT, "Left Ventricle") (111031, DCM, "Image View") = (G-0395, SRT, "Apical long axis") (G-C036, SRT, "Measurement Method") = (125205, DCM, "Area-Length Single Plane")</p>
<p>CI(A-L LAX) Alias: CI A-L LAX</p>	<p>(F-32110, SRT, "Cardiac Index")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T-32600, SRT, "Left Ventricle") (111031, DCM, "Image View") = (G-0395, SRT, "Apical long axis") (G-C036, SRT, "Measurement Method") = (125205, DCM, "Area-Length Single Plane")</p>
<p>EF(MOD LAX) Alias: LVEF MOD LAX</p>	<p>(18043-0, LN, "Left Ventricular Ejection Fraction")</p>	<p>(111031, DCM, "Image View") = (G-0395, SRT, "Apical long axis") (G-C036, SRT, "Measurement Method") = (125208,</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		DCM, "Method of Disks, Single Plane")
SV(MOD LAX) Alias: SV MOD LAX	(F-32120, SRT, "Stroke Volume")	(G-C0E3, SRT, "Finding Site") = (T-32600, SRT, "Left Ventricle") (111031, DCM, "Image View") = (G-0395, SRT, "Apical long axis") (G-C036, SRT, "Measurement Method") = (125208, DCM, "Method of Disks, Single Plane")
SI(MOD LAX) Alias: SI MOD LAX	(F-00078, SRT, "Stroke Index")	(G-C0E3, SRT, "Finding Site") = (T-32600, SRT, "Left Ventricle") (111031, DCM, "Image View") = (G-0395, SRT, "Apical long axis") (G-C036, SRT, "Measurement Method") = (125208, DCM, "Method of Disks, Single Plane")
CO(MOD LAX) Alias: CO MOD LAX	(F-32100, SRT, "Cardiac Output")	(G-C0E3, SRT, "Finding Site") = (T-32600, SRT, "Left Ventricle") (111031, DCM, "Image View") = (G-0395, SRT, "Apical long axis") (G-C036, SRT, "Measurement Method") = (125208, DCM, "Method of Disks, Single Plane")

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>CI(MOD LAX) Alias: CI MOD LAX</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32600, SRT, “Left Ventricle”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (125208, DCM, “Method of Disks, Single Plane”)</p>
<p>LVEDV(MOD BP) Alias: LVEDV MOD BP</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>LVEDV Index(MOD BP) Alias: LVEDV Index MOD BP</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>LVESV(MOD BP) Alias: LVESV MOD BP</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>LVESV Index(MOD BP) Alias: LVESV Index MOD BP</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>EF(Biplane) Alias: EF Biplane</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>SV(Biplane) Alias: SV Biplane</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>SI(Biplane) Alias: SI Biplane</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>CO(Biplane) Alias: CO Biplane</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>CI(Biplane) Alias: CI Biplane</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125207, DCM, “Method of Disks, Biplane”)</p>
<p>LVLd(apical) Alias: LVLd apical</p>	<p>(18077-8, LN, “Left Ventricle diastolic major axis”)</p>	
<p>LVLd(apical epi) Alias : LVLd(apical epi)</p>	<p>(18077-8, LN, "Left Ventricle diastolic major axis")</p>	<p>(G-C0E3, SRT, "Finding Site") = (GEU, GEU-106-0157, "Epicardium")</p>
<p>LVLs(apical) Alias: LVLs apical</p>	<p>(18076-0, LN, “Left Ventricle systolic major axis”)</p>	
<p>LVAd(sax MV) Alias: LVAd sax MV</p>	<p>(G-0375, SRT, “Left Ventricular Diastolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-039A, SRT, “Parasternal short axis at the Mitral Valve level”)</p>
<p>LVA(sax MV) Alias: LVA(sax MV)</p>	<p>(G-0374, SRT, “Left Ventricular Systolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-039A, SRT, “Parasternal short axis at the Mitral Valve level”)</p>
<p>LVAd(sax PM) Alias: LVAd sax PM</p>	<p>(G-0375, SRT, “Left Ventricular Diastolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-039B, SRT, “Parasternal short axis at the Papillary Muscle level”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVAs(sax PM) Alias: LVAs sax PM</p>	<p>(G-0374, SRT, “Left Ventricular Systolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-039B, SRT, “Parasternal short axis at the Papillary Muscle level”)</p>
<p>LVAd(sax epi) Alias: LVAd sax EPI</p>	<p>(G-0379, SRT, “Left Ventricle Epicardial Diastolic Area, psax pap view”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode</p>
<p>LVAs(sax epi) Alias: LVAs sax EPI</p>	<p>(59093-5, LN, “Epicardial Area”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-03A2, SRT, “2D mode”) or (R-409E2, SRT, “Doppler Color Flow”) depending on scan mode (111031, DCM, “Image View”) = (G-039B, SRT, “Parasternal short axis at the Papillary Muscle level”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”)</p>
<p>LVAd(sax) Alias: LVAd sax</p>	<p>(G-0375, SRT, “Left Ventricular Diastolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-0397, SRT, “Parasternal short axis”)</p>
<p>LVAs(sax) Alias: LVAs sax</p>	<p>(G-0374, SRT, “Left Ventricular Systolic Area”)</p>	<p>(111031, DCM, “Image View”) = (G-0397, SRT, “Parasternal short axis”)</p>
<p>EDV(mod sim) Alias: EDV mod sim</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>ESV(mod sim) Alias: ESV mod sim</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	
<p>EF(mod sim) Alias: EF mod sim</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	
<p>SV(mod sim) Alias: SV mod sim</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)</p>
<p>SI(mod sim) Alias: SI mod sim</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)</p>
<p>CO(mod sim) Alias: CO mod sim</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)</p>
<p>CI(mod sim) Alias: CI mod sim</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)</p>
<p>EDV(bullet) Alias: EDV bullet</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	
<p>EDV(bullet epi) Alias : EDV(bullet epi)</p>	<p>(18026-5, LN, "Left Ventricular End Diastolic Volume")</p>	<p>(G-C0E3, SRT, "Finding Site") = (GEU, GEU-106-0157, "Epicardium") (G-C036, SRT, "Measurement Method") = (DCM, 125228, "Bullet Method") (R-4089A, SRT, "Cardiac Cycle Point") = (SRT, F-32011, "End Diastole")</p>
<p>ESV(bullet) Alias: ESV bullet</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>EF(bullet) Alias: LVEF Bullet</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	
<p>SV(bullet) Alias: SV bullet</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, «Finding Site») = (T-32600, SRT, «Left Ventricle»)</p>
<p>SI(bullet) Alias: SI bullet</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)</p>
<p>CO(bullet) Alias: CO bullet</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)</p>
<p>CI(bullet) Alias: CI bullet</p>	<p>(F-32110, SRT, “Cardiac Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”)</p>
<p>EDV(bp el) Alias: EDV bp el</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125211, DCM, “Biplane Ellipse”)</p>
<p>ESV(bp el) Alias: ESV bp el</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125211, DCM, “Biplane Ellipse”)</p>
<p>EF(bp el) Alias: LVEF BP-EL</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (125211, DCM, “Biplane Ellipse”)</p>
<p>SV(bp el) Alias: SV bp el</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-C036, SRT, “Measurement Method”) = (125211,</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		DCM, “Biplane Ellipse”)
SI(bp el) Alias: SI bp el	(F-00078, SRT, “Stroke Index”)	(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-C036, SRT, “Measurement Method”) = (125211, DCM, “Biplane Ellipse”)
CO(bp el) Alias: CO bp el	(F-32100, SRT, “Cardiac Output”)	(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-C036, SRT, “Measurement Method”) = (125211, DCM, “Biplane Ellipse”)
CI(bp el) Alias: CI bp el	(F-32110, SRT, “Cardiac Index”)	(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-C036, SRT, “Measurement Method”) = (125211, DCM, “Biplane Ellipse”)
LVd Mass(A-L) Alias: LVd Mass A-L	(18087-7, LN, “Left Ventricle Mass”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32010, SRT, “Diastole”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
LVs Mass(A-L) Alias: LVs Mass A-L	(18087-7, LN, “Left Ventricle Mass”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT,

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		“Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
LVd Mass Index(A-L) Alias: LVd Mass I A-L	(GEU-106-0028, 99GEMS, “Left Ventricle Mass Index”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32010, SRT, “Diastole”) (G-C036, SRT, “Measurement Method”) = (125205, DCM, “Area-Length Single Plane”)
MM/IVSd Alias: IVSd	(18154-5, LN, “Interventricular Septum Diastolic Thickness”)	(G-0373, SRT, «Image Mode») = (G-0394, SRT, «M mode»)
MM/IVSs Alias: IVSs	(18158-6, LN, “Interventricular Septum Systolic Thickness”)	(G-0373, SRT, «Image Mode») = (G-0394, SRT, «M mode»)
MM/LVIDd Alias: LVIDd	(29436-3, LN, “Left Ventricle Internal End Diastolic Dimension”)	(G-0373, SRT, «Image Mode») = (G-0394, SRT, «M mode»)
MM/LVIDs Alias: LVIDs	(29438-9, LN, “Left Ventricle Internal Systolic Dimension”)	(G-0373, SRT, «Image Mode») = (G-0394, SRT, «M mode»)
MM/LVPWd Alias: LVPWd	(18152-9, LN, “Left Ventricle Posterior Wall Diastolic Thickness”)	(G-0373, SRT, «Image Mode») = (G-0394, SRT, «M mode»)
MM/LVPWs Alias: LVPWs	(18156-0, LN, “Left Ventricle Posterior Wall Systolic Thickness”)	(G-0373, SRT, «Image Mode») = (G-0394, SRT, «M mode»)
MM/EDV(Teich) Alias: EDV(Teich)	(18026-5, LN, “Left Ventricular End Diastolic Volume”)	(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125209, DCM, “Teichholz”)

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>MM/ESV(Teich) Alias: ESV(Teich)</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125209, DCM, “Teichholz”)</p>
<p>MM/EF(Teich) Alias: EF(Teich)</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125209, DCM, “Teichholz”)</p>
<p>MM/SV(Teich) Alias: SV(Teich)</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125209, DCM, “Teichholz”)</p>
<p>MM/SI(Teich) Alias: SI(Teich)</p>	<p>(F-00078, SRT, “Stroke Index”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125209, DCM, “Teichholz”)</p>
<p>MM/CO(Teich) Alias: CO(Teich)</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		(G-C036, SRT, “Measurement Method”) = (125209, DCM, “Teichholz”)
MM/CI(Teich) Alias: CI(Teich)	(F-32110, SRT, “Cardiac Index”)	(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125209, DCM, “Teichholz”)
MM/EDV(Cube) Alias: EDV(Cube)	(18026-5, LN, “Left Ventricular End Diastolic Volume”)	(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)
MM/ESV(Cube) Alias: ESV(Cube)	(18148-7, LN, “Left Ventricular End Systolic Volume”)	(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)
MM/EF(Cube) Alias: EF(Cube)	(18043-0, LN, “Left Ventricular Ejection Fraction”)	(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)
MM/SV(Cube) Alias: SV(Cube)	(F-32120, SRT, “Stroke Volume”)	(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”)

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		(G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)
MM/SI(Cube) Alias: SI(Cube)	(F-00078, SRT, “Stroke Index”)	(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)
MM/CO(Cube) Alias: CO(Cube)	(F-32100, SRT, “Cardiac Output”)	(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)
MM/CI(Cube) Alias: CI(Cube)	(F-32110, SRT, “Cardiac Index”)	(G-C0E3, SRT, “Finding Site”) = (T-32600, SRT, “Left Ventricle”) (G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”) (G-C036, SRT, “Measurement Method”) = (125206, DCM, “Cube Method”)
MM/%FS Alias: %FS	(18051-3, LN, “Left Ventricular Fractional Shortening”)	(G-0373, SRT, «Image Mode») = (G-0394, SRT, «M mode»)
MM/IVSd/LVPWd Alias: IVSd/LVPWd	(18155-2, LN, “Interventricular Septum to Posterior	(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”)

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

	Wall Thickness Ratio")	(R-4089A, SRT, "Cardiac Cycle Point") = (F-32010, SRT, "Diastole")
MM/%LVPW Thck Alias: %LVPW Thck	(18053-9, LN, "Left Ventricle Posterior Wall % Thickening")	(G-0373, SRT, «Image Mode») = (G-0394, SRT, «M mode») = (F-32010, SRT, «Diastole»)
MM/LVd Mass Alias: LVd Mass	(18087-7, LN, "Left Ventricle Mass")	(G-0373, SRT, "Image Mode") = (G-0394, SRT, "M mode") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32010, SRT, "Diastole")
MM/LVs Mass Alias: LVs Mass	(18087-7, LN, "Left Ventricle Mass")	(G-0373, SRT, "Image Mode") = (G-0394, SRT, "M mode") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole")
MM/LVd Mass/ASE Alias: LVd Mass (ASE)	(18087-7, LN, "Left Ventricle Mass")	(G-0373, SRT, "Image Mode") = (G-0394, SRT, "M mode") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32010, SRT, "Diastole") (G-C036, SRT, "Measurement Method") = (125221, DCM, "Left Ventricle Mass by M-mode")
MM/LVs Mass/ASE Alias: LVs Mass (ASE)	(18087-7, LN, "Left Ventricle Mass")	(G-0373, SRT, "Image Mode") = (G-0394, SRT, "M mode") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole") (G-C036, SRT, "Measurement Method") = (125221, DCM, "Left Ventricle Mass by M-mode")

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		DCM, “Left Ventricle Mass by M-mode”)
MM/LVPEP Alias: LVPEP	(18068-7, LN, “Left Ventricle Pre Ejection Period”)	(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”)
MM/LVPEP/ET Ratio Alias: LVPEP ET Ratio	(59088-5, LN, “Left Ventricular Pre-ejection time/Ejection time”)	(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”)
MM/LVET Alias: LVET	(20222-6, LN, “Ejection Time”)	(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”)
MM/Vcf Mean Alias: Vcf Mean	(59117-2, LN, “Mean Velocity of Circumferential Fibe Shortening (Mean VcFv”)	(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”)
MM/Vcf Mean (corr) Alias: Vcf Mean (corr)	(59118-0, LN, “HR-Corrected Mean Velocity of Circumferential Fiber Shortening”)	(G-0373, SRT, “Image Mode”) = (G-0394, SRT, “M mode”)
MM/HeartRate Alias: HR	(8867-4, LN, “Heart rate”)	(G-0373, SRT, «Image Mode») = (G-0394, SRT, «M mode»)
SD/HeartRate Alias: HR	(8867-4, LN, “Heart rate”)	(G-0373, SRT, «Image Mode») = (R-409E4, SRT, «Doppler Pulsed»)
SD/HeartRate/Calc Alias: HR	(8867-4, LN, “Heart rate”)	(G-0373, SRT, «Image Mode») = (R-409E4, SRT, «Doppler Pulsed»)
IVCT Alias: IVCT	(G-037E, SRT, “Left Ventricular Isovolumic Contraction Time”)	
IVRT Alias: IVRT	(18071-1, LN, “Left Ventricular Isovolumic Relaxation Time”)	

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>MP/LVOT Diam Alias: LVOT Diam</p>	<p>(G-038F, SRT, "Cardiovascular Orifice Diameter")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T- 32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>MP/LVOT VTI Alias: LVOT VTI</p>	<p>(20354-7, LN, "Velocity Time Integral")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T- 32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>LVOT Vmax Alias: LVOT Vmax</p>	<p>(11726-7, LN, "Peak Velocity")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T- 32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>LVOT Vmax P Alias: LVOT Vmax</p>	<p>(11726-7, LN, "Peak Velocity")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T- 32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>LVOT Vmax VM Alias : LVOT Vmax VM</p>	<p>(11726-7, LN, "Peak Velocity")</p>	<p>(G-C0E3, SRT, "Finding Site") = (SRT, T-32650, "Left Ventricle Outflow Tract") (18139-6, LN, "Stage") = (SRT, R-40928, "Valsalva maneuver")</p>
<p>LVOT maxPG Alias: LVOT maxPG</p>	<p>(20247-3, LN, "Peak Gradient")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T- 32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>LVOT maxPG VM Alias : LVOT maxPG VM</p>	<p>(20247-3, LN, "Peak Gradient")</p>	<p>(G-C0E3, SRT, "Finding Site") = (SRT, T-32650, "Left Ventricle Outflow Tract") (18139-6, LN, "Stage") = (SRT, R-40928, "Valsalva maneuver")</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVOT Vmean Alias: LVOT Vmean</p>	<p>(20352-1, LN, "Mean Velocity")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T-32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>LVOT Vmean VM Alias : LVOT Vmean VM</p>	<p>(20352-1, LN, "Mean Velocity")</p>	<p>(G-C0E3, SRT, "Finding Site") = (SRT, T-32650, "Left Ventricle Outflow Tract") (18139-6, LN, "Stage") = (SRT, R-40928, "Valsalva maneuver")</p>
<p>LVOT meanPG Alias: LVOT meanPG</p>	<p>(20256-4, LN, "Mean Gradient")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T-32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>LVOT meanPG VM Alias : LVOT meanPG VM</p>	<p>(20256-4, LN, "Mean Gradient")</p>	<p>(G-C0E3, SRT, "Finding Site") = (SRT, T-32650, "Left Ventricle Outflow Tract") (18139-6, LN, "Stage") = (SRT, R-40928, "Valsalva maneuver")</p>
<p>LVOT VTI Alias: LVOT VTI</p>	<p>(20354-7, LN, "Velocity Time Integral")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T-32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>LVOT HR Alias: HR</p>	<p>(8867-4, LN, "Heart rate")</p>	
<p>LVOT SV Alias: LVSV Dopp</p>	<p>(F-32120, SRT, "Stroke Volume")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T-32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>LVOT SI Alias: LVSI Dopp</p>	<p>(F-00078, SRT, "Stroke Index")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T-32650, SRT, "Left</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		Ventricle Outflow Tract")
<p>LVOT CO Alias: LVCO Dopp</p>	<p>(F-32100, SRT, "Cardiac Output")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T-32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>LVOT CI Alias: LVCI Dopp</p>	<p>(F-32110, SRT, "Cardiac Index")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T-32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>LVOT Env. Ti Alias: LVOT Env. Ti</p>	<p>(GEU-106-0081, 99GEMS, "Tie duration of the VTI trace on LVOT")</p>	<p>(G-0373, SRT, "Image Mode") = (R-409E4, SRT, "Doppler Pulsed")</p>
<p>LIMP Alias: LIMP</p>	<p>(G-037F, SRT, "Left Ventricular Index of Myocardial Performance")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T-32600, SRT, "Left Ventricle") (G-0373, SRT, "Image Mode") = (R-409E4, SRT, "Doppler Pulsed")</p>
<p>LVPEP Alias: LVPEP</p>	<p>(79989-0, LN, "Left Ventricle Pre Ejection Period by US doppler")</p>	<p>(G-0373, SRT, "Image Mode") = (R-409E4, SRT, "Doppler Pulsed") (G-C0E3, SRT, "Finding Site") = (T-32650, SRT, "Left Ventricle Outflow Tract")</p>
<p>LVET Alias: Ejection Time</p>	<p>(20222-6, LN, "Ejection Time")</p>	<p>(G-0373, SRT, "Image Mode") = (R-409E4, SRT, "Doppler Pulsed") (G-C0E3, SRT, "Finding Site") = (T-32650, SRT, "Left Ventricle Outflow Tract")</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVPEP/ET Ratio Alias: LVPEP/ET Ratio</p>	<p>(59088-5, LN, “Ventricular Pre ejection time/ Ejection time by US”)</p>	<p>(G-0373, SRT, “Image Mode”) = (R-409E4, SRT, “Doppler Pulsed”) (G-C0E3, SRT, “Finding Site”) = (T- 32650, SRT, “Left Ventricle Outflow Tract”)</p>
<p>AP/LVOT Diam Alias: LVOT Diam</p>	<p>(G-038F, SRT, “Cardiovascular Orifice Diameter”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32650, SRT, “Left Ventricle Outflow Tract”)</p>
<p>AP/LVOT VTI Alias: LVOT VTI</p>	<p>(20354-7, LN, “Velocity Time Integral”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32650, SRT, “Left Ventricle Outflow Tract”)</p>
<p>LVAd(avg) Alias: LVAd(avg)</p>	<p>(G-0375, SRT, “Left Ventricular Diastolic Area”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (GEU-106- 0017, 99GEMS, “Triplane”)</p>
<p>LVAs(avg) Alias: LVAs(avg)</p>	<p>(G-0374, SRT, “Left Ventricular Systolic Area”)</p>	<p>(G-C036, SRT, “Measurement Method”) = (GEU-106- 0017, 99GEMS, “Triplane”)</p>
<p>ECG/HeartRate Alias: HR</p>	<p>(8867-4, LN, “Heart rate”)</p>	
<p>LVEDV(4D) Alias: EDV</p>	<p>(18026-5, LN, “Left Ventricular End Diastolic Volume”)</p>	<p>(G-0373, SRT, “Image Mode”) = (125231, DCM, “3D mode”) (G-C036, SRT, “Measurement Method”) = (GEU-106- 0023, 99GEMS, “4D Auto Left Ventricle Quantification”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>LVESV(4D) Alias: ESV</p>	<p>(18148-7, LN, “Left Ventricular End Systolic Volume”)</p>	<p>(G-0373, SRT, “Image Mode”) = (125231, DCM, “3D mode”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0023, 99GEMS, “4D Auto Left Ventricle Quantification”)</p>
<p>EF(4D) Alias: EF</p>	<p>(18043-0, LN, “Left Ventricular Ejection Fraction”)</p>	<p>(G-0373, SRT, “Image Mode”) = (125231, DCM, “3D mode”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0023, 99GEMS, “4D Auto Left Ventricle Quantification”)</p>
<p>SV(4D) Alias: SV</p>	<p>(F-32120, SRT, “Stroke Volume”)</p>	<p>(G-0373, SRT, “Image Mode”) = (125231, DCM, “3D mode”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0023, 99GEMS, “4D Auto Left Ventricle Quantification”)</p>
<p>CO(4D) Alias: CO</p>	<p>(F-32100, SRT, “Cardiac Output”)</p>	<p>(G-0373, SRT, “Image Mode”) = (125231, DCM, “3D mode”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0023, 99GEMS, “4D Auto Left Ventricle Quantification”)</p>
<p>AWMA/GpeakSysSL(A2C) Alias: G peak SL(A2C)</p>	<p>(GEU-106-0001, 99GEMS, “Global Peak Longitudinal Strain”)</p>	<p>(111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”) (G-C036, SRT, “Measurement</p>

DIRECTION DOC2652554 REV 3

		Method”) = (GEU-106-0018, 99GEMS, “AFI”)
AWMA/GpeakSysSL(A4C) Alias: G peak SL(A4C)	(GEU-106-0001, 99GEMS, “Global Peak Longitudinal Strain”)	(111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)
AWMA/GpeakSysSL(APLAX) Alias: G peak SL(APLAX)	(GEU-106-0001, 99GEMS, “Global Peak Longitudinal Strain”)	(111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)
AWMA/GpeakSysSL(Avg) Alias: G peak SL(Avg)	(GEU-106-0001, 99GEMS, “Global Peak Longitudinal Strain”)	(G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)
AWMA/AVC Alias: AVC	(GEU-106-0003, 99GEMS, “Aortic Valve Closure”)	(G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)
AWMA/BS PeakSysSL Alias: BS peak sys SL	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(G-C0E3, SRT, “Finding Site”) = (R-10076, SRT, “left ventricle basal inferoseptal segment”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)
AWMA/MS PeakSysSL Alias: MS peak sys SL	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(G-C0E3, SRT, “Finding Site”) = (R-10078, SRT, “left ventricle mid

		<p>inferoseptal segment") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole") (G-C036, SRT, "Measurement Method") = (GEU-106- 0018, 99GEMS, "AFI")</p>
<p>AWMA/AS PeakSysSL Alias: AS peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, "Peak Longitudinal Strain")</p>	<p>(G-C0E3, SRT, "Finding Site") = (T- 32614, SRT, "left ventricle apical septal segment") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole") (G-C036, SRT, "Measurement Method") = (GEU-106- 0018, 99GEMS, "AFI")</p>
<p>AWMA/BL PeakSysSL Alias: BL peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, "Peak Longitudinal Strain")</p>	<p>(G-C0E3, SRT, "Finding Site") = (R- 1007A, SRT, "left ventricle basal anterolateral segment") (R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole") (G-C036, SRT, "Measurement Method") = (GEU-106- 0018, 99GEMS, "AFI")</p>
<p>AWMA/ML PeakSysSL Alias: ML peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, "Peak Longitudinal Strain")</p>	<p>(G-C0E3, SRT, "Finding Site") = (R- 1007C, SRT, "left ventricle mid anterolateral segment") (R-4089A, SRT, "Cardiac Cycle Point")</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		= (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)
AWMA/AL PeakSysSL Alias: AL peak sys SL	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(G-C0E3, SRT, “Finding Site”) = (T-3261C, SRT, “left ventricle apical lateral segment”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)
AWMA/BI PeakSysSL Alias: BI peak sys SL	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(G-C0E3, SRT, “Finding Site”) = (T-32615, SRT, “left ventricle basal inferior segment”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)
AWMA/MI PeakSysSL Alias: MI peak sys SL	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(G-C0E3, SRT, “Finding Site”) = (T-32616, SRT, “left ventricle mid inferior segment”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		Method”) = (GEU-106-0018, 99GEMS, “AFI”)
<p>AWMA/AI PeakSysSL</p> <p>Alias: AI peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32618, SRT, “left ventricle apical inferior segment”)</p> <p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”)</p> <p>(G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)</p>
<p>AWMA/BA PeakSysSL</p> <p>Alias: BA peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32619, SRT, “left ventricle basal anterior segment”)</p> <p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”)</p> <p>(G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)</p>
<p>AWMA/MA PeakSysSL</p> <p>Alias: MA peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T-32617, SRT, “left ventricle mid anterior segment”)</p> <p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”)</p> <p>(G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)</p>

<p>AWMA/AA PeakSysSL Alias: AA peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (T- 32613, SRT, “left ventricle apical anterior segment”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106- 0018, 99GEMS, “AFI”)</p>
<p>AWMA/BP PeakSysSL Alias: BP peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (R- 10079, SRT, “left ventricle basal inferolateral segment”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106- 0018, 99GEMS, “AFI”)</p>
<p>AWMA/MP PeakSysSL Alias: MP peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (R- 1007B, SRT, “left ventricle mid inferolateral segment”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106- 0018, 99GEMS, “AFI”)</p>
<p>AWMA/AP PeakSysSL Alias: AP peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (GEU-106-0025, 99GEMS, “left ventricle apical</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		<p>posterior segment”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)</p>
<p>AWMA/BAS PeakSysSL Alias: BAS peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (R-10075, SRT, “left ventricle basal anteroseptal segment”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)</p>
<p>AWMA/MAS PeakSysSL Alias: MAS peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (R-10077, SRT, “left ventricle mid anteroseptal segment”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)</p>
<p>AWMA/AAS PeakSysSL Alias: AAS peak sys SL</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(G-C0E3, SRT, “Finding Site”) = (GEU-106-0026, 99GEMS, “left ventricle apical anteroseptal segment”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT,</p>

		<p>“Systole” (G-C036, SRT, “Measurement Method”) = (GEU-106- 0018, 99GEMS, “AFI”)</p>
<p>AFI/BA PeakSysSL_ASE18 Alias: BA PeakSysSL ASE</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (T- 32619, SRT, “left ventricle basal anterior segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106- 0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)</p>
<p>AFI/BAS PeakSysSL_ASE18 Alias: BAS PeakSysSL ASE</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (R- 10075, SRT, “left ventricle basal anteroseptal segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106- 0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)</p>
<p>AFI/BIS PeakSysSL_ASE18 Alias: BIS PeakSysSL ASE</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (R- 10076, SRT, “left</p>

		<p>ventricle basal inferoseptal segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)</p>
<p>AFI/BI PeakSysSL_ASE18 Alias: BI PeakSysSL ASE</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (T-32615, SRT, “left ventricle basal inferior segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)</p>
<p>AFI/BIL PeakSysSL_ASE18 Alias: BIL PeakSysSL ASE</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (R-10079, SRT, “left ventricle basal inferolateral segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)</p>
<p>AFI/BAL PeakSysSL_ASE18 Alias: BAL PeakSysSL ASE</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		<p>= (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (R-1007A, SRT, “left ventricle basal anterolateral segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)</p>
<p>AFI/MA PeakSysSL_ASE18 Alias: MA PeakSysSL ASE</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (T-32617, SRT, “left ventricle mid anterior segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)</p>
<p>AFI/MAS PeakSysSL_ASE18 Alias: MAS PeakSysSL ASE</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (R-10077, SRT, “left ventricle mid anteroseptal segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI</p>

		with 18 segments following 2015 ASE recommendations")
AFI/MIS PeakSysSL_ASE18 Alias: MIS PeakSysSL ASE	(GEU-106-0002, 99GEMS, "Peak Longitudinal Strain")	(R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole") (G-C0E3, SRT, "Finding Site") = (R-10078, SRT, "left ventricle mid inferoseptal segment") (G-C036, SRT, "Measurement Method") = (GEU-106-0128, 99GEMS, "AFI with 18 segments following 2015 ASE recommendations")
AFI/MI PeakSysSL_ASE18 Alias: MI PeakSysSL ASE	(GEU-106-0002, 99GEMS, "Peak Longitudinal Strain")	(R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole") (G-C0E3, SRT, "Finding Site") = (T-32616, SRT, "left ventricle mid inferior segment") (G-C036, SRT, "Measurement Method") = (GEU-106-0128, 99GEMS, "AFI with 18 segments following 2015 ASE recommendations")
AFI/MIL PeakSysSL_ASE18 Alias: MIL PeakSysSL ASE	(GEU-106-0002, 99GEMS, "Peak Longitudinal Strain")	(R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole") (G-C0E3, SRT, "Finding Site") = (R-1007B, SRT, "left ventricle mid inferolateral segment")

		(G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)
AFI/MAL PeakSysSL_ASE18 Alias: MAL PeakSysSL ASE	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (R-1007C, SRT, “left ventricle mid anterolateral segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)
AFI/AA PeakSysSL_ASE18 Alias: AA PeakSysSL ASE	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (T-32613, SRT, “left ventricle apical anterior segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)
AFI/AAS PeakSysSL_ASE18 Alias: AAS PeakSysSL ASE	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”)

		<p>(G-C0E3, SRT, “Finding Site”) = (GEU-106-0026, 99GEMS, “left ventricle apical anteroseptal segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)</p>
<p>AFI/AIS PeakSysSL_ASE18 Alias: AIS PeakSysSL ASE</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (GEU-106-0125, 99GEMS, “left ventricle apical inferoseptal segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)</p>
<p>AFI/AI PeakSysSL_ASE18 Alias: AI PeakSysSL ASE</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (T-32618, SRT, “left ventricle apical inferior segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		following 2015 ASE recommendations”)
AFI/AIL PeakSysSL_ASE18 Alias: AIL PeakSysSL ASE	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (GEU-106-0126, 99GEMS, “left ventricle apical inferolateral segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)
AFI/AAL PeakSysSL_ASE18 Alias: AAL PeakSysSL ASE	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (GEU-106-0127, 99GEMS, “left ventricle apical anterolateral segment”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0128, 99GEMS, “AFI with 18 segments following 2015 ASE recommendations”)
AFI/BA PeakSysSL_Endo Alias: BA PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (R-4089A, SRT, “Cardiac Cycle Point”)

		= (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (T-32619, SRT, “left ventricle basal anterior segment”)
AFI/BAS PeakSysSL_Endo Alias: BAS PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (R-10075, SRT, “left ventricle basal anteroseptal segment”)
AFI/BS PeakSysSL_Endo Alias: BS PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C0E3, SRT, “Finding Site”) = (R-10076, SRT, “left ventricle basal inferoseptal segment”)
AFI/BI PeakSysSL_Endo Alias: BI PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”)

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		(G-C0E3, SRT, “Finding Site”) = (T-32615, SRT, “left ventricle basal inferior segment”)
AFI/BP PeakSysSL_Endo Alias: BP PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (G-C0E3, SRT, “Finding Site”) = (R-10079, SRT, “left ventricle basal inferolateral segment”)
AFI/BL PeakSysSL_Endo Alias: BL PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (G-C0E3, SRT, “Finding Site”) = (R-1007A, SRT, “left ventricle basal anterolateral segment”)
AFI/MA PeakSysSL_Endo Alias: MA PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”)

		(G-C0E3, SRT, “Finding Site”) = (R-32617, SRT, “left ventricle mid anterior segment”)
AFI/MAS PeakSysSL_Endo Alias: MAS PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (G-C0E3, SRT, “Finding Site”) = (R-10077, SRT, “left ventricle mid anteroseptal segment”)
AFI/MS PeakSysSL_Endo Alias: MS PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (G-C0E3, SRT, “Finding Site”) = (R-10078, SRT, “left ventricle mid inferoseptal segment”)
AFI/MI PeakSysSL_Endo Alias: MI PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (G-C0E3, SRT, “Finding Site”) = (T-

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		32616, SRT, “left ventricle mid inferior segment”)
AFI/MP PeakSysSL_Endo Alias: MP PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (G-C0E3, SRT, “Finding Site”) = (R-1007B, SRT, “left ventricle mid inferolateral segment”)
AFI/ML PeakSysSL_Endo Alias: ML PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (G-C0E3, SRT, “Finding Site”) = (R-1007C, SRT, “left ventricle mid anterolateral segment”)
AFI/AA PeakSysSL_Endo Alias: AA PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (G-C0E3, SRT, “Finding Site”) = (T-32613, SRT, “left

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		ventricle apical anterior segment”)
AFI/AAS PeakSysSL_Endo Alias: AAS PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (G-C0E3, SRT, “Finding Site”) = (GEU-106-0026, 99GEMS, “left ventricle apical anteroseptal segment”)
AFI/AS PeakSysSL_Endo Alias: AS PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (G-C0E3, SRT, “Finding Site”) = (T-32614, SRT, “left ventricle apical septal segment”)
AFI/AI PeakSysSL_Endo Alias: AI PeakSysSL Endo	(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)	(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0129, 99GEMS, “AFI on endocardium”) (G-C0E3, SRT, “Finding Site”) = (T-32618, SRT, “left

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		ventricle apical inferior segment")
AFI/AP PeakSysSL_Endo Alias: AP PeakSysSL Endo	(GEU-106-0002, 99GEMS, "Peak Longitudinal Strain")	(R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole") (G-C036, SRT, "Measurement Method") = (GEU-106-0129, 99GEMS, "AFI on endocardium") (G-C0E3, SRT, "Finding Site") = (GEU-106-0025, 99GEMS, "left ventricle apical posterior segment")
AFI/AL PeakSysSL_Endo Alias: AL PeakSysSL Endo	(GEU-106-0002, 99GEMS, "Peak Longitudinal Strain")	(R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole") (G-C036, SRT, "Measurement Method") = (GEU-106-0129, 99GEMS, "AFI on endocardium") (G-C0E3, SRT, "Finding Site") = (T-3261C, SRT, "left ventricle apical lateral segment")
AFI/PSD_Endo_ASE18 Alias: PSD Endo ASE18	(GEU-106-0131, 99GEMS, "Peak Strain Dispersion")	(R-4089A, SRT, "Cardiac Cycle Point") = (F-32020, SRT, "Systole") (G-C036, SRT, "Measurement Method") = (GEU-106-0130, 99GEMS, "AFI on endocardium with 18 segments following 2015 ASE recommendations")

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>AFI/BA PeakSysSL_Endo_ASE18 Alias: BA PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (T-32619, SRT, “left ventricle basal anterior segment”)</p>
<p>AFI/BAS PeakSysSL_Endo_ASE18 Alias: BAS PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (R-10075, SRT, “left ventricle basal anteroseptal segment”)</p>
<p>AFI/BIS PeakSysSL_Endo_ASE18 Alias: BIS PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		<p>2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (R-10076, SRT, “left ventricle basal inferoseptal segment”)</p>
<p>AFI/BI PeakSysSL_Endo_ASE18 Alias: BI PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (T-32615, SRT, “left ventricle basal inferior segment”)</p>
<p>AFI/BIL PeakSysSL_Endo_ASE18 Alias: BIL PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (R-10079, SRT, “left ventricle basal inferolateral segment”)</p>
<p>AFI/BAL PeakSysSL_Endo_ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>Alias: BAL PeakSysSL Endo ASE18</p>		<p>= (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (R-1007A, SRT, “left ventricle basal anterolateral segment”)</p>
<p>AFI/MA PeakSysSL_Endo_ASE18 Alias: MA PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (T-32617, SRT, “left ventricle mid anterior segment”)</p>
<p>AFI/MAS PeakSysSL_Endo_ASE18 Alias: MAS PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		<p>2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (R-10077, SRT, “left ventricle mid anteroseptal segment”)</p>
<p>AFI/MIS PeakSysSL_Endo_ASE18 Alias: MIS PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (R-10078, SRT, “left ventricle mid inferoseptal segment”)</p>
<p>AFI/MI PeakSysSL_Endo_ASE18 Alias: MI PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (T-32616, SRT, “left ventricle mid inferior segment”)</p>
<p>AFI/MIL PeakSysSL_Endo_ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”)</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

<p>Alias: MIL PeakSysSL Endo ASE18</p>		<p>= (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (R-1007B, SRT, “left ventricle mid inferolateral segment”)</p>
<p>AFI/MAL PeakSysSL_Endo_ASE18 Alias: MAL PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (R-1007C, SRT, “left ventricle mid anterolateral segment”)</p>
<p>AFI/AA PeakSysSL_Endo_ASE18 Alias: AA PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		<p>2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (T-32613, SRT, “left ventricle apical anterior segment”)</p>
<p>AFI/AAS PeakSysSL_Endo_ASE18 Alias: AAS PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (GEU-106-0026, 99GEMS, “left ventricle apical anteroseptal segment”)</p>
<p>AFI/AIS PeakSysSL_Endo_ASE18 Alias: AIS PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (GEU-106-0125, 99GEMS, “left ventricle apical inferoseptal segment”)</p>

<p>AFI/AI PeakSysSL_Endo_ASE18 Alias: AI PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (T-32618, SRT, “left ventricle apical inferior segment”)</p>
<p>AFI/AIL PeakSysSL_Endo_ASE18 Alias: AIL PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with 18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (GEU-106-0126, 99GEMS, “left ventricle apical inferolateral segment”)</p>
<p>AFI/AAL PeakSysSL_Endo_ASE18 Alias: AAL PeakSysSL Endo ASE18</p>	<p>(GEU-106-0002, 99GEMS, “Peak Longitudinal Strain”)</p>	<p>(R-4089A, SRT, “Cardiac Cycle Point”) = (F-32020, SRT, “Systole”) (G-C036, SRT, “Measurement Method”) = (GEU-106-0130, 99GEMS, “AFI on endocardium with</p>

GE HEALTHCARE
STATEMENT

DIRECTION DOC2652554 REV 3

		18 segments following 2015 ASE recommendations”) (G-C0E3, SRT, “Finding Site”) = (GEU-106-0127, 99GEMS, “left ventricle apical anterolateral segment”)
AFI/AVC Alias: AVC	(GEU-106-0003, 99GEMS, “Aortic Valve Closure”)	(G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)
AFI/GPeakSysSL(APLAX) Alias: GpeakSysSL(APLAX)	(GEU-106-0001, 99GEMS, “Global Peak Longitudinal Strain”)	(G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”) (111031, DCM, “Image View”) = (G-0395, SRT, “Apical long axis”)
AFI/GpeakSysSL(A4C) Alias: GpeakSysSL(A4C)	(GEU-106-0001, 99GEMS, “Global Peak Longitudinal Strain”)	(G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”) (111031, DCM, “Image View”) = (G-A19C, SRT, “Apical four chamber”)
AFI/GpeakSysSL(A2C) Alias: GpeakSysSL(A2C)	(GEU-106-0001, 99GEMS, “Global Peak Longitudinal Strain”)	(G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”) (111031, DCM, “Image View”) = (G-A19B, SRT, “Apical two chamber”)
AFI/GpeakSysSL(Avg) Alias: GpeakSysSL(Avg)	(GEU-106-0001, 99GEMS, “Global Peak Longitudinal Strain”)	(G-C036, SRT, “Measurement Method”) = (GEU-106-0018, 99GEMS, “AFI”)