

## Parathyroid hormone (parathormone, parathyrin) - PTH, intact

REF	$\Sigma$	SYSTEM
11972103 122	100	Elecsys 2010 MODULAR ANALYTICS E170 <b>cobas e 411</b> <b>cobas e 601</b> <b>cobas e 602</b>

## English

## Intended use

Immunoassay for the in vitro quantitative determination of intact parathyroid hormone in human serum and plasma for the differential diagnosis of hypercalcemia and hypocalcemia. This assay can be used intraoperatively.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on Elecsys and **cobas e** immunoassay analyzers.

## Summary

Parathyroid hormone (PTH) is formed in the parathyroid glands and secreted into the blood stream. Intact PTH consists of a single polypeptide chain containing 84 amino acids and has a molecular weight of approximately 9500 daltons.

The biologically active N-terminal fragment has a half-life of only a few minutes. Selective measurement of the (mainly) intact parathyroid hormone permits direct ascertainment of the secretory activity of the parathyroid glands.<sup>1,2</sup>

PTH, together with vitamin D and calcitonin, brings about mobilization of calcium and phosphate from the skeletal system and increases the uptake of calcium in the intestine and the excretion of phosphate via the kidneys. The constancy of the blood calcium level is ensured by the interaction of PTH and calcitonin. The secretion of PTH is inhibited by high calcium concentrations and promoted by low calcium concentrations.

Parathyroid gland disorders lead to elevated or depressed blood calcium levels (hypercalcemia or hypocalcemia) brought about by a change in the secretion of PTH.

Detection of subfunctioning parathyroid glands (hypoparathyroidism) requires the use of a highly sensitive test in order to be able to measure PTH levels well below normal.<sup>3,4</sup>

Hyperfunctioning of the parathyroid glands results in an increased secretion of PTH (hyperparathyroidism). Primary causes are adenomas of the parathyroid glands. In secondary hyperparathyroidism the blood calcium level is low as a result of other pathological states (e.g. vitamin D deficiency).

Today, great significance is attached to the determination of the PTH and calcium concentrations when assessing hyperparathyroidism.

The determination of PTH intraoperatively during adenoma resection in the parathyroid glands has also been reported for primary hyperparathyroidism,<sup>5,6,7</sup> secondary hyperparathyroidism relating to renal failure,<sup>8,9</sup> and tertiary hyperparathyroidism post renal transplant surgery.<sup>10</sup> Because PTH has a reported half life of 3-5 minutes,<sup>11</sup> a significant drop in PTH levels after resection of the abnormal gland or glands enables the surgeon to assess the completeness of resection and whether all hyperfunctioning parathyroid tissue has been removed from the patient.<sup>12</sup>

The NACB guidelines recommend that baseline samples be obtained preoperation and pre-excision of the suspected hyperfunctioning gland.<sup>13</sup> Specimens for PTH testing should be drawn at 5 and 10 minutes post resection and that a > 50 % reduction in PTH levels from the highest baseline be used as criteria for surgical success. Additional samples may be necessary as it has been shown that sensitivity can increase with time.<sup>14</sup> Failure of PTH to drop below recommended levels indicates that either 1) residual hyperfunctioning tissue is still present and further exploration may be necessary, as was in the case of two patients, both with a fifth ectopic parathyroid gland requiring further surgery,<sup>7</sup> or 2) a spike in PTH levels during adenoma mobilization occurred.<sup>15</sup> Intraoperative PTH measurements offer fast, reliable assessment when all hyperfunctioning parathyroid tissue has been removed during the surgical process.

The Elecsys assay for determining intact PTH employs a sandwich test principle in which a biotinylated monoclonal antibody reacts with the N-terminal fragment (1-37) and a monoclonal antibody labeled with a ruthenium complex<sup>a)</sup> reacts with the C-terminal fragment (38-84).

The antibodies used in this assay are reactive with epitopes in the amino acid regions 26-32 and 37-42.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)<sub>3</sub><sup>2+</sup>)

## Test principle

Sandwich principle. Total duration of assay: 18 minutes.

- 1st incubation: 50 µL of sample, a biotinylated monoclonal PTH-specific antibody, and monoclonal PTH-specific antibody labeled with a ruthenium complex form a sandwich complex.
- 2nd incubation: After addition of streptavidin-coated microparticles, the complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the reagent barcode.

## Reagents - working solutions

The reagent rackpack is labeled as PTH.

- M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL:  
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 Anti-PTH-Ab~biotin (gray cap), 1 bottle, 7 mL:  
Biotinylated monoclonal anti-PTH antibody (mouse) 2.3 mg/L;  
phosphate buffer 100 mmol/L, pH 7.0; preservative.
- R2 Anti-PTH-Ab~Ru(bpy)<sub>3</sub><sup>2+</sup> (black cap), 1 bottle, 7 mL:  
Monoclonal anti-PTH antibody (mouse) labeled with ruthenium complex 2.0 mg/L; phosphate buffer 100 mmol/L, pH 7.0;  
preservative.

## Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

## Reagent handling

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in from the respective reagent barcodes.

## Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability:	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	12 weeks

**Parathyroid hormone (parathormone, parathyrin) - PTH, intact**

Stability:	
on the analyzers	8 weeks

**Specimen collection and preparation**

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes.

K<sub>2</sub>-EDTA and K<sub>3</sub>-EDTA plasma.

Because of the short half-life of PTH, it is recommended that, when serum is needed, the blood should be centrifuged immediately.

Preference should be given to EDTA plasma, as it is stable longer than serum.

Criterion: Method comparison serum versus plasma, slope 0.9-1.1 + intercept within  $< \pm 2 \times$  analytical sensitivity (LDL) + coefficient of correlation  $> 0.95$ .

Serum: Stable for 8 hours at 15-25 °C, 2 days at 2-8 °C, 6 months at -20 °C.

Plasma: Stable for 2 days at 15-25 °C, 3 days at 2-8 °C, 6 months at -20 °C.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

Do not use samples and controls stabilized with azide.

Ensure the samples, calibrators and controls are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples, calibrators and controls on the analyzers should be analyzed/measured within 2 hours.

**Materials provided**

See "Reagents – working solutions" section for reagents.

**Materials required (but not provided)**

- [REF] 11972219122, PTH CalSet, for 4 x 1 mL
- [REF] 05618860190, PreciControl Varia, for 2 x 3 mL each of PreciControl Varia 1 and 2
- General laboratory equipment
- Elecsys 2010, MODULAR ANALYTICS E170 or **cobas e** analyzer

Accessories for Elecsys 2010 and **cobas e** 411 analyzers:

- [REF] 11662988122, ProCell, 6 x 380 mL system buffer
  - [REF] 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
  - [REF] 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
  - [REF] 11933159001, Adapter for SysClean
  - [REF] 11706802001, Elecsys 2010 AssayCup, 60 x 60 reaction vessels
  - [REF] 11706799001, Elecsys 2010 AssayTip, 30 x 120 pipette tips
- Accessories for MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers:
- [REF] 04880340190, ProCell M, 2 x 2 L system buffer
  - [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
  - [REF] 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use
  - [REF] 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change
  - [REF] 03004899190, PreClean M, 5 x 600 mL detection cleaning solution
  - [REF] 12102137001, AssayTip/AssayCup Combimagazine M, 48 magazines x 84 reaction vessels or pipette tips, waste bags
  - [REF] 03023150001, WasteLiner, waste bags
  - [REF] 03027651001, SysClean Adapter M

Accessories for all analyzers:

- [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

**Assay**

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers.

MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers: PreClean M solution is necessary.

Bring the cooled reagents to approximately 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

**Calibration**

**Traceability:** This method has been standardized against a commercial PTH test (RIA). The recovery of the NIBSC 95/646 (WHO) standard was assessed by testing dilutions in human serum covering the measuring range (40-4000 pg/mL) on 16 analyzers (**cobas e** 411 and **cobas e** 601 analyzers). The mean recovery was 100 %  $\pm$  4 %.

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

**Calibration frequency:** Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer). Renewed calibration is recommended as follows:

- after 12 weeks when using the same reagent lot
- after 7 days (when using the same reagent kit on the analyzer)
- as required: e.g. quality control findings outside the defined limits

**Quality control**

For quality control, use PreciControl Varia.

In addition, other suitable control material can be used.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

**Calculation**

The analyzer automatically calculates the analyte concentration of each sample (either in pg/mL or pmol/L).

Conversion factors:  $\text{pg/mL} \times 0.106 = \text{pmol/L}$   
 $\text{pmol/L} \times 9.43 = \text{pg/mL}$

**Limitations - interference**

The assay is unaffected by icterus (bilirubin  $< 1112 \mu\text{mol/L}$  or  $< 65 \text{mg/dL}$ ), lipemia (Intralipid  $< 1500 \text{mg/dL}$ ) and biotin ( $< 205 \text{nmol/L}$  or  $< 50 \text{ng/mL}$ ).

The assay is affected by hemolysis  $\geq 0.15 \text{g/dL}$ . Do not analyze samples that show visible signs of hemolysis.

Criterion: Recovery within  $\pm 10 \%$  of initial value.

Samples should not be taken from patients receiving therapy with high biotin doses (i.e.  $> 5 \text{mg/day}$ ) until at least 8 hours following the last biotin administration.

No interference was observed from rheumatoid factors up to a concentration of 1500 IU/mL.

There is no high-dose hook effect at PTH concentrations up to 17000 pg/mL (1802 pmol/L).

## Parathyroid hormone (parathormone, parathyrin) - PTH, intact

In vitro tests were performed on 16 commonly used pharmaceuticals. No interference with the assay was found.

In rare cases, interference due to extremely high titers of antibodies to analyte-specific antibodies, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

### Limits and ranges

#### Measuring range

1.20-5000 pg/mL or 0.127-530 pmol/L (defined by the lower detection limit and the maximum of the master curve). Values below the lower detection limit are reported as < 1.20 pg/mL (< 0.127 pmol/L). Values above the measuring range are reported as > 5000 pg/mL (> 530 pmol/L).

#### Lower limits of measurement

##### Lower detection limit of the test

Lower detection limit: 1.20 pg/mL (0.127 pmol/L)

The lower detection limit represents the lowest measurable analyte level that can be distinguished from zero. It is calculated as the value lying two standard deviations above that of the lowest standard (master calibrator, standard 1 + 2 SD, repeatability study, n = 21).

#### Dilution

Not necessary due to the broad measuring range.

#### Expected values

15-65 pg/mL (1.6-6.9 pmol/L)<sup>16,17</sup>

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

#### Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

#### Precision

Precision was determined using Elecsys reagents and pooled human sera in a modified protocol (EP5-A) of the CLSI (Clinical and Laboratory Standards Institute): 6 times daily for 10 days (n = 60); repeatability on MODULAR ANALYTICS E170 analyzer, n = 21. The following results were obtained:

Elecsys 2010 and <b>cobas e 411</b> analyzers								
Sample	Repeatability					Intermediate precision		
	Mean		SD		CV	SD		CV
	pg/mL	pmol/L	pg/mL	pmol/L	%	pg/mL	pmol/L	%
HS <sup>b)</sup> 1	26.7	2.83	0.711	0.075	2.7	1.73	0.184	6.5
HS 2	52.5	5.56	0.853	0.091	1.6	2.07	0.220	3.9
HS 3	261	27.7	4.0	0.424	1.5	7.81	0.829	3.0

b) HS = human serum

MODULAR ANALYTICS E170, <b>cobas e 601</b> and <b>cobas e 602</b> analyzers					
Sample	Repeatability				
	Mean		SD		CV
	pg/mL	pmol/L	pg/mL	pmol/L	%
HS 1	21.9	2.32	0.44	0.05	2.0
HS 2	35.0	3.71	0.43	0.05	1.2
HS 3	123	13.04	1.31	0.14	1.1

MODULAR ANALYTICS E170, <b>cobas e 601</b> and <b>cobas e 602</b> analyzers					
Sample	Intermediate precision				
	Mean		SD		CV
	pg/mL	pmol/L	pg/mL	pmol/L	%
HS 1	23.2	2.46	0.79	0.08	3.4
HS 2	80.9	8.58	2.01	0.21	2.5
HS 3	240	25.4	6.72	0.71	2.8

Precision was determined using Elecsys reagents and controls in a separate study according to protocol EP5-A2 of the CLSI (Clinical and Laboratory Standards Institute): 2 runs per day in duplication each for 21 days (n = 84). The following results were obtained:

Elecsys 2010 and <b>cobas e 411</b> analyzers								
Sample	Repeatability					Intermediate precision		
	Mean		SD		CV	SD		CV
	pg/mL	pmol/L	pg/mL	pmol/L	%	pg/mL	pmol/L	%
PC <sup>c)</sup> Varia 1	54.6	5.79	0.657	0.070	1.2	1.11	0.118	2.0
PC Varia 2	182	19.3	2.43	0.258	1.3	3.14	0.333	1.7

c) PC = PreciControl

MODULAR ANALYTICS E170, <b>cobas e 601</b> and <b>cobas e 602</b> analyzers								
Sample	Repeatability					Intermediate precision		
	Mean		SD		CV	SD		CV
	pg/mL	pmol/L	pg/mL	pmol/L	%	pg/mL	pmol/L	%
PC Varia 1	54.7	5.80	0.943	0.100	1.7	0.977	0.104	1.8
PC Varia 2	184	19.5	2.65	0.281	1.4	3.11	0.330	1.7

#### Method comparison

A comparison of the Elecsys PTH assay (y) with a commercial PTH test (x) using clinical samples gave the following correlations (pg/mL):

Number of samples measured: 152

Passing/Bablok<sup>18</sup> Linear regression

$$y = 1.01x + 4.86$$

$$y = 0.83x + 19.4$$

$$r = 0.886$$

$$r = 0.991$$

The sample concentrations were between approximately 1.4 and 1880 pg/mL (0.15 and 199 pmol/L).

#### Analytical specificity

No cross-reactivities were found for: Osteocalcin, PTH fragment 1-37, PTH-related protein (1-86), bone-specific alkaline phosphatase, and  $\beta$ -CrossLaps.

#### Functional sensitivity

6.0 pg/mL (0.64 pmol/L)

The functional sensitivity is the lowest analyte concentration that can be reproducibly measured with an intermediate precision CV of  $\leq 20\%$ .

#### Clinical investigations in intraoperative use

In 2006, the National Academy of Clinical Biochemistry published their Laboratory Medicine Practice Guidelines for point of care testing, entitled Evidence Based Practice for Point of Care Testing.<sup>13</sup> The guidelines recommend the use of intraoperative parathyroid hormone testing 1) for patients undergoing surgery for hyperparathyroidism, especially in minimally invasive or directed procedures, 2) for patients undergoing reoperation, and 3) as a replacement for traditional laboratory

## Parathyroid hormone (parathormone, parathyrin) - PTH, intact

measurements of PTH during venous localization in order to help the angiography team guide sampling.

The guidelines further recommend for patients undergoing parathyroidectomy for hyperparathyroidism that baseline samples be obtained preoperation exploration and pre-excision of the gland, and that post-excision sampling be drawn at 5 and 10 minutes post resection with a 50 % reduction in PTH concentrations from the highest baseline level. The guidelines also caution that additional samples may be necessary.<sup>13</sup>

PTH testing during parathyroid surgery was conducted by several groups of investigators using the Elecsys PTH immunoassay.<sup>6,7,8,9,10</sup>

The overall sensitivity and specificity of the assay to demonstrate successful surgery as defined by postoperative reduction of calcium levels was 99.6 % and 93.7 %, respectively.

## References

- Silverman R, Yalow RS. Heterogeneity of parathyroid hormone: Clinical and physiologic implications. *J Clin Invest* 1973;52:1958-1971.
- Flehtje D, Schmidt-Gayk H, Fischer S, et al. Intact parathyroid hormone in primary hyperparathyroidism. *Br J Surg* 1990;77:168-172.
- Nussbaum S, Potts JT. Advances in Immunoassays for Parathyroid Hormone. Clinical Applications to Skeletal Disorders of Bone and Mineral Metabolism. In Bilezikian JP, Levine MA, Marcus R (eds). *The Parathyroids: Basic and Clinical Concepts*. Raven Press, New York 1994:157-169.
- Berson SA, Yalow RS, Aurbach GD, et al. Immunoassay of bovine and human parathyroid hormone. *Proc Natl Acad Sci USA* 1963;49:613-617.
- Bergenfels A, Nordén NE, Ahrén B. Intraoperative fall in plasma levels of intact parathyroid hormone after removal of one enlarged parathyroid gland in hyperthyroid patients. *Eur J Surg* 1991;157:109-112.
- Ohe MN, Santos RO, Kunii IS, et al. Usefulness of a rapid immunometric assay for intra-operative parathyroid hormone measurements. *Braz J Med Biol Res* 2003;36(6):715-721.
- Ohe MN, Santos RO, Kunii IS, et al. Usefulness of intra-operative PTH measurement in primary and secondary hyperparathyroidism: experience with 109 patients. *Arq Bras Endocrinol Metab* 2006;50(5):869-875.
- Seehofer D, Rayes N, Ulrich F, et al. Intra-operative measurement of intact parathyroid hormone in renal hyperparathyroidism by an inexpensive routine assay. *Langenbecks Arch Surg* 2001;386(6):440-443.
- Seehofer D, Rayes N, Klupp J, et al. Predictive value of intact parathyroid hormone measurement during surgery for renal hyperparathyroidism. *Langenbecks Arch Surg* 2005;390(3):222-229.
- Haustein SV, Mack E, Starling JR, et al. The role of intra-operative parathyroid hormone testing in patients with tertiary hyperparathyroidism after renal transplantation. *Surgery* 2005;138(6):1066-1071.
- Maier GW, Kreis ME, Renn W, et al. Parathyroid hormone after adenectomy for primary hyperparathyroidism: A study of peptide hormone elimination kinetics in humans. *Jour Clin Endocrinol Metab* 1998;83(11):3853-3856.
- Carter AB, Howanitz TJ. Intra-operative testing for parathyroid hormone: a comprehensive review of the use of the assay and the relevant literature. *Arch Pathol Lab Med* 2003;127:1424-1442.
- Nichols JH, Christenson RH, Clarke W, et al. National Academy of Clinical Biochemistry Laboratory Medicine Practice Guidelines: Evidence Based Practice for Point of Care Testing. AACC Press:2006.
- Bergenfels A, Isaksson A, Lindblom P, et al. Measurement of parathyroid hormone in patients with primary hyperparathyroidism undergoing first and reoperative surgery. *Br J Surg* 1998;85:1129-1132.
- Yang GP, Levine S, Weigel RJ. A spike in parathyroid hormone during neck exploration may cause a false-negative intra-operative assay result. *Arch Surg* 2001;136:945-949.
- Blind E. Measurement of Intact Parathyroid Hormone by an Extracting Two-Site Immunometric Assay. In: Schmidt-Gayk H, Armbruster FP, Bouillon R, (eds). *Calcium regulating hormones, vitamin D metabolites, an cyclic AMP*. Heidelberg: Springer 1990:151.
- Thomas L. Parathyroid hormone (PTH). *Clinical Laboratory Diagnosis. TH-Books*, Frankfurt. 1st english edition 1998:248-250.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. *J Clin Chem Clin Biochem* 1988 Nov;26(11):783-790.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

## Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

	Contents of kit
	Analyzers/Instruments on which reagents can be used
	Reagent
	Calibrator
	Volume after reconstitution or mixing
	Global Trade Item Number

COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2015, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
www.roche.com

