

# Certificate of Analysis

RAIDYNAS LTD J. SAVICKIO STR. 4 01108 VILNIUS LITHUANIA	Customer Order: 20190111ASHLAND 1 Ordered Quantity: 11.000 EA Shipped From: ASI US FISKEVILLE PLT Order Number: 8298607/000010 Delivery: 851924364/000010 Date Shipped: January 31, 2019 Sold To Number: 697165
--	--

RAD-SURE 25 GY INDICATOR@200EA  
 Rad-Sure(TM) 25 GY INDICATOR blood irradiation indicators  
 Ashland Material Number: 831379

Batch: 0328510U25

Characteristics	Specification	Results
Visual Density , before irradiation, du	<= 0.60	0.49
Visual Density, after irradiation, du	>= 1.52	1.61
Date of Manufacture		July 24, 2018
Retest Date		July 23, 2021
Ordered Quantity		11.000 EA

Notes:

Retest Date is for Ashland's internal purposes only.

Expiry Date is assigned to the last day of the month as shown above.

## Certificate of Analysis

RAIDYNAS LTD J. SAVICKIO STR. 4 01108 VILNIUS LITHUANIA		Customer Order: 20190111ASHLAND 1 Ordered Quantity: 11.000 EA Shipped From: ASI US FISKEVILLE PLT Order Number: 8298607/000010 Delivery: 851924364/000010 Date Shipped: January 31, 2019 Sold To Number: 697165
RAD-SURE 25 GY INDICATOR@200EA Rad-Sure(TM) 25 GY INDICATOR blood irradiation indicators Ashland Material Number: 831379		
		Batch: 0328510U25
Characteristics	Specification	Results



Sean Quarles

Quality Assurance Manager

This COA is printed from a secure computer system ensuring the batch was properly released by the Quality Department and is valid with electronic signature.

Manufactured by:  
ASHLAND SPECIALTY INGREDIENTS G.P.  
AFFILIATE OF ASHLAND  
538 MAIN STREET  
FISKEVILLE RI 02823

Contact:  
BIBIANA KING  
Phone: 908-243-4233  
Email: BKING@ASHLAND.COM