

To whom it may concern,

We, STEPHANIX, who are official manufacturer of medical X-ray equipment, having factory at 10, rue Jean Moulin – ZI du Bayon – 42150 La Ricamarie hereby, confirm that:

1. As we constantly research most optimal equipment solutions, recently the software for the use camera integrated in the vertical column in our XTREME PREMIUM unit has been added to PRIMO S software and the technical documentation regarding this change has not been publicly announced yet. We are currently working on adjustments for the manual of the equipment where we will combine all the small changes and changes regarding the integrated camera use. As we declared before, our equipment has video camera integrated into vertical column which could be controlled from operator workstation and STEPHANIX will thoroughly instruct our partners on the operation of said integrated camera.
2. Quality control tests are performed by the manufacturer, (i.e. STEPHANIX) during the production of the XTREME PREMIUM unit with needed equipment and phantom (PMMA layers (25 cm x 25 cm x 20 cm)). We confirm that the client or final user will not have to perform these tests during installation or use of the equipment. There are no other quality control or calibration tests that must be performed by our partners during installation or the end user during the use of equipment.
3. As it was indicated in XTREME PREMIUM unit manual, the equipment recognizes the collimated area and displays only the region that was indicated before the procedure. This process may as well be called virtual collimation (different manufacturers call it differently). This virtual collimation may be performed automatically via afore mentioned algorithm which virtually indicates collimated area only or, if preferred, manually.

**Yannick
CHAPOTOT**

Signature
numérique de
Yannick CHAPOTOT
Date : 2024.08.28
15:12:09 +02'00'