

STATEMENT

To: Vilnius University Hospital Santaros Clinics

Subject: Confirmation of specifications of technical requirement, purchase no. 686593:
„Ultragarsinis aparatas (Nr.7383)“

To whom it may concern,

We, Philips Medical Systems Nederland B.V. , a well-established manufacturer and/or supplier of medical equipment, having its head office at Veenpluis 6, 5684PC Best, The Netherlands ("Philips"), hereby confirm that ultrasound system: **„5500G“**, offered by UAB **„Arbor Medical Corporation LT“**, fulfil all tender requirements, including the below described requirements:

2.1.2. Linear probe „eL18-4“ can be used for adult patients

3.9. Maximum framerate in 2D mode is 1900 frames/sec.

Eindhoven, The Netherlands, 13 September 2023

Philips Medical Systems Nederland B.V.



Mr. Frank Geerings

Tender Business Support Manager

International Business Team

High Tech Campus 52, Building HTC52.5

5656 AG Eindhoven, the Netherlands

frank.geerings@philips.com

www.healthcare.philips.com

PATVIRTINIMAS

Kam: Vilniaus Universiteto ligoninei Santaros klinikoms
Tema: Techninių specifikacijų patvirtinimas, pirkimo nr. 686593: „**Ultragarsinis aparatas (Nr. 7383)**“

Gerbiamieji,

Mes, Philips Medical Systems Nederland B.V., gerai žinomas medicinos įrangos gamintojas ir (arba) tiekėjas, kurio pagrindinė buveinė yra Veenpluis 6, 5684PC Best, Nyderlandai („Philips“), patvirtiname, kad ultragarso sistema: „5500G“, siūloma UAB „Arbor Medical Corporation LT“, atitinka visus konkurso reikalavimus, įskaitant žemiau nurodytus reikalavimus:

- 2.1.2.** Linijinis daviklis „eL18-4“ gali būti naudojamas skenuoti suaugusiu pacientus
- 3.9.** Maksimalus kadrų dažnis 2D režime yra 1900 kadrų/sek.

Eindhovenas, Nyderlandai, Rugsėjo 13d, 2023m.
Philips Medical Systems Nederland B.V.

Frank Geerings
Viešųjų pirkimų skyriaus vadovas
Tarptautinio verslo padalinys
High Tech Campus 52, Pastatas HTC52.5
5656 AG Eindhovenas, Nyderlandai
Frank.geerings@philips.com
www.healthcare.philips.com

Vertimas tikras
Direktorius Mindaugas Daugėl

