

Declaration of EC Conformity



GANSHORN
Medizin Electronic GmbH

Manufacturer: **Ganshorn Medizin Electronic GmbH**
Address: **Industriestraße 6-8**
97618 Niederlauer, Germany

declares, on it's own responsibility, that the following product:

Product designation: **Pulmonary function diagnostic system**

Product name: **SpiroScout**

Is in conformity with the Medical Device Directive:

Medical Device Directive 93 / 42 / EEC

According to Annex IX of this Directive the SpiroScout is classified as:

Device: Active medical device - class IIa

A Conformity Assessment Procedure according to Annex II of the Medical Device Directive 93/42/EEC (with exception clause 4) was conducted with the Notified Body:

TÜV SÜD
Product Services GmbH
Notified Body Regulatory Code 0123
Ridlerstraße 65
80339 München



GANSHORN MEDIZIN
ELECTRONIC GMBH



GANSHORN
MEDIZIN ELECTRONIC


Jürgen Behringer (CEO)
Chief Executive Officer

GANSHORN Medizin Electronic GmbH
Industriestrasse 6 - 8 • D-97618 Niederlauer
Tel: +49 9771 6222 0 • Fax: +49 9771 6222 55

Niederlauer, November 01, 2018

Note: This document version is valid until November 2023
Declaration of EC-Conformity_SpiroScout_eng_rev02.doc