



WS TECH S.C. R. Lorens, W. Pielech, J. Mazur

Blood drawing chair FK-02




WS TECH S.C.
R. LORENS, W. PIELECH, J. MAZUR
38-500 Sanok, ul. Okulickiego 43
NIP 687-17-22-142 REGON 370513655



WS TECH S.C. R. Lorens, W. Pielech, J. Mazur

Universal chair FK-02 is intended for injection procedures. The frame of the medical treatment chair is made of metal constructions painted with powder polymer paint and has four height-adjustable legs. Two adjustable armrests make the injection procedure easier on both the left and right side of the patient. The chair has adjustment mechanisms for backrest and footrest. All parts of the chair are made of non-hazardous, non-allergenic materials, easy to clean and disinfect, fire and swelling resistant. The legs of the chair are covered with clear plastic, which is resistant to abrasion and disinfectants.

Parameters	Model FK-02
Backrest adjustment angle	0° - 95°
The total length of the chair	170 cm
Seat width	50 cm
Overall width of the chair	71 cm
Trendelenburg position	from 0° to -3°
Adjustment of the handles by means of a ball bearing	from -15° to + 65°
Permissible load	150 kg



WS TECH S.C.
R. LORENS, W. PIELECH, J. MAZUR
38-500 Sanok, ul. Okulickiego 43
NIP 687-17-22-142 REGON 370513655



WS TECH S.C. R. Lorens, W. Pielech, J. Mazur



Chair table 25x35 cm; load 10 kg

WS TECH S.C.
R. LORENS, W. PIELECH, J. MAZUR
38-500 Sanok, ul. Okulickiego 43
NIP 687-17-12142 REGON 370513655

