

7.1 Reguluojamas moterų šlapimo nelaikymo raištis

Titel:/ Title:	Produktbeschreibung und -spezifikation / Product Description and Specification		
Dokument Nr.: / Document No.:	F90_001B_PBS_Slings_A		

8 Product Specification

8.1 Technical Details

5. Visas ilgis 450mm
6. Plotis 12 mm

8.1.1 Slings

Criteria	PFR5021	TOA5121	TVA5021	SUI5001	SUI5011	SUI5021
Length of the implant	450 mm	450 mm	450 mm	67 mm	450 mm	450 mm
Width of the implant	12 mm	12 mm	12 mm	14 mm	12 mm	12 mm
Thickness of mesh	0,62 mm	0,62 mm	0,62 mm	0,62 mm	0,62 mm	0,62 mm
Pore size	>0,4 mm ²	>0,4 mm ²	>0,4 mm ²	>0,4 mm ²	>0,4 mm ²	>0,4 mm ²
Thickness of suture used for the mesh	0,15 mm	0,15 mm	0,15 mm	0,15 mm	0,15 mm	0,15 mm
Total length (including pulling threads)	1380 mm	1400 mm	1400 mm	n.a.	1400 mm	1400 mm
Size of pull-in sutures	USP0	USP0	USP0	n.a.	USP0	USP0
Size of adjustment sutures	n.a.	USP0	USP0	n.a.	n.a.	USP0
Length of the central part	n.a.	n.a.	n.a.	n.a.	35 mm	35 mm

Table 7: Slings

8.1.2 Tunneller

Criteria	TOA5130	TOA5170	TOA5140	TOA5180	TVA5030	TVA5040
Max. diameter of the Tunnellers needle	4 mm	3,2 mm	4 mm	3,2 mm	4 mm	3,2 mm
Width of the tip	5 mm	4 mm	5 mm	4 mm	5 mm	4 mm
Size of the slot	8 x 1,7 mm	8 x 1,7 mm	8 x 1,7 mm	8 x 1,7 mm	8 x 1,7 mm	8 x 1,7 mm