

28 January 2025

## Statement

Regarding: Complementary confirmation on Integrated cell preparation system for cryopreservation for Vilnius university hospital Santaros klinikos published tender Nr. 569941

It is hereby stated for Sepax C-Pro system (29264741):

4. Stalinio tipo įrenginys, kuris gali būti pernešamas vieno darbuotojo jėgomis

1) System is benchtop (Desktop-type) type device that can be lifted up and carried by one person. Please refer to system brochure where we state instrument dimensions and weight that are as follows: Width – 27 cm, Length – 40 cm and Height – 46 cm. Weight of the device – 16,3 kg. We trust that this clearly shows that the device can be carried by one person if needed without any trouble or using heavy equipment.

3.2.3. Dedikuota programinė skirta prietaiso valdymui, kokybės ir atsekamumo užtikrinimui, suderinama su geros gamybos praktikos (angl. GMP) reikalavimais

2) System has a dedicated software for instrument management, quality assurance and traceability. We confirm that this software fully complies to Good Manufacturing Practices (GMP) requirements.

3) Authorized and trained service engineer will provide full device installation and IQ/OQ will be on customer demand.

6. Autorizuotas ir apmokytas serviso inžinierius atliks sistemos instaliavimą ir IQ/OQ

4) The ability to connect to an internal information system that integrates process management and traceability to a higher organization

3.5. Galimybė prijungti prie aukštesnės organizacijos procesų valdymą ir atsekamumą integruojančios informacinės sistemos



Zone Lead, authorized company representor  
Cytiva Finland Oy