

REF	$\Sigma$	SYSTEM
11820796 190	100	Elecsys 2010 MODULAR ANALYTICS E170 <b>cobas e 411</b> <b>cobas e 601</b> <b>cobas e 602</b>

## English

### Intended use

Immunoassay for the in vitro quantitative determination of digoxin in human serum and plasma. Measurements are used in the diagnosis and treatment of digoxin overdose and in monitoring levels of digoxin to ensure proper therapy.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on Elecsys and **cobas e** immunoassay analyzers.

### Summary

Digoxin is a widely prescribed steroidal cardio-active glycoside. It acts by binding and inhibiting the Na<sup>+</sup>/K<sup>+</sup>-ATPase which in the end increases the intracellular Ca<sup>2+</sup> concentration.<sup>1,2</sup> This results in a positive inotrope effect which makes digoxin a beneficial drug for heart failure. It improves the strength of myocardial contraction and results in the beneficial effects of increased cardiac output, increased Left Ventricular Ejection Fraction, and decreased Pulmonary Capillary Wedge pressure.<sup>3,4</sup> Digoxin therapy also results in stabilized and slowed ventricular pulse rate.<sup>5</sup>

Although the availability of crystalline digoxin has permitted the standardization of drug dosage, therapeutic administration inadvertently, yet frequently, results in toxicity. Importantly, symptoms of digoxin toxicity often mimic the cardiac arrhythmias for which the drug was originally prescribed. Digoxin concentrations of 0.9-2.0 ng/mL in serum or plasma are normally considered to be therapeutic.<sup>6,7</sup> However, later studies observed an increased risk for mortality for digoxin concentrations of 1.2 ng/mL and higher.<sup>8,9</sup> These observations are also reflected in the "ESC Guidelines for the diagnosis and treatment of acute and chronic heart failure 2008" which recommend a therapeutic concentration range for digoxin of 0.6-1.2 ng/mL<sup>10</sup> and other guidelines even recommend limits < 1.0 ng/mL.<sup>11</sup>

Toxicity of digoxin may reflect several factors:

1. The drug has a low therapeutic ratio (i.e. a very small difference exists between therapeutic and toxic tissue levels);
2. Individuals vary in their response to digoxin;
3. Absorption of various tablet forms of digoxin may vary over a two-fold range;<sup>12,13</sup>
4. Susceptibility to digitalis toxicity apparently increases with age mainly associated with renal impairment.<sup>4</sup>

In combination with other clinical data, monitoring serum or plasma levels may provide the physician with useful information to aid in adjusting patient dosage, and achieving optimal therapeutic effect, while avoiding both subtherapeutic and harmful toxic drug levels.<sup>14</sup>

The Elecsys Digoxin assay employs a competitive test principle using a monoclonal antibody specifically directed against digoxin. Digoxin in the sample competes with the added digoxin derivative labeled with biotin for the binding sites on the ruthenylated antibody-complex<sup>a)</sup>.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)<sub>3</sub><sup>2+</sup>)

### Test principle

Competition principle. Total duration of assay: 18 minutes.

- 1st incubation: By incubating the sample (10 µL) with a digoxin-specific ruthenium-labeled antibody, an immunocomplex is formed, the amount of which is dependent upon the analyte concentration in the sample.
- 2nd incubation: After addition of streptavidin-coated microparticles and a digoxin derivative labeled with biotin, the still-vacant sites of the ruthenium labeled antibodies become occupied, with formation of an antibody-hapten complex. The entire complex becomes bound to the solid phase via interaction of biotin and streptavidin.

- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the reagent barcode.

### Reagents - working solutions

The reagent rackpack is labeled as DIGO.

- M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL:  
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 Anti-digoxin-Ab~Ru(bpy)<sub>3</sub><sup>2+</sup> (gray cap), 1 bottle, 10 mL:  
Monoclonal anti-digoxin antibody (mouse) labeled with ruthenium complex 15 µg/L; phosphate buffer 100 mmol/L, pH 7.0; preservative.
- R2 Digoxin-derivative~biotin (black cap), 1 bottle, 10 mL:  
Biotinylated digoxigenin 1.06 ng/mL; biotin 15 µg/L; phosphate buffer 100 mmol/L, pH 7.0; preservative.

### Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

### Reagent handling

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in from the respective reagent barcodes.

### Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability:	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	12 weeks
on the analyzers	8 weeks

### Specimen collection and preparation

Blood samples for digoxin analyses should be collected at trough levels which is just prior to the next drug dose or at least 12 hours, and preferably 24 hours after the previous digoxin dose. Considering a blood elimination half-life of 1.5 days for digoxin, steady state blood concentrations require approximately 1 week after initiation of therapy – or longer in case of abnormal kidney function.<sup>15</sup>

Only the specimens listed below were tested and found acceptable. Serum collected using standard sampling tubes or tubes containing separating gel.

Li-heparin, K<sub>2</sub>-EDTA and K<sub>3</sub>-EDTA plasma. Li-heparin plasma tubes containing separating gel can be used.

# Digoxin

Criterion: Slope 0.9-1.1 + intercept within  $\pm 1 \times$  Limit of Blank + coefficient of correlation  $\geq 0.95$ .

Stable for 7 days at 15-25 °C, 14 days at 2-8 °C, 6 months at -20 °C. Freeze only once.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay. Heat-inactivated serum can be used.

Ensure the samples, calibrators and controls are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples, calibrators and controls on the analyzers should be analyzed/measured within 2 hours.

## Materials provided

See "Reagents – working solutions" section for reagents.

## Materials required (but not provided)

- [REF] 11820907322, Digoxin CalSet, 4 x 1.5 mL
  - [REF] 04917049190, PreciControl Cardiac II, for 2 x 2 mL each of PreciControl Cardiac II 1 and 2
  - [REF] 11732277122, Diluent Universal, 2 x 16 mL sample diluent or [REF] 03183971122, Diluent Universal, 2 x 36 mL sample diluent
  - General laboratory equipment
  - Elecsys 2010, MODULAR ANALYTICS E170 or **cobas e** analyzer
- Accessories for Elecsys 2010 and **cobas e** 411 analyzers:

- [REF] 11662988122, ProCell, 6 x 380 mL system buffer
- [REF] 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
- [REF] 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
- [REF] 11933159001, Adapter for SysClean
- [REF] 11706802001, Elecsys 2010 AssayCup, 60 x 60 reaction vessels
- [REF] 11706799001, Elecsys 2010 AssayTip, 30 x 120 pipette tips

Accessories for MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers:

- [REF] 04880340190, ProCell M, 2 x 2 L system buffer
- [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- [REF] 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use
- [REF] 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change
- [REF] 03004899190, PreClean M, 5 x 600 mL detection cleaning solution
- [REF] 12102137001, AssayTip/AssayCup Combimagazine M, 48 magazines x 84 reaction vessels or pipette tips, waste bags
- [REF] 03023150001, WasteLiner, waste bags
- [REF] 03027651001, SysClean Adapter M

Accessories for all analyzers:

- [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

## Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers: PreClean M solution is necessary.

Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in

exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers.

Bring the cooled reagents to approximately 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

## Calibration

**Traceability:** This method has been standardized by weighing United States Pharmacopoeia (USP) digoxin reference material into analyte free human serum.

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

**Calibration frequency:** Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer). Renewed calibration is recommended as follows:

- after 1 month (28 days) when using the same reagent lot
- after 7 days (when using the same reagent kit on the analyzer)
- as required: e.g. quality control findings outside the defined limits

## Quality control

For quality control, use PreciControl Cardiac II.

In addition, other suitable control material can be used.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

## Calculation

The analyzer automatically calculates the analyte concentration of each sample (either in nmol/L or ng/mL).

Conversion factors:  $\text{nmol/L} \times 0.78 = \text{ng/mL}$   
 $\text{ng/mL} \times 1.28 = \text{nmol/L}$

## Limitations - interference

The assay is unaffected by icterus (bilirubin  $\leq 1129 \mu\text{mol/L}$  or  $\leq 66 \text{ mg/dL}$ ), hemolysis (Hb  $\leq 0.621 \text{ mmol/L}$  or  $\leq 1.0 \text{ g/dL}$ ), lipemia (Intralipid  $\leq 1500 \text{ mg/dL}$ ) and biotin ( $\leq 409 \text{ nmol/L}$  or  $\leq 100 \text{ ng/mL}$ ).

Criterion: Recovery within  $\pm 0.08 \text{ ng/mL}$  ( $\pm 0.10 \text{ nmol/L}$ ) for digoxin concentrations  $\leq 0.8 \text{ ng/mL}$  ( $\leq 1.02 \text{ nmol/L}$ ) or  $\pm 10 \%$  for concentrations  $> 0.8\text{-}4.0 \text{ ng/mL}$  ( $> 1.02\text{-}5.12 \text{ nmol/L}$ ) or  $\pm 12 \%$  for concentrations  $> 4.0 \text{ ng/mL}$  ( $> 5.12 \text{ nmol/L}$ ).

Samples should not be taken from patients receiving therapy with high biotin doses (i.e.  $> 5 \text{ mg/day}$ ) until at least 8 hours following the last biotin administration.

No interference was observed from rheumatoid factors up to a concentration of 1630 IU/mL.

In vitro tests were performed on a panel of commonly used pharmaceuticals. While 34 of these showed no interference with the assay, uzara, nabumetone, hydrocortisone, pentoxifylline and canrenone were identified to cause falsely elevated digoxin values at concentrations of the recommended daily dose.

Spironolactone causes elevated digoxin results above (drug) levels of 10000 ng/mL. Canrenone causes elevated digoxin results above (drug) levels of 80000 ng/mL.

Digoxin-like immunoreactive substances (DLIS) have been identified in blood from patients in renal failure, liver failure, and pregnant women in their third trimester. Studies have shown that the presence of DLIS in a sample can result in a false elevation of digoxin when assayed by commercially available immunoassays.<sup>16,17,18</sup>

As stated by the manufacturers of digitalis antidotes, the therapeutic antibody fragments against digitalis (e.g. DigiFab, DigiBind) will interfere

with digitalis immunoassay measurements.<sup>19</sup> Therefore, the manufacturer of DigiFab recommends to obtain samples for determination of digoxin concentration prior to antidote administration.<sup>19</sup> As a consequence Elecsys Digoxin concentrations may be falsely elevated if measured in the presence of the antidote until the Fab fragments are eliminated from the body.<sup>19</sup>

In rare cases, interference due to extremely high titers of antibodies to analyte-specific antibodies, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

## Limits and ranges

### Measuring range

0.2-5.0 ng/mL or 0.26-6.4 nmol/L (defined by the Limit of Detection and the maximum of the master curve). Values below the Limit of Detection are reported as < 0.2 ng/mL or < 0.26 nmol/L. Values above the measuring range are reported as > 5.0 ng/mL or > 6.4 nmol/L (or up to 10.0 ng/mL or 12.8 nmol/L for 2-fold diluted samples).

### Lower limits of measurement

*Limit of Blank, Limit of Detection and Limit of Quantitation*

Limit of Blank = 0.15 ng/mL (0.19 nmol/L)

Limit of Detection = 0.2 ng/mL (0.26 nmol/L)

Limit of Quantitation = 0.4 ng/mL (0.51 nmol/L) with a total allowable error of ≤ 20 %

The Limit of Blank, Limit of Detection and Limit of Quantitation were determined in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP17-A2 requirements.

The Limit of Blank is the 95<sup>th</sup> percentile value from  $n \geq 60$  measurements of analyte-free samples over several independent series. The Limit of Blank corresponds to the concentration below which analyte-free samples are found with a probability of 95 %.

The Limit of Detection is determined based on the Limit of Blank and the standard deviation of low concentration samples. The Limit of Detection corresponds to the lowest analyte concentration which can be detected (value above the Limit of Blank with a probability of 95 %).

The Limit of Quantitation is defined as the lowest amount of analyte in a sample that can be accurately quantitated with a total allowable error of ≤ 20 %.

### Dilution

Samples with digoxin concentrations above the measuring range can be diluted with Diluent Universal. The recommended dilution is 1:2 (either automatically by the MODULAR ANALYTICS E170, Elecsys 2010 or **cobas e** analyzers or manually). The concentration of the diluted sample must be > 2.5 ng/mL or > 3.2 nmol/L.

After manual dilution, multiply the result by the dilution factor.

After dilution by the analyzers, the MODULAR ANALYTICS E170, Elecsys 2010 and **cobas e** software automatically takes the dilution into account when calculating the sample concentration.

### Expected values

The recommended therapeutic range for digoxin is 0.6-1.2 ng/mL (0.77-1.5 nmol/L) (ESC Guideline 2008<sup>10</sup>) or even 0.5-1.0 ng/mL (0.64-1.3 nmol/L).<sup>20</sup> Particularly the upper end of the therapeutic range is controversial and concentrations up to 2.0 ng/mL (2.6 nmol/L) may still be applied.<sup>6,7</sup> Concentrations > 2.0 ng/mL are generally considered toxic.<sup>21</sup> Some overlap of toxic and non-toxic values has been reported.<sup>22</sup>

Therefore, clinical diagnosis should be based on clinical and laboratory data. Each laboratory should establish an acceptable reporting format and identify procedures for the reporting of abnormal results.

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

### Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

### Precision

Precision was determined using Elecsys reagents, samples and controls in a protocol (EP5-A3) of the CLSI (Clinical and Laboratory Standards

Institute): 2 runs per day in duplication each for 21 days ( $n = 84$ ). The following results were obtained:

Elecsys 2010 and <b>cobas e</b> 411 analyzers					
Sample	Mean nmol/L	Repeatability		Intermediate precision	
		SD nmol/L	CV %	SD nmol/L	CV %
Human serum 1	0.724	0.025	3.4	0.046	6.4
Human serum 2	1.39	0.035	2.5	0.080	5.8
Human serum 3	2.37	0.049	2.1	0.106	4.5
Human serum 4	3.05	0.071	2.3	0.117	3.8
Human serum 5	5.98	0.152	2.5	0.382	6.4
PC <sup>b)</sup> CARDII1	1.54	0.045	2.9	0.066	4.3
PC CARDII2	3.51	0.131	3.7	0.143	4.1

b) PC = PreciControl

Elecsys 2010 and <b>cobas e</b> 411 analyzers					
Sample	Mean ng/mL	Repeatability		Intermediate precision	
		SD ng/mL	CV %	SD ng/mL	CV %
Human serum 1	0.565	0.019	3.4	0.036	6.4
Human serum 2	1.09	0.027	2.5	0.063	5.8
Human serum 3	1.85	0.039	2.1	0.083	4.5
Human serum 4	2.38	0.055	2.3	0.092	3.8
Human serum 5	4.67	0.119	2.5	0.299	6.4
PC CARDII1	1.20	0.035	2.9	0.051	4.3
PC CARDII2	2.74	0.102	3.7	0.111	4.1

MODULAR ANALYTICS E170, <b>cobas e</b> 601 and <b>cobas e</b> 602 analyzers					
Sample	Mean nmol/L	Repeatability		Intermediate precision	
		SD nmol/L	CV %	SD nmol/L	CV %
Human serum 1	0.712	0.045	6.3	0.058	8.2
Human serum 2	1.36	0.038	2.8	0.058	4.3
Human serum 3	2.33	0.058	2.5	0.084	3.6
Human serum 4	2.94	0.069	2.4	0.111	3.8
Human serum 5	5.49	0.147	2.7	0.276	5.0
PC CARDII1	1.55	0.037	2.4	0.056	3.6
PC CARDII2	3.50	0.047	1.3	0.082	2.3

MODULAR ANALYTICS E170, <b>cobas e</b> 601 and <b>cobas e</b> 602 analyzers					
Sample	Mean ng/mL	Repeatability		Intermediate precision	
		SD ng/mL	CV %	SD ng/mL	CV %
Human serum 1	0.556	0.035	6.3	0.046	8.2
Human serum 2	1.07	0.030	2.8	0.045	4.3
Human serum 3	1.82	0.045	2.5	0.066	3.6
Human serum 4	2.29	0.054	2.4	0.087	3.8
Human serum 5	4.29	0.115	2.7	0.216	5.0
PC CARDII1	1.21	0.029	2.4	0.044	3.6

MODULAR ANALYTICS E170, <b>cobas e 601</b> and <b>cobas e 602</b> analyzers					
		Repeatability		Intermediate precision	
Sample	Mean ng/mL	SD ng/mL	CV %	SD ng/mL	CV %
PC CARDII2	2.74	0.037	1.3	0.064	2.3

### Analytical specificity

For the Co-analytes tested, the following relative Co-analyte reactivities were found:

Co-analyte	Concentration ED50 ng/mL	Relative Co-analyte reactivity %
α-acetyldigoxin	1.18	77.9
β-acetyldigoxin	1.09	84.4
β-methyldigoxin	1.05	87.9
Lanatoside C	1.31	65.2
Deslanoside	1.08	85.6
Digoxigenin-bis-digitoxoside	0.853	108
Digoxigenin-mono-digitoxoside	0.603	141

For the substances tested, the following cross-reactivities were found:

Substances	Concentration tested ng/mL	Cross-reactivity %
Digitoxin	250	0.522
Digitoxigenin	250	0.529
Digoxigenin	6.00	31.3
Dihydrodigoxin	1000	0.201
K-strophanthine	1250	0.137

No significant cross-reactivity (< 0.01 %) was found for the following substances (tested concentration 5000 ng/mL):

Cortisol, prednisone, β-estradiol, d-aldosterone, DHEA, dexamethasone, furosemide, sulthiame, quinidine (free base) and oleandrin. For testosterone and ouabain a cross-reactivity of < 0.1 % was found at 5000 ng/mL. For progesterone a cross-reactivity of < 0.05 % was found at 5000 ng/mL.

### References

- Hauptmann PJ, Kelly RA Digitalis. *Circulation* 1999;99:1265-1270.
- Katz A, Lifshitz Y, Bab-Dinitz E, et al. Selectivity of Digitalis Glycosides for Isoforms of Human Na,K-ATPase. *J Biol Chem* 2010 Jun;285(25):19582-19592.
- Eichhorn EJ, Gheorghide M Digoxin. *Progress Cardiovasc Diseases* 2002 Jan/Feb;44(4):251-266.
- Gheorghide M, van Veldhuisen DJ, Colucci WS. Contemporary Use of Digoxin in the Management of Cardiovascular Disorders. *Circulation* 2006;113:2556-2564.
- Hoffman BF, Bigger JT Jr. In: Gilman AG, Goodman LS, Gilman A, eds. *The Pharmacological Basis of Therapeutics*. 6th ed. New York, NY: MacMillan; 1980:729-760.
- Oellerich M. Pharmaka (Drug monitoring). In: Thomas L (ed.). *Labor und Diagnose, TH-Books, Frankfurt*, 5. edition, 1998:1174. Englisch: *Clinical Laboratory*. 1st English Edition 1998:1151.
- Jortani SA, Valdes R Jr. Digoxin and Its Related Endogenous Factors. *Critical Reviews Clin Lab Sci* 1997;34(3):225-274.
- Rathore SS, Curtis JP, Wang Y, et al. Association of Serum Digoxin Concentration and Outcome in Patients with Heart Failure. *JAMA* 2003 Feb;289(7):871-878.

- Adams KF, Patterson JH, Gattis WA, et al. Relationship of Serum Digoxin Concentration to Mortality and Morbidity in Women in the Digitalis Investigation Group Trial. *J Am Coll Cardiology* 2005;46(3):497-504.
- Dickstein K, Cohen-Solal A, Filippatos G, et al. ESC Guidelines for the diagnosis and treatment of acute and chronic heart failure 2008: the Task Force for the Diagnosis and Treatment of Acute and Chronic Heart Failure 2008 of the European Society of Cardiology. Developed in collaboration with the Heart Failure Association of the ESC (HFA) and endorsed by the European Society of Intensive Care Medicine (ESICM). *Eur Heart J* 2008;29:2388-2442.
- Goldberger ZD, Goldberger AL Therapeutic Ranges of Serum Digoxin Concentrations in Patients with Heart Failure. *Am J Cardiol* 2012;109:1818-1821.
- Lindenbaum J, Mellow MH, Blackstone MO, et al. Variation in Biologic Availability of Digoxin from Four Preparations. *New Engl J Med* 1971;285:1344-1347.
- Lindenbaum J, Butler VP Jr., Murphy JE, et al. Correlation of Digoxin-Tablet Dissolution Rate with Biological Availability. *Lancet* 1973;1:1215-1217.
- Butler VP Jr., Lindenbaum J. Serum Digitalis Measurements in the Assessment of Digitalis Resistance and Sensitivity. *Am J Med* 1975;58:460-469.
- Frendl G, Sodickson AC, Chung MK, et al. 2014 AATS guidelines for the prevention and management of perioperative atrial fibrillation and flutter for thoracic surgical procedures. *J Thorac Cardiovasc Surg* 2014;148(3):e153-93.
- Keys PW, Stafford RW. In: Taylor WJ, Finn AL, eds. *Individualizing Drug Therapy: Practical Applications of Drug Monitoring*. New York, Gross, Townsend, Frank, Inc; 1981; vol 3:1-21.
- Valdes R Jr. Endogenous digoxin-like immunoreactive factors: impact on digoxin measurements and potential physiological implications. *Clin Chem* 1985;31(9):1525-1532.
- Valdiva R, Hornig Y, Gross S, et al. Digoxin-like Immunoreactive Factor Cross-reactivity in the CEDIA Digoxin R Assay on the RA-1000. *Clin Chem* 1990;36(6):1111.
- DigiFab® Package Insert. P12011D (1-Aug-2014). BTG International Inc.
- Terra SG, Washam JB, Dunham GD, et al. Therapeutic Range of Digoxin's Efficacy in Heart Failure: What Is The Evidence? *Pharmacotherapy* 1999;19(10):1123-1126.
- Matzuk MM, Shlomchik M, Shaw LM. *Therapeutic Drug Monitoring* 1991;13:215-219.
- Beller GA, Smith TW, Abelmann WH, et al. Digitalis Intoxication: A Prospective Clinical Study with Serum Level Correlations. *New Engl J Med* 1971;284:989-997.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

### Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

	Contents of kit
	Analyzers/Instruments on which reagents can be used
	Reagent
	Calibrator
	Volume after reconstitution or mixing
	Global Trade Item Number

# Digoxin



COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2015, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
[www.roche.com](http://www.roche.com)

