

Folate III

Folate

| REF | Σ | SYSTEM |
|--------------|----------|--|
| 04476433 190 | 100 | Elecsys 2010 MODULAR ANALYTICS E170 cobas e 411 cobas e 601 cobas e 602 |

English

Intended use

Binding assay for the in vitro quantitative determination of folate in human serum and plasma.

The binding assay is intended for use on Elecsys and **cobas e** immunoassay analyzers.

Summary

Nutritional and macrocytic anemias can be caused by a deficiency of folate. This deficiency can result from diets devoid of raw fruits, vegetables or other foods rich in folic acid, as may be the case with chronic alcoholics, drug addicts, the elderly or persons of low socioeconomic status, etc. In addition, low serum folate during pregnancy has been associated with neural tube defects in the fetus.¹ Dietary deficiency and malabsorption are the major causes of folate deficiency in humans.² Folate is necessary for normal metabolism, DNA synthesis and red blood cell regeneration. Untreated deficiencies may lead to megaloblastic anemia.

Since a deficiency of either vitamin B₁₂ or folate can cause megaloblastic anemia, it is advisable to determine the concentration of both vitamin B₁₂ and folate in order to properly diagnose the etiology of anemia. Radioassays were first reported for folate in 1973.^{3,4,5,6}

The majority utilize ¹²⁵I-folate radiolabeled tracers and natural binding proteins (milk binding protein, folate binding protein). The various commercial assays differ in their free versus bound separation techniques and choice of specimen pretreatment.

The Elecsys Folate III assay employs a competitive test principle using natural folate binding protein (FBP) specific for folate. Folate in the sample competes with the added folate (labeled with biotin) for the binding sites on FBP (labeled with ruthenium complex^{a)}).

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)₃²⁺)

Test principle

Competition principle. Total duration of assay: 27 minutes.

- 1st incubation: By incubating 25 µL of sample with the folate pretreatment reagents 1 and 2, bound folate is released from endogenous folate binding proteins.
- 2nd incubation: By incubating the pretreated sample with the ruthenium labeled folate binding protein, a folate complex is formed, the amount of which is dependent upon the analyte concentration in the sample.
- 3rd incubation: After addition of streptavidin-coated microparticles and folate labeled with biotin, the unbound sites of the ruthenium labeled folate binding protein become occupied, with formation of a ruthenium labeled folate binding protein-folate biotin complex. The entire complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the reagent barcode.

Reagents - working solutions

The reagent rackpack (M, R1, R2) and the pretreatment reagents (PT1, PT2) are labeled as Fol III.

PT1 Pretreatment reagent 1 (white cap), 1 bottle, 4 mL:

Sodium 2-mercaptoethanesulfonate (MESNA) 40 g/L, pH 5.5.

PT2 Pretreatment reagent 2 (gray cap), 1 bottle, 5 mL:

Sodium hydroxide 25 g/L.

M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL:

Streptavidin-coated microparticles 0.72 mg/mL; preservative.

R1 Folate binding protein~Ru(bpy)₃²⁺ (gray cap), 1 bottle, 9 mL:

Ruthenium labeled folate binding protein 75 µg/L; human serum albumin (stabilizer); borate/phosphate/citrate buffer 70 mmol/L, pH 5.5; preservative.

R2 Folate~biotin (black cap), 1 bottle, 8 mL:

Biotinylated folate 17 µg/L; biotin 120 µg/L; human serum albumin (stabilizer); borate buffer 100 mmol/L, pH 9.0; preservative.

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:



Danger

H290 May be corrosive to metals.

H314 Causes severe skin burns and eye damage.

Prevention:

P280 Wear protective gloves/ protective clothing/ eye protection/ face protection.

Response:

P301 + P330 IF SWALLOWED: Rinse mouth. Do NOT induce vomiting. + P331

P303 + P361 IF ON SKIN (or hair): Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower. + P353

P304 + P340 IF INHALED: Remove person to fresh air and keep comfortable for breathing. + P310 Immediately call a POISON CENTER or doctor/physician.

P305 + P351 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. + P338

P390 Absorb spillage to prevent material damage.

Product safety labeling primarily follows EU GHS guidance.

Contact phone: all countries: +49-621-7590

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of

Folate III

Folate

donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV. The testing methods applied were FDA-approved or cleared in compliance with the European Directive 98/79/EC, Annex II, List A.

However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.^{7,8}

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Reagent handling

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in from the respective reagent barcodes.

Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

| Stability: | |
|-------------------------|----------------------------------|
| unopened at 2-8 °C | up to the stated expiration date |
| after opening at 2-8 °C | 8 weeks |
| on the analyzers | 2 weeks |

Specimen collection and preparation

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel.

Li-heparin plasma. Li-heparin plasma tubes containing separating gel can be used.

Criterion: Method comparison serum versus Li-heparin plasma, slope 0.9-1.1 + intercept within $\pm 2x$ analytical sensitivity (LDL/LoB), coefficient of correlation ≥ 0.95 .

Serum: Stable for 2 hours at 20-25 °C, 2 days at 2-8 °C, 1 month at -20 °C. Freeze only once. Protect from light. Store the samples at 2-8 °C if they cannot be measured immediately.

Li-heparin plasma: Stable for 2 hours at 20-25 °C, 2 days at 2-8 °C. Do not freeze samples containing Li-heparin. Protect from light. Store the samples at 2-8 °C if they cannot be measured immediately.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Samples should not subsequently be altered with additives (biocides, anti-oxidants or substances possibly changing the pH of the sample) in order to avoid erroneous folate recovery.

Centrifuge samples containing precipitates before performing the assay.

Do not use heat-inactivated samples.

Ensure the samples, calibrators and controls are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples, calibrators and controls on the analyzers should be analyzed/measured within 2 hours.

Note: Hemolysis may significantly increase folate values due to high concentrations of folate in red blood cells. Therefore, hemolyzed samples are not suitable for use in this assay. Samples for folate determinations should be collected from fasting persons.

Materials provided

See "Reagents – working solutions" section for reagents.

Materials required (but not provided)

- [REF] 04874072190, Folate III CalSet, for 4 x 1.0 mL

- [REF] 05618860190, PreciControl Varia, for 2 x 3 mL each of PreciControl Varia 1 and 2
 - [REF] 11732277122, Diluent Universal, 2 x 16 mL sample diluent or [REF] 03183971122, Diluent Universal, 2 x 36 mL sample diluent
 - General laboratory equipment
 - Elecsys 2010, MODULAR ANALYTICS E170 or **cobas e** analyzer
- Accessories for Elecsys 2010 and **cobas e** 411 analyzers:
- [REF] 11662988122, ProCell, 6 x 380 mL system buffer
 - [REF] 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
 - [REF] 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
 - [REF] 11933159001, Adapter for SysClean
 - [REF] 11706802001, Elecsys 2010 AssayCup, 60 x 60 reaction vessels
 - [REF] 11706799001, Elecsys 2010 AssayTip, 30 x 120 pipette tips
- Accessories for MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers:
- [REF] 04880340190, ProCell M, 2 x 2 L system buffer
 - [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
 - [REF] 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use
 - [REF] 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change
 - [REF] 03004899190, PreClean M, 5 x 600 mL detection cleaning solution
 - [REF] 12102137001, AssayTip/AssayCup Combimagazine M, 48 magazines x 84 reaction vessels or pipette tips, waste bags
 - [REF] 03023150001, WasteLiner, waste bags
 - [REF] 03027651001, SysClean Adapter M
- Accessories for all analyzers:
- [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers.

MODULAR ANALYTICS E170, **cobas e** 601 and **cobas e** 602 analyzers: PreClean M solution is necessary.

Bring the cooled reagents to approximately 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

Calibration

Traceability: This method has been standardized against the Folate II assay ([REF] 03253678).

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

Calibration frequency: Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer). Renewed calibration is recommended as follows:

- after 1 month (28 days) when using the same reagent lot
- after 7 days (when using the same reagent kit on the analyzer)
- as required: e.g. quality control findings outside the defined limits

Quality control

For quality control, use PreciControl Varia.

In addition, other suitable control material can be used.

Folate III

Folate



| Elecsys 2010 and cobas e 411 analyzers | | | | | | | | |
|---|--------|-------|---------------|-------|-----|------------------------|-------|-----|
| Sample | Mean | | Repeatability | | | Intermediate precision | | |
| | nmol/L | ng/mL | nmol/L | ng/mL | % | nmol/L | ng/mL | % |
| PC V ^d 1 | 9.19 | 4.05 | 0.533 | 0.235 | 5.8 | 0.867 | 0.382 | 9.4 |
| PC V2 | 26.8 | 11.8 | 1.28 | 0.562 | 4.8 | 1.92 | 0.845 | 7.2 |

c) HS = human serum
d) PC V = PreciControl Varia

| MODULAR ANALYTICS E170, cobas e 601 and cobas e 602 analyzers | | | | | | | | |
|---|--------|-------|---------------|-------|-----|------------------------|-------|-----|
| Sample | Mean | | Repeatability | | | Intermediate precision | | |
| | nmol/L | ng/mL | nmol/L | ng/mL | % | nmol/L | ng/mL | % |
| HS 1 | 9.99 | 4.40 | 0.341 | 0.150 | 3.4 | 0.400 | 0.176 | 4.0 |
| HS 2 | 14.6 | 6.43 | 0.661 | 0.291 | 4.5 | 0.926 | 0.408 | 6.3 |
| HS 3 | 16.3 | 7.19 | 0.556 | 0.245 | 3.4 | 0.824 | 0.363 | 5.0 |
| HS 4 | 34.3 | 15.1 | 0.822 | 0.362 | 2.4 | 1.56 | 0.688 | 4.6 |
| PC V1 | 8.88 | 3.91 | 0.449 | 0.198 | 5.1 | 0.624 | 0.275 | 7.0 |
| PC V2 | 27.5 | 12.1 | 0.842 | 0.371 | 3.1 | 0.940 | 0.414 | 3.4 |

Method comparison

a) A comparison of the Elecsys Folate III assay (MODULAR ANALYTICS E170 analyzer; calibrated with Folate III CalSet; y) and the Elecsys Folate II assay (MODULAR ANALYTICS E170 analyzer; calibrated with Folate II CalSet II; x) using clinical samples gave the following correlations (ng/mL):

Number of samples measured: 98

Passing/Bablok¹⁰ Linear regression
 $y = 1.11x + 0.317$ $y = 1.12x + 0.236$
 $r = 0.939$ $r = 0.994$

The sample concentrations were between approximately 1.96 and 14.6 ng/mL (approximately 4.45 and 33.1 nmol/L).

b) A comparison of the Elecsys Folate III assay on the MODULAR ANALYTICS E170 analyzer (y) with the Elecsys Folate III assay on the Elecsys 2010 analyzer (x) (both tests have been calibrated with Folate III CalSet) using clinical samples gave the following correlations (ng/mL):

Number of samples measured: 149

Passing/Bablok¹⁰ Linear regression
 $y = 0.956x + 0.129$ $y = 0.925x + 0.314$
 $r = 0.858$ $r = 0.976$

The sample concentrations were between approximately 3.10 and 14.6 ng/mL (approximately 7.04 and 33.1 nmol/L).

c) A comparison of the Elecsys Folate III assay on the MODULAR ANALYTICS E170 analyzer (y) with another commercially available assay (x) using clinical samples gave the following correlations (ng/mL):

Number of samples measured: 170

Passing/Bablok¹⁰ Linear regression
 $y = 0.794x + 2.15$ $y = 0.645x + 2.95$
 $r = 0.673$ $r = 0.843$

The sample concentrations were between approximately 1.19 and 18.9 ng/mL (approximately 2.70 and 42.9 nmol/L).

Analytical specificity

The following cross-reactivities were found:

Aminopterin 2.7 %

Folinic acid 2.3 %
Amethopterin 2.3 %

Functional sensitivity

≤ 2.0 ng/mL (≤ 4.54 nmol/L)

The functional sensitivity is the lowest analyte concentration that can be reproducibly measured with an intermediate precision CV of 20 %.

References

- 1 Rush D. Folate Supplements Prevent Recurrence of Neural Tube Defects, FDA Dietary Supplement Task Force. Nutrition Reviews 1992;50(1):22-28.
- 2 Herbert V. Drugs effective in megaloblastic anemias. In Goodman LS and Gilman A (eds): The Pharmacological Basis of Therapeutics, 5th Ed, MacMillan Co, 1975;1324-1349.
- 3 Dunn RT, Foster LB. Radioassay of serum Folate. Clin Chem 1973;19:1101-1105.
- 4 Rothenberg SP, DaCosta M, Rosenberg BS. A radioassay for serum Folate: Use of a two phase sequential incubation, ligand-binding system. New Eng J Med 1972;285(25):1335-1339.
- 5 Gutcho S, Mansbach L. Simultaneous radioassay of serum Folate and folic acid. Clin Chem 1977;23:1609-1614.
- 6 BIO RAD Quantaphase B-12/Folate Radioassay Instruction Manual. March 1995.
- 7 Occupational Safety and Health Standards: bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- 8 Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.
- 9 Pfeiffer CM, Johnson CL, Jain RB, et al. Trends in blood folate and vitamin B-12 concentrations in the United States, 1988-2004. Am J Clin Nutr 2007;86:718-727.
- 10 Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. J Clin Chem Clin Biochem 1988 Nov;26(11):783-790.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

| | |
|--|---|
| | Contents of kit |
| | Analyzers/Instruments on which reagents can be used |
| | Reagent |
| | Calibrator |
| | Volume after reconstitution or mixing |
| | Global Trade Item Number |

COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2015, Roche Diagnostics

ms_04476433190V13.0

Folate II

Folate



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com



cobas®