

# HCG STAT



Human chorionic gonadotropin, STAT (Short Turn Around Time)

REF	$\Sigma$		SYSTEM
03300811 190	100		Elecsys 2010 <b>cobas e 411</b> <b>cobas e 601</b> <b>cobas e 602</b>

**English**

**Intended use**

Immunoassay for the in vitro quantitative determination of human chorionic gonadotropin in human serum and plasma. This assay is intended for use especially in the diagnosis and monitoring of pregnancy.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on Elecsys and **cobas e** immunoassay analyzers.

**Summary**

References<sup>1,2,3,4</sup>

Similarly to LH, FSH and TSH, human chorionic gonadotropin (hCG) is a member of the glycoprotein family and consists of 2 subunits ( $\alpha$ - and  $\beta$ -chains) which are associated to the intact hormone. The  $\alpha$ -chains in all four of these glycoprotein hormones are virtually identical, whereas the  $\beta$ -chains have greatly differing structures and are responsible for the respective specific hormonal functions.

Human chorionic gonadotropin consists of a number of isohormones with differing molecular size. The biological action of hCG serves to maintain the corpus luteum during pregnancy. It also influences steroid production. The serum of pregnant women contains mainly intact hCG.

Measurement of the hCG concentration permits the diagnosis of pregnancy just one week after conception. The determination of hCG in the 1st trimester of pregnancy is of particular importance. Elevated values here serve as an indication of chorionic carcinoma, hydatiform mole or multiple pregnancy. Depressed values indicate threatening or missed abortion, ectopic pregnancy, gestosis or intra-uterine death.

The specific monoclonal antibodies used recognize the holo-hormone. The Elecsys HCG STAT assay should therefore be used in particular in the diagnosis and monitoring of pregnancy. The ruthenium-labeled and biotinylated antibodies used are directed against different epitopes of the hCG molecule.

**Test principle**

Sandwich principle. Total duration of assay: 9 minutes.

Elecsys 2010 and **cobas e 411** analyzers:

- 1st incubation: 10  $\mu$ L of sample, a biotinylated monoclonal hCG-specific antibody, and a monoclonal hCG-specific antibody labeled with a ruthenium complex<sup>a)</sup> react to form a sandwich complex.
- 2nd incubation: After addition of streptavidin-coated microparticles, the complex becomes bound to the solid phase via interaction of biotin and streptavidin.

**cobas e 601** and **cobas e 602** analyzers:

- During a 9 minute incubation, antigen in the sample (10  $\mu$ L), a biotinylated monoclonal hCG-specific antibody, a monoclonal hCG-specific antibody labeled with a ruthenium complex and streptavidin-coated microparticles react to form a sandwich complex, which is bound to the solid phase.

All analyzers:

- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.

- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the reagent barcode.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)<sub>3</sub><sup>2+</sup>)

**Reagents - working solutions**

The reagent rackpack is labeled as HCGSTAT.

- M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL: Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 Anti-hCG-Ab~biotin (gray cap), 1 bottle, 9 mL: Biotinylated monoclonal anti-hCG antibody (mouse) 2.3 mg/L; phosphate buffer 40 mmol/L, pH 7.5; preservative.
- R2 Anti-hCG-Ab~Ru(bpy)<sub>3</sub><sup>2+</sup> (black cap), 1 bottle, 10 mL: Monoclonal anti-hCG antibody (mouse) labeled with ruthenium complex 6.0 mg/L; phosphate buffer 40 mmol/L, pH 6.5; preservative.

**Precautions and warnings**

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines.

Safety data sheet available for professional user on request.

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

**Reagent handling**

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in automatically from the respective reagent barcodes.

**Storage and stability**

Store at 2-8 °C.

Do not freeze.

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability:	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	12 weeks
on the analyzers	4 weeks

**Specimen collection and preparation**

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel.

Li-, Na-, NH<sub>4</sub><sup>+</sup>-heparin, K<sub>3</sub>-EDTA, sodium citrate and sodium fluoride/potassium oxalate plasma.

Criterion: Recovery within 90-110 % of serum value or slope 0.9-1.1 + intercept within  $\pm 2x$  analytical sensitivity (LDL) + coefficient of correlation > 0.95.

Stable for 3 days at 2-8 °C, 12 months at -20 °C. Freeze only once.<sup>5</sup>

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could



# HCG STAT

## Human chorionic gonadotropin, STAT (Short Turn Around Time)

affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

Do not use heat-inactivated samples.

Do not use samples and controls stabilized with azide.

Ensure the samples, calibrators and controls are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples, calibrators and controls on the analyzers should be analyzed/measured within 2 hours.

### Materials provided

See "Reagents – working solutions" section for reagents.

### Materials required (but not provided)

- [REF](#) 03303071190, HCG STAT CalSet, for 4 x 1 mL
- [REF](#) 11731416190, PreciControl Universal, for 2 x 3 mL each of PreciControl Universal 1 and 2
- [REF](#) 11731416160, PreciControl Universal, for 2 x 3 mL each of PreciControl Universal 1 and 2 (for USA)
- [REF](#) 11732277122, Diluent Universal, 2 x 16 mL sample diluent or [REF](#) 03183971122, Diluent Universal, 2 x 36 mL sample diluent
- General laboratory equipment
- Elecsys 2010 or **cobas e** analyzer

Accessories for Elecsys 2010 and **cobas e** 411 analyzers:

- [REF](#) 11662988122, ProCell, 6 x 380 mL system buffer
- [REF](#) 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
- [REF](#) 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
- [REF](#) 11933159001, Adapter for SysClean
- [REF](#) 11706802001, Elecsys 2010 AssayCup, 60 x 60 reaction vessels
- [REF](#) 11706799001, Elecsys 2010 AssayTip, 30 x 120 pipette tips

Accessories for **cobas e** 601 and **cobas e** 602 analyzers:

- [REF](#) 04880340190, ProCell M, 2 x 2 L system buffer
- [REF](#) 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- [REF](#) 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use
- [REF](#) 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change
- [REF](#) 03004899190, PreClean M, 5 x 600 mL detection cleaning solution
- [REF](#) 12102137001, AssayTip/AssayCup Combimagazine M, 48 magazines x 84 reaction vessels or pipette tips, waste bags
- [REF](#) 03023150001, WasteLiner, waste bags
- [REF](#) 03027651001, SysClean Adapter M

Accessories for all analyzers:

- [REF](#) 11298500316, Elecsys SysClean, 5 x 100 mL system cleaning solution
- [REF](#) 11298500160, Elecsys SysClean, 5 x 100 mL system cleaning solution (for USA)

### Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles prior to use and the reading in of the test-specific parameters via the reagent barcode take place automatically. No manual input is necessary. If in exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers.

**cobas e** 601 and **cobas e** 602 analyzers: PreClean M solution is necessary.

Bring the cooled reagents to approx. 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

### Calibration

**Traceability:** This method has been standardized against the 4th International Standard for Chorionic Gonadotropin from the National Institute for Biological Standards and Control (NIBSC) code 75/589.

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

**Calibration frequency:** Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer). Renewed calibration is recommended as follows:

- after 1 month (28 days) when using the same reagent lot
- after 7 days (when using the same reagent kit on the analyzer)
- as required: e.g. quality control findings outside the defined limits

### Quality control

For quality control, use PreciControl Universal.

In addition, other suitable control material can be used.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

### Calculation

The analyzer automatically calculates the analyte concentration of each sample (either in mIU/mL or IU/L).

### Limitations - interference

The assay is unaffected by icterus (bilirubin < 496 µmol/L or < 29 mg/dL), hemolysis (Hb < 0.932 mmol/L or < 1.5 g/dL), lipemia (Intralipid < 2400 mg/dL) and biotin (< 164 nmol/L or < 40 ng/mL).

Criterion: Recovery within ± 10 % of initial value.

Samples should not be taken from patients receiving therapy with high biotin doses (i.e. > 5 mg/day) until at least 8 hours following the last biotin administration.

No interference was observed from rheumatoid factors up to a concentration of 667 IU/mL and samples from dialysis patients.

There is no high-dose hook effect at hCG concentrations up to 500000 mIU/mL.

In vitro tests were performed on 26 commonly used pharmaceuticals. No interference with the assay was found.

In rare cases, interference due to extremely high titers of antibodies to analyte-specific antibodies, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.





# HCG STAT

**Human chorionic gonadotropin, STAT (Short Turn Around Time)**

- 2 Hoermann R, Berger P, Spoettl G, et al. Immunological Recognition and Clinical Significance of Nicked Human Chorionic Gonadotropin in Testicular Cancer. Clin Chem 1994;40(12):2306-2312.
- 3 Schwarz S, Berger P, Wick G. The Antigenic Surface of Human Chorionic Gonadotropin as Mapped by Murine Monoclonal Antibodies. Endocrinology 1986;118(1):189-197.
- 4 Runnebaum B, Rabe T. Gynäkologische Endokrinologie, Grundlagen, Physiologie, Pathologie, Prophylaxe, Diagnostik, Therapie. Berlin, Heidelberg, New York, London, Paris, Tokyo: Springer Verlag, 1987:8,43,489-541.
- 5 Guder WG, Narayanan S, Wisser H, et al. List of Analytes; Preanalytical Variables. Brochure in: Samples: From the Patient to the Laboratory. GIT-Verlag, Darmstadt 1996:16-19. ISBN 3-928865-22-6.
- 6 Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. J Clin Chem Clin Biochem 1988 Nov;26(11):783-790.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

**Symbols**

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

	Contents of kit
	Analyzers/Instruments on which reagents can be used
	Reagent
	Calibrator
	Volume after reconstitution or mixing

**FOR US CUSTOMERS ONLY: LIMITED WARRANTY**

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Significant additions or changes are indicated by a change bar in the margin.

© 2013, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
www.roche.com

Distribution in USA by:  
Roche Diagnostics, Indianapolis, IN  
US Customer Technical Support 1-800-428-2336

