

# Myoglobin STAT CalSet



REF 11820893 122

4 x 1.0 mL

For USA: Elecsys Myoglobin STAT CalSet

**English**

**Intended use**

Myoglobin STAT CalSet is used for calibrating the quantitative Elecsys Myoglobin STAT assay on the Elecsys and **cobas e** immunoassay analyzers.

**Summary**

Myoglobin STAT CalSet is a buffer/protein matrix with added myoglobin in two concentration ranges.

The CalSet can be used with all reagent lots.

**Reagents - working solutions**

- MYO-STAT Cal1: 2 bottles, each containing 1.0 mL of calibrator 1
  - MYO-STAT Cal2: 2 bottles, each containing 1.0 mL of calibrator 2
- Myoglobin (human) in two concentration ranges (approximately 35 ng/mL and approximately 1200 ng/mL) in a liquid buffer/protein (bovine serum albumin) matrix.

The exact lot-specific calibrator values are encoded in the barcode.

**Calibrator values**

Traceability: The Elecsys Myoglobin STAT assay has been standardized against an inhouse reference preparation.

**Precautions and warnings**

For in vitro diagnostic use.  
Exercise the normal precautions required for handling all laboratory reagents.  
Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

For USA: For prescription use only.

All human material should be considered potentially infectious. Materials of human origin have been tested for HIV, hepatitis B, and hepatitis C infections. The findings were negative.

However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.<sup>1,2</sup>

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

**Handling**

The calibrators are supplied ready-for-use in bottles compatible with the system.

Elecsys 2010 and **cobas e** 411 analyzers: The calibrators should only be left on the analyzers during calibration at 20-25 °C. After use, close the bottles as soon as possible and store upright at 2-8 °C.

Due to possible evaporation effects, not more than 5 calibration procedures per bottle set should be performed.

**cobas e** 601 and **cobas e** 602 analyzers: Unless the entire volume is necessary for calibration on the analyzers, transfer aliquots of the ready-for-use calibrators into empty snap-cap bottles (CalSet Vials). Attach the supplied labels to these additional bottles. Store the aliquots at 2-8 °C for later use.

Perform **only one** calibration procedure per aliquot.

*Please note:* Both the vial labels, and the additional labels (if available) contain 2 different barcodes. The barcode between the yellow markers is for **cobas** 8000 systems only. If using a **cobas** 8000 system, please turn the vial cap 180° into the correct position so the barcode can be read by the system. Place the vial on the instrument as usual.

**Storage and stability**

Store at 2-8 °C.

Stability:	
unopened at 2-8 °C	up to the stated expiration date
after opening/in aliquots at 2-8 °C	12 weeks

Stability:	
on Elecsys 2010 and <b>cobas e</b> 411 analyzers at 20-25 °C	up to 5 hours
on <b>cobas e</b> 601 and <b>cobas e</b> 602 analyzers at 20-25 °C	use only once

Store calibrators **upright** in order to prevent the calibrator solution from adhering to the snap-cap.

**Materials provided**

- Myoglobin STAT CalSet, barcode card, 2 x 6 bottle labels

**Materials required (but not provided)**

- REF 11776576322, CalSet Vials, 2 x 56 empty snap-cap bottles
- Elecsys 2010 or **cobas e** immunoassay analyzers and Elecsys Myoglobin STAT assay reagents.

See the assay Method Sheet and the operator's manual for additionally required material.

**Assay**

Place the bottles in the sample zone.

Read in all the information necessary for calibrating the assay.

Ensure the calibrators are at 20-25 °C prior to measurement.

**References**

- Occupational Safety and Health Standards: bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets, the product information and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

**Symbols**

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

	Contents of kit
	Analyzers/Instruments on which reagents can be used
	Reagent
	Calibrator
	Volume after reconstitution or mixing
	Global Trade Item Number

**FOR US CUSTOMERS ONLY: LIMITED WARRANTY**

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS E and ELECSYS are trademarks of Roche.

All other product names and trademarks are the property of their respective owners. Significant additions or changes are indicated by a change bar in the margin.

© 2014, Roche Diagnostics

ms\_11820893122V15.0

# Myoglobin STAT CalSet

**cobas**<sup>®</sup>



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
[www.roche.com](http://www.roche.com)



Distribution in USA by:  
Roche Diagnostics, Indianapolis, IN  
US Customer Technical Support 1-800-428-2336