



## *EC Declaration of Conformity*

**Manufacturer:**

Name: Acro Biotech, Inc.

Address: 9500, 7th str., Unit M, Rancho Cucamonga, CA 91730, USA

**European Representative:**

Name: MedNet GmbH

Address: Borkstrasse 10, 48163 Muenster , Germany

Product Name: Norovirus Rapid Test

Model: Cassette

Classification: Other Device of IVDD 98/79/EC

Conformity Assessment Route: IVDD 98/79/EC Annex III

EDMA Code: 15 70 90 90 00

We herewith declare that the above mentioned products meet the transposition into national law, the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer.

### DIRECTIVES

**General applicable directives:**

DIRECTIVE 98/79/EC OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 27 October 1998 on in vitro diagnostic medical devices

Standard Applied: EN ISO13485:2012/AC:2012, EN ISO14971:2012, EN 13975:2003, EN ISO 18113-1:2011, EN ISO 18113-2:2011, EN 13612:2002/AC:2002, EN ISO 17511:2003, EN ISO 15193:2009, EN ISO 15194:2009, EN 13640:2002, EN 13641:2002, EN 1041:2008, ISO 15223-1:2012

Place, Date of Issue: in Rancho Cucamonga on 13/09/2016

Signature: \_\_\_\_\_

Name: Joseph Fan

Position: President

ACRO BIOTECH, INC