

Declaration of Conformity to 98/79/EC

Manufacturer: Ventana Medical Systems, Inc.
1910 E. Innovation Park Drive
Tucson, AZ USA 85755

European Authorized Representative: ROCHE DIAGNOSTICS GmbH
Sandhofer Strasse 116
D-68305 Mannheim
Germany

Manufacturing Site: Tucson, AZ USA

		Ventana	Roche
		<input type="checkbox"/> REF	
Product name(s)/ Number(s):	HER2 Dual ISH 3 in 1 Xenograft Slides	783-4422	05640300001

Technical Data File: TDF 0408

Classification: General IVD


Conformity Assessment: 98/79/EC Annex III

Ventana Medical Systems, Inc. declares that the product(s) listed is/are in conformity with the essential requirements of Annex I of Council Directive 98/79/EC. All supporting documentation is retained under the premise of the Manufacturer.

Ventana Medical Systems, Inc. maintains a quality system based on ISO 13485.

Place of Issue: Tucson, AZ USA 85755

Name of Authorized Signatory: Sam M. Rua
Senior Director, Regulatory Affairs

Signature:  Date: 10 JUN 2011