Juventas Bogusław Ziółkowski
Srogów Górny 172
38-507 Jurowce
tel. kom. +48 608 074 052
tel. kom. +48 664 975 612
e-mail: biuro@juventas.pl
web: www.juventas.pl
NIP: 6631799694
REGON: 260108888

Key features and technical description
Examination chair JFG5

- Chair JFG 5 has been designed to perform gynecological and urological examination procedures and is used in rehabilitation, medical and healthcare centers. 1.
- It has epoxy polyester coated steel frame and a comfortable sleeping area. 2.
- The three-section chair also has a height-adjustable head section which is adjustable by the means of gas springs. Foot section can be removed or placed to extend the laying surface. 5. 6.
- The height of the chair is operated by electrical motor, adjustable by hand remote control. 4.
- High density polyurethane padding lying surface is 5 cm thick and covered with seamless durable, stain resistant, easy to clean, disinfectant, blood and urine stains resistant, liquid repellent synthetic leather cover. Its color is available in many different choices. 13.
- Couple padded and height adjustable footrests, which could be rotated any direction by help of ball bearing
- Stainless steel removable tray under the seat and head pillow comes as standard options.
- Stainless steel guides on both sides of the seat. 17.4

Technical parameters:

Width : 65 cm 9.
Height : from 70 cm to 105 cm 10.
Load capacity : 150 kg 16.
Weight : 83 kg
Back rest inclination angle : from -25 ° to + 80 ° 11.
Length (with leg support) : 188 cm 8.
Length (without leg support) : 138 cm
Trendelenburg tilt angle gas spring: 0 ° to +25 ° 7. 12.
Anti-Trendelenburg heel angle gas spring: from 0 ° to -25 ° 7. 12.
Permissible leg support load : 40 kg
Voltage: 220V/50Hz 15.

Frame color – signal white RAL 9003, other colors available under request 14.
Additional options: paper roll holder, metal shelf, bumpers, wheels.

Signature: Ziółkowski B.

Bogusław Ziółkowski
Company Owner

JUVENTAS
JUVENTAS Bogusław Ziółkowski
Srogów Górny 172, 38-507 Jurowce
tel./fax +48 13 463 80 35
NIP 6631799694 REGON 260108888

Pagrindinės savybės ir techninis aprašymas
Apžiūrų kėdė JFG 5

• Kėdė JFG 5 buvo sukurta ginekologinėms ir urologinėms apžiūros procedūroms ir gydymui atlikti, naudojama reabilitacijos, medicinos ir sveikatos priežiūros centruose.
• Ji turi epoksio poliesterio dažis milteliniu būdu dengtą rėmą ir patogų gulimąjį paviršių.
• Trijų sekcijų kėdė taip pat turi reguliuojamo aukščio galvos dalį dujinės spyruoklės pagalba. Kojų dalis gali būti lengvai uždedama ir nuimama norint prailginti gulimąjį paviršių.
Kėdės aukštis reguliuojamas elektros Varikliu, reguliuojamas rankiniu pulteliu
• Didelio tankio poliuretano paminkštinimas yra 5 cm storio ir padengtas patvaria besiūle, dėmėms atsparia, lengvai valoma, atsparia dezinfekavimo, kraujo ir šlapimo dėmėms atsparia, skysčiams nelaidžia sintetinės odos danga. Jos spalvą galima pasirinkti iš daugybės skirtingų variantų.
Pora paminkštintų kojų atramų, reguliuojamo aukščio ir gylio, kurie gali būti rotuojami visomis kryptimis rutulinio guolio deka
Nerūdijančio plieno išimamas indas ir pagalvėlė galvai -standartiniai priedai
Nerūdijančio plieno bėgeliai abiejose sėdynės pusėse.

Techniniai parametrai:
Plotis: 65 cm
Aukštis: nuo 70 cm iki 105 cm
Keliamoji galia: 150 kg
Svoris: 83 kg
Nugaros atramos pasvirimo kampas: nuo -25 ° iki + 80 °
Ilgis (su kojų atrama): 188 cm
Ilgis (be kojos atramos): 138 cm
Trendelenburgo pasvirimo kampas: nuo 0 ° iki +25 °
Anti-Trendelenburg kulno kampas: nuo 0 ° iki -25 °
Leistina kojų atramos apkrova: 40 kg

Rėmo spalva – signal white RAL 9003 arba kita spalva pagal pareikalavimą
Papildomi priedai: popieriaus rulono laikiklis, metalinė lentynėlė, buferiai, ratukai