



Product Service

CERTIFICATE

No. Z1 12 08 67069 002

Holder of Certificate: **Linet spol. s.r.o.**

Želevcice 5
274 01 Slaný
CZECH REPUBLIC

**Production
Facility(ies):**

67069

Certification Mark:**Product:****Medical Beds****Model(s):****Eleganza 1****Parameters:**


Application environment:	1 / 2 / 3
Safe working load:	250 kg
Max. patient weight:	185 kg
Dimensions:	218 x 100 cm
Mattress support platform:	200 x 90 cm
Height adjustment:	40 and 76 cm
Rated voltage:	230 VAC
Rated frequency:	50/60 Hz
Rated power:	max. 370 VA
Degree of protection:	IPX4
Protection class:	I
Applied part:	Type B
Mode of operation:	10%, max 2min/18min
Angle A:	73°
Angle B:	18,5°
Angle C:	90°
Angle D:	23°

Tested according to: EN 60601-2-52:2010
EN 60601-1:2006

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 713010512

Date, 2012-08-23


(Torsten Zimmer)

Page 1 of 1

