

| REF | CONTENT | SYSTEM |
|--------------|---------|-------------|
| 06334601 001 | ▽ 400 | cobas u 601 |

English**Caution**

Do not open the inner bag prior to use.
Immediately insert cassette into analyzer!

Intended use

The cobas u pack is a cassette with teststrips for the in vitro qualitative or semi-quantitative determination of pH, leukocytes, nitrite, protein, glucose, ketones, urobilinogen, bilirubin, color and erythrocytes in urine with the cobas u 601 urine analyzer. These measurements are useful in the evaluation of renal, urinary, hepatic and metabolic disorders.

For professional use only.

Test principle

Reference¹

pH: The test paper contains the indicators methyl red, phenolphthalein and bromothymol blue and reacts specifically with H⁺-ions. The most frequent pH values of fresh urine from healthy subjects lie between 5 and 6.

Leukocytes (LEU): The test reveals the presence of granulocyte esterases. These esterases cleave an indoxyl ester, and the indoxyl so liberated reacts with a diazonium salt to produce a violet dye. Bacteria, trichomonads or erythrocytes present in the urine do not affect the reaction.

Nitrite (NIT): The test is based on the principle of the Griess test and is specific for nitrite. The reaction reveals the presence of nitrite and hence indirectly nitrite-forming bacteria in the urine by a pink-to-red coloration of the test patch. Even a slight pink coloration is indicative of significant bacteriuria.

Protein (PRO): The test is based on the principle of the protein error of a pH indicator. It is particularly sensitive to albumin. An elevated pH (up to 9) does not affect the test.

Glucose (GLU): The glucose determination is based on the specific glucose-oxidase/peroxidase reaction (GOD/POD method). The test is independent of the pH and specific gravity of the urine and is not affected by the presence of ketone bodies.

Ketone bodies (KET): This test is based on the principle of Legal's test and is more sensitive to acetoacetic acid than to acetone.

Urobilinogen (UBG): A stable diazonium salt reacts almost immediately with urobilinogen to give a red azo dye. The test is specific for urobilinogen and is not susceptible to the interfering factors known to affect the Ehrlich's test.

Bilirubin (BIL): The test is based on the coupling of bilirubin with a diazonium salt. Even the slightest pink coloration constitutes a positive, i.e. pathologic, result. Other urinary constituents produce a more or less intense yellow coloration.

Blood (ERY/Hb): The peroxidase-like action of hemoglobin and myoglobin specifically catalyzes the oxidation of the indicator by means of the organic hydroperoxide contained in the test paper to give a blue-green coloration.

Compensation area (COMP): This white area, which is not impregnated with reagents, allows instrumental compensation for the intrinsic color of the urine while testing leukocytes, nitrite, glucose, ketone bodies, urobilinogen, bilirubin, erythrocytes and determination of the urine color (COL).

Reagents

Each test contains per 1 cm² test patch area the following:

pH: Bromothymol blue 13.9 µg; methyl red 1.2 µg; phenolphthalein 8.6 µg

Leukocytes: Indoxylcarbonic acid ester 15.5 µg; methoxymorpholinobenzene diazonium salt 5.5 µg

Nitrite: 3-hydroxy-1,2,3,4-tetrahydro-7,8-benzoquinoline 33.5 µg; sulfanilamide 29.1 µg

Protein: 3',3'',5',5''-tetrachlorophenol-3,4,5,6-tetrabromosulfophthalein 13.9 µg

Glucose: 3,3',5,5'-tetramethylbenzidine 103.5 µg; GOD 6 U, POD 35 U

Ketone bodies: Sodium nitroprusside 157.2 µg

Urobilinogen: 4-methoxybenzene-diazonium-tetrafluoroborate 67.7 µg

Bilirubin: 2,6-dichlorobenzene-diazonium-tetrafluoroborate 16.7 µg

Blood: 3,3',5,5'-tetramethylbenzidine 52.8 µg; 2,5-dimethyl-2,5-dihydroperoxyhexane 297.2 µg

Precautions and warnings

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Note: If the cassette has been stored refrigerated, it must be left at room temperature for a minimum of one hour prior to use.

The cassette contains a non-toxic silicate-based desiccant which must not be removed. If ingested by accident, drink large quantities of water.

Disposal of all waste material should be in accordance with local guidelines.

Safety data sheet available for professional user on request.

Reagent handling

Ready for use.

Storage and stability

Store the cassette at 2-30 °C.

After loading the cassette into the analyzer, the test strips are stable within the tightly closed cassette compartment for 14 days. After this period, the cassette has to be replaced by a new one.

Do not use the cassette after the specified expiry date.

Specimen collection and preparation

Reference¹

For specimen collection and preparation only use suitable tubes or collection containers.

Use fresh urine that has not been centrifuged. The urine specimen should not stand for more than 2 hours before testing. In case of longer standing, mix before use.

Use only clean, well-rinsed vessels to collect urine.

Do not add preservatives to the urine.

Do not expose urine specimens to sunlight as this induces oxidation of bilirubin and urobilinogen and hence leads to artificially low results for these two parameters.

Materials provided

- [REF] 06334601001, Cassette with 400 test strips

Materials required (but not provided)

- [REF] 06390498001, cobas u 601 urine analyzer
- [REF] 06390579001, cobas u calibration strip
- Controls as indicated below
- General laboratory equipment

Assay

1. Unfold and cut open aluminum bag with scissors (illustration 1).
2. Remove test strip cassette from the packaging and remove the two protection pads (illustration 2).
3. Immediately place test strip cassette into the cobas u 601 urine analyzer (illustration 3).

Follow the instructions in the operator's manual of the instrument for correct insertion and positioning. These instructions also contain information on further handling precautions of the cassette.



Note: If the cassette is stored in the opened bag or exposed to air (humidity, nitrogen oxides) for more than 3 minutes, environmental conditions may cause a color change of the test patches and damage of the reagents.

This has to be avoided. Do not use the cassette if the packaging shows severe damages, or the test strip layers in the cassette are not correctly aligned, or the test strips show unusual coloring.

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Calibration

cobas u calibration strips are used for the calibration of the **cobas u 601** urine analyzer. For details see the operator's manual of the analyzer.

Quality control

For quality control, use commercially available urine controls, or other suitable control material.

Following quality controls from BIO-RAD are recommended to use:

- qUAntify Plus Control
- Liquichek Urinalysis Control

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

Important note for reporting results:

According to the regulations from the German Medical Association for quality assurance of medical laboratory analyses dated 11/23/2007, the decision to classify a laboratory test result to either part B1 or B2 depends on the way the test results are expressed in the report (scale level).

The specification in the report defines whether a determination is quantitative or qualitative and therefore which legal requirements for quality assurance (B1 for quantitative or B2 for qualitative) need to be followed.

Examples of qualitative characteristics include titer levels, concentrations/color ranges (+ to +++) or a defined range of values. A characteristic of a quantitative value is when the value has a corresponding measured unit value.

Limitations - interference

Therapeutic drugs and endogenous substances were tested for a potential interference to the test parameters of the **cobas u pack**.

All parameters were tested with negative urine samples and samples spiked to the first positive concentration range.

Therapeutic drugs were tested at concentrations in urine occurring under medication with the therapeutic dosage and above.

The following therapeutic drugs and concentrations were tested:

| Therapeutic drugs | |
|--------------------|------------------------------|
| Substance | Maximum tested concentration |
| Acetaminophen | 3000 mg/L |
| N-Acetylcystein | 200 mg/L |
| Amoxicillin | 10000 mg/L |
| Amlodipin-besylat | 33.3 mg/L |
| Ascorbic Acid | 4000 mg/L |
| Cefoxitin | 12000 mg/L |
| Cetirizin | 66.6 mg/L |
| Cotrimoxazol | 6000 mg/L |
| Cyclosporin | 80 mg/L |
| Furosemid | 3333 mg/L |
| Gentamycin-sulfat | 400 mg/L |
| Hydrochlorothiazid | 333 mg/L |
| Hydroxychloroquin | 1333 mg/L |
| Ibuprofen | 2500 mg/L |
| Levodopa | 1250 mg/L |
| Levothyroxine | 1.0 mg/L |

| Therapeutic drugs | |
|-------------------|------------------------------|
| Substance | Maximum tested concentration |
| Lisinopril | 133.3 mg/L |
| Methyldopa | 2000 mg/L |
| Ofloxazin | 900 mg/L |
| Phenazopyridin | 300 mg/L |
| Salicyluric Acid | 6000 mg/L |
| Tetracyclin | 500 mg/L |

For drug testing:

Within this range the following observations were made:

| Parameter | Therapeutic drug | No interference up to | Effect above stated concentration |
|-----------|-------------------|-----------------------|--|
| LEU | Cotrimoxazol | 500 mg/L | elevated positive results |
| | Furosemide | 1000 mg/L | elevated positive results |
| | Salicyl Uric Acid | 3000 mg/L | false negative results |
| NIT | Ascorbic Acid | 1500 mg/L | false negative results |
| PRO | Cotrimoxazol | 1200 mg/L | false positive results |
| | Hydrochloroquine | 600 mg/L | false positive results |
| | Phenazopyridine | 200 mg/L | false positive results |
| GLU | Ascorbic Acid | 400 mg/L | false normal results |
| KET | N-Acetylcysteine | 30 mg/L | false positive results and elevated positive results |
| | Levodopa | 250 mg/L | elevated positive results |
| | Methyldopa | 700 mg/L | elevated positive results |
| UBG | Phenazopyridine | 200 mg/L | false positive results and elevated positive results |
| BIL | Cotrimoxazol | 5000 mg/L | false negative results |
| | Phenazopyridine | 100 mg/L | false positive results and elevated positive results |
| ERY | Ascorbic Acid | 700 mg/L | false negative results |
| | Cotrimoxazol | 3000 mg/L | false negative results |
| | Furosemide | 1500 mg/L | false negative results |
| | Hydrochloroquine | 200 mg/L | false negative results |
| | Ibuprofen | 800 mg/L | false negative results |

Endogenous substances were tested at abnormal high concentrations.

| Endogenous substances | |
|-----------------------|------------------------------|
| Substance | Maximum tested concentration |
| Ammonium | 25000 mg/L |
| Calcium | 3000 mg/L |
| Creatinin | 15000 mg/L |
| Glucose | 50000 mg/L |
| Hemoglobin | 750 mg/L |
| β-Hydroxybutyrat | 4500 mg/L |
| Immunglobulin G | 5000 mg/L |
| Nitrit | 110 mg/L |
| Urea | 200000 mg/L |
| Uric acid | 1550 mg/L |
| Urobilinogen | 3000 mg/L |
| pH | 4.5 - 9.0 |

For endogenous substance testing within this range the following observations were made:

| Parameter | Endogenous substance | No interference up to | Effect above stated concentration |
|-----------|----------------------|---|--|
| LEU | Calcium | 2200 mg/L | elevated positive results |
| | Urobilinogen | 150 mg/L | false positive results and elevated positive results |
| NIT | Hemoglobin | 259 mg/L | false positive results |
| | Urobilinogen | 120 mg/L | false positive results |
| PRO | Hemoglobin | false positive and false elevated results due to unspecific protein detection may occur | |
| | Ammonium | 5000 mg/L | false negative results |
| | Creatinin | 7500 mg/L | false positive results and elevated positive results |
| | Urea | 120000 mg/L | false positive results and elevated positive results |
| | Urobilinogen | 1000 mg/L | false positive results |
| GLU | Ammonium | 15000 mg/L | false normal results |
| | Urea | 120000 mg/L | false normal results |
| | Urobilinogen | 2500 mg/L | false normal results |
| KET | Creatinin | 3000 mg/L | elevated positive results |
| | Hemoglobin | 259 mg/L | elevated positive results |
| | Urobilinogen | 1500 mg/L | false negative results |
| UBG | Nitrite | 10 mg/L | false normal results |
| BIL | Nitrite | 20 mg/L | false negative results |
| ERY | Nitrite | 10 mg/L | false negative results |
| | Urobilinogen | 120 mg/L | false positive results |

Common limitations:

NIT:² Prolonged urinary retention in the bladder (4-8 hours) is essential in order to obtain an accurate result. Administration of antibiotics or chemical drugs should be discontinued 3 days before the test.

Attention: Nitrogen oxides present in the atmosphere may have an influence on the stability of the nitrite test pad.²

PRO: False positive readings may be found after infusion of polyvinylpyrrolidone (blood substitute).

ERY: The result values refer to intact erythrocytes. At concentrations of about 5-50 Ery/μL, significant hemolysis (such as may occur on prolonged standing of the urine) leads to values which are higher than the corresponding concentrations given for intact erythrocytes. In women the test for blood may be falsified from 3 days before to 3 days after menstruation. It is therefore advisable not to perform the test during this time. After physical activity, e.g. strenuous jogging, raised values for erythrocytes and protein may occur without being signs of disease.

Note:

Knowledge of the effects of drugs or their metabolites upon the individual tests is not yet complete. In doubtful cases, it is therefore advisable to repeat the test after discontinuing a particular drug.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

Expected values

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

For the **cobas u 601** urine analyzer please refer to Appendix 1.

Result values

Please refer to Appendix 1.

Specific performance data

Representative performance data are given below. Results obtained in individual laboratories may differ.

The values for NEG and POS indicate the rate of concordant negative or positive results.

For the **cobas u 601** urine analyzer please refer to Appendix 2.

Precision

Precision experiments comprised an assessment of repeatability (within-run precision) and intermediate precision.

Repeatability was checked in 2 separate runs with 21 measurements each for the tested controls. In total there were 42 measurements performed per used control.

Intermediate precision was assessed over 21 days with 2 runs per day and duplicate measurements per used control. In total there were 84 measurements performed per used control.

The following results were obtained:

| Repeatability | | | |
|---------------|-------------------|------------|-----------------|
| Parameter | Control | Result | Exact agreement |
| | BIO-RAD Liquichek | | |
| pH | Level 1 | 6.5 | 100 % |
| | Level 2 | 7 | 100 % |
| LEU | Level 1 | NEG | 100 % |
| | Level 2 | 500 Leu/μL | 100 % |
| NIT | Level 1 | NEG | 100 % |
| | Level 2 | POS | 100 % |
| PRO | Level 1 | NEG | 100 % |
| | Level 2 | 150 mg/dL | 100 % |
| GLU | Level 1 | NORM | 100 % |
| | Level 2 | 1000 mg/dL | 100 % |

| Repeatability | | | |
|---------------|------------------------------|------------------|--------------------|
| Parameter | Control BIO-RAD Liquichek | Result | Exact agreement |
| KET | Level 1 | NEG | 100 % |
| | Level 2 | 150 mg/dL | 100 % |
| UBG | Level 1 | NORM | 100 % |
| | Level 2 | 12 mg/dL | 100 % |
| BIL | Level 1 | NEG | 100 % |
| | Level 2 | 6 mg/dL | 100 % |
| ERY | Level 1 | NEG | 100 % |
| | Level 2 | 250 Ery/ μ L | 100 % |
| COL | Level 1 | Yellow | 100 % |
| | Level 2 | Brown | 100 % |

| Intermediate precision | | | |
|------------------------|------------------------------|------------------|--------------------|
| Parameter | Control BIO-RAD Liquichek | Result | Exact agreement |
| pH | Level 1 | 6.5 | 100 % |
| | Level 2 | 7 | 100 % |
| LEU | Level 1 | NEG | 100 % |
| | Level 2 | 500 Leu/ μ L | 100 % |
| NIT | Level 1 | NEG | 100 % |
| | Level 2 | POS | 100 % |
| PRO | Level 1 | NEG | 100 % |
| | Level 2 | 150 mg/dL | 100 % |
| GLU | Level 1 | NORM | 100 % |
| | Level 2 | 1000 mg/dL | 100 % |
| KET | Level 1 | NEG | 100 % |
| | Level 2 | 150 mg/dL | 100 % |
| UBG | Level 1 | NORM | 100 % |
| | Level 2 | 12 mg/dL | 100 % |
| BIL | Level 1 | NEG | 100 % |
| | Level 2 | 6 mg/dL | 100 % |
| ERY | Level 1 | NEG | 100 % |
| | Level 2 | 250 Ery/ μ L | 100 % |
| COL | Level 1 | Yellow | 100 % |
| | Level 2 | Brown | 100 % |

Method comparison

Results of a method comparison of **cobas u 601** urine analyzer with **cobas u 411** urine analyzer and **cobas u 601** urine analyzer with URISYS 2400 Analyzer (for COL) using at least 1348 clinical samples are presented below in Appendix 2.

The method comparison results for specific gravity and clarity are presented in the **cobas u 601** urine analyzer Operator Manual.

Analytical sensitivity

The values specified for the analytical sensitivity are defined as the concentration of the analyte which leads to a positive result in > 90 % of the examined urines.

Please refer to Appendix 2.

Appendix 1

| Parameter | Expected values | Result values |
|-----------|--------------------|---|
| pH | 4.8 – 7.4 | 5, 6, 6.5, 7, 8, 9 |
| LEU | < 10 Leu/ μ L | NEG, 25, 100, 500 Leu/ μ L |
| NIT | – | NEG, POS |
| PRO | < 10 mg/dL | NEG, 25, 75, 150, 500 mg/dL |
| | < 0.1 g/L | NEG, 0.25, 0.75, 1.5, 5.0 g/L |
| GLU | < 30 mg/dL | NORM, 50, 100, 300, 1000 mg/dL |
| | < 1.7 mmol/L | NORM, 3, 6, 17, 56 mmol/L |
| KET | < 5 mg/dL | NEG, 5, 15, 50, 150 mg/dL |
| | < 0.5 mmol/L | NEG, 0.5, 1.5, 5, 15 mmol/L |
| UBG | < 1 mg/dL | NORM, 1, 4, 8, 12 mg/dL |
| | < 17 μ mol/L | NORM, 17, 68, 135, 203 μ mol/L |
| BIL | < 0.2 mg/dL | NEG, 1, 3, 6 mg/dL |
| | < 3.4 μ mol/L | NEG, 17, 50, 100 μ mol/L |
| ERY/Hb | 0 – 5 Ery/ μ L | NEG, 10, 25, 50, 150, 250 Ery/ μ L |
| COL | – | pale yellow, yellow, amber, brown, orange, red, green, others |

Appendix 2

| Parameter | Analytical sensitivity | Method comparison |
|-----------|---------------------------|---|
| pH | N. A. | Ident.: 77 % pH 5+6: 98 % pH 8+9: 88% |
| LEU | 20 - 30 Leu/ μ L | NEG: 91 % POS: 97 % |
| NIT | 0.05 - 0.06 mg/dL | NEG: 95 % POS: 94 % |
| PRO | 8 - 12 mg/dL albumin | NEG: 96 % POS: 93 % |
| GLU | 30 - 40 mg/dL | NEG: 98 % POS: 100 % |
| KET | 3 - 6 mg/dL | NEG: 94 % POS: 96 % |
| UBG | 1.0 - 1.4 mg/dL | NEG: 96 % POS: 98 % |
| BIL | 0.4 - 0.6 mg/dL | NEG: 93 % POS: 95 % |
| ERY | 5 - 10 Ery/ μ L | NEG: 95 % POS: 96 % |

cobas u pack

cobas u pack



| Parameter | Analytical sensitivity | Method comparison |
|-----------|------------------------|--|
| COL | N. A. | Pale yellow+yellow: 94 % Amber: 73 % Brown: 92 % Red: 100 % |

References

- 1 REF 1225432001, Compendium of urinalysis
- 2 European Urinalysis Guidelines

For further information, please refer to the appropriate operator's manual for the relevant analyzer and Method Sheets of all necessary components.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard.

| | |
|---|---|
|  | Contents of kit |
|  | Analyzers/Instruments on which reagents can be used |
|  | Reagent |
|  | Calibrator |
|  | Volume after reconstitution or mixing |

COBAS, COBAS U and URISYS 2400 are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Significant additions or changes are indicated by a change bar in the margin.

© 2013, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

